


LOT PX 016
Rev 1


CONTROL

 (Exp.) 2016-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.3	2.4	2.5	± 0.4	7.4	7.4	7.3	7.4	7.3	± 1.0	18.0	17.7	17.8	17.7	18.0	± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.30	2.34	2.30	2.34	± 0.16	4.61	4.57	4.60	4.57	4.53	± 0.20	5.15	5.16	5.16	5.16	5.17	± 0.25	
HB	HGB	g/dl	6.8	6.7	6.7	6.7	6.7	± 0.4	13.9	13.8	13.8	13.8	13.7	± 0.5	16.7	16.7	16.8	16.7	16.7	± 0.6	
		g/l	68	67	67	67	67	± 4	139	138	138	138	137	± 5	167	167	168	167	167	± 6	
		mmol/l	4.22	4.16	4.16	4.16	4.16	± 0.25	8.63	8.57	8.57	8.57	8.51	± 0.31	10.37	10.37	10.43	10.37	10.37	± 0.37	
HT	HCT	%	19.0	18.9	18.5	18.9	18.7	± 1.5	37.8	37.5	37.3	37.5	37.1	± 2.0	45.3	45.4	44.9	45.4	45.5	± 2.5	
		l/l	0.190	0.189	0.185	0.189	0.187	± 0.015	0.378	0.375	0.373	0.375	0.371	± 0.020	0.453	0.454	0.449	0.454	0.455	± 0.025	
VGM	MCV	µm ³ ; fl	80	82	79	82	80.0	± 5	82	82	81	82	82.0	± 5	88	88	87	88	88.0	± 5	
TGMH	MCH	pg	28.7	29.1	28.6	29.1	28.6	± 2.0	30.2	30.2	30.0	30.2	30.2	± 2.0	32.4	32.4	32.6	32.4	32.3	± 2.5	
		fmol	1.78	1.81	1.78	1.81	1.78	± 0.12	1.87	1.88	1.86	1.88	1.88	± 0.12	2.01	2.01	2.02	2.01	2.01	± 0.16	
CCMH	MCHC	g/dl	35.9	35.5	36.2	35.5	35.8	± 3.0	36.8	36.8	37.0	36.8	36.9	± 3.0	36.8	36.8	37.4	36.8	36.7	± 3.0	
		g/l	359	355	362	355	358	± 30	368	368	370	368	369	± 30	368	368	374	368	367	± 30	
		mmol/l	22.27	22.06	22.51	22.06	22.23	± 1.86	22.83	22.87	23.00	22.87	22.90	± 1.86	22.88	22.84	23.24	22.84	22.79	± 1.86	
IDR	RDW	%	13.3	12.8	12.7	12.8	12.0	± 4.0	12.6	12.7	12.0	12.7	12.3	± 4.0	12.5	12.6	11.7	12.6	12.5	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	66	67	66	70	± 20	252	255	260	255	255	± 30	495	495	520	495	498	± 50	
VPM	MPV	µm ³ ; fl	9.5	9.5	9.4	9.5	8.9	± 2.0	9.0	9.1	9.0	9.1	8.4	± 2.0	8.7	8.9	8.7	8.9	8.2	± 2.0	
NEUT	#		1.32	1.32	1.27	1.32	1.43	± 0.35	4.18	4.23	4.20	4.23	4.20	± 0.90	12.65	12.32	12.64	11.95	12.87	± 1.90	
		%	54.8	55.0	55.3	55.0	57.0	± 10.0	56.5	57.2	57.5	57.2	57.5	± 10.0	70.3	69.6	71.0	67.5	71.5	± 10.0	
LYMPHO	#		0.77	0.77	0.72	0.77	0.78	± 0.33	2.37	2.32	2.26	2.32	2.34	± 0.70	2.88	2.90	2.85	2.74	2.79	± 1.50	
		%	32.0	32.0	31.4	32.0	31.2	± 12.0	32.0	31.3	31.0	31.3	32.0	± 8.0	16.0	16.4	16.0	15.5	15.5	± 8.0	
MONO	#		0.07	0.07	0.07	0.07	0.08	± 0.07	0.32	0.30	0.31	0.30	0.26	± 0.26	0.72	0.71	0.71	0.55	0.54	± 0.54	
		%	3.0	2.8	3.0	2.8	3.0	± 2.8	4.3	4.0	4.3	4.0	3.5	± 3.5	4.0	4.0	4.0	3.1	3.0	± 3.0	
EOS	#		0.17	0.17	0.16	0.17	0.14	± 0.14	0.27	0.30	0.27	0.30	0.26	± 0.26	0.94	1.06	0.89	1.68	1.08	± 0.89	
		%	7.0	7.0	7.0	7.0	5.5	± 5.5	3.7	4.0	3.7	4.0	3.5	± 3.5	5.2	6.0	5.0	9.5	6.0	± 5.0	
BASO	#		0.08	0.08	0.08	0.08	0.08	± 0.08	0.26	0.26	0.26	0.26	0.26	± 0.26	0.81	0.71	0.71	0.78	0.72	± 0.71	
		%	3.2	3.2	3.3	3.2	3.3	± 3.2	3.5	3.5	3.5	3.5	3.5	± 3.5	4.5	4.0	4.0	4.4	4.0	± 4.0	

LOT PX 016
Rev 1

CONTROL

 (Exp.) 2016-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120	DX120	DX NEXUS		120	DX120	DX NEXUS		120	DX120	DX NEXUS					
120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS										
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.6	7.6	7.6		± 1.0	18.7	18.7	18.7		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.37	2.37		± 0.16	4.61	4.61	4.61		± 0.20	5.24	5.24	5.24		± 0.25	
HB	HGB	g/dl	6.8	6.8	6.8		± 0.4	13.8	13.8	13.8		± 0.5	16.6	16.6	16.6		± 0.6	
		g/l	68	68	68		± 4	138	138	138		± 5	166	166	166		± 6	
		mmol/l	4.22	4.22	4.22		± 0.25	8.57	8.57	8.57		± 0.31	10.31	10.31	10.31		± 0.37	
HT	HCT	%	19.7	19.7	19.7		± 1.5	38.7	38.7	38.7		± 2.0	46.6	46.6	46.6		± 2.5	
		l/l	0.197	0.197	0.197		± 0.015	0.387	0.387	0.387		± 0.020	0.466	0.466	0.466		± 0.025	
VGM	MCV	µm ³ ; fl	83	83	83		± 5	84	84	84		± 5	89	89	89		± 5	
TGMH	MCH	pg	28.7	28.7	28.7		± 2.0	29.9	29.9	29.9		± 2.0	31.7	31.7	31.7		± 2.5	
		fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86		± 0.12	1.97	1.97	1.97		± 0.16	
CCMH	MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.6	35.6	35.6		± 3.0	35.6	35.6	35.6		± 3.0	
		g/l	346	346	346		± 30	356	356	356		± 30	356	356	356		± 30	
IDR	RDW	%	15.0	15.0	15.0		± 4.0	15.0	15.0	15.0		± 4.0	13.8	13.8	13.8		± 4.0	
		µm ³ ; fl	9.4	9.4	9.4		± 2.0	9.0	9.0	9.0		± 2.0	8.6	8.6	8.6		± 2.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	73	73		± 20	265	265	265		± 30	505	505	505		± 50	
VPM	MPV	µm ³ ; fl	9.4	9.4	9.4		± 2.0	9.0	9.0	9.0		± 2.0	8.6	8.6	8.6		± 2.0	
NEUT	#	%	1.39	1.45	1.45		± 0.35	4.35	4.50	4.50		± 0.90	13.40	13.60	13.60		± 1.90	
		%	55.7	58.1	58.1		± 10.0	57.3	59.2	59.2		± 10.0	71.4	72.9	72.9		± 10.0	
LYMPHO	#	%	0.79	0.72	0.72		± 0.33	2.33	2.17	2.17		± 0.70	2.92	2.71	2.71		± 1.50	
		%	31.7	28.8	28.8		± 12.0	30.6	28.6	28.6		± 8.0	15.6	14.5	14.5		± 8.0	
MONO	#	%	0.10	0.10	0.10		± 0.10	0.41	0.42	0.42		± 0.41	1.07	1.05	1.05		± 1.05	
		%	3.9	4.0	4.0		± 3.9	5.4	5.5	5.5		± 5.4	5.7	5.6	5.6		± 5.6	
EOS	#	%	0.16	0.17	0.17		± 0.16	0.27	0.28	0.28		± 0.27	0.84	0.84	0.84		± 0.84	
		%	6.2	6.6	6.6		± 6.2	3.5	3.7	3.7		± 3.5	4.5	4.5	4.5		± 4.5	
BASO	#	%	0.06	0.06	0.06		± 0.06	0.24	0.23	0.23		± 0.23	0.52	0.47	0.47		± 0.47	
		%	2.5	2.5	2.5		± 2.5	3.2	3.0	3.0		± 3.0	2.8	2.5	2.5		± 2.5	

9930096-A
FRONT / RECTO
Ref: TEMP-0821 Rev.36