

ABX Difftrol



LOT PX 036
Rev 4

CONTROL

(Exp.) 2016-05-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.4	2.5	2.5	± 0.4	7.4	7.4	7.3	7.4	7.5	± 1.0	17.7	17.5	17.5	17.5	18.1	± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.28	2.23	2.27	2.23	2.26	± 0.16	4.58	4.60	4.56	4.60	4.55	± 0.20	5.04	5.07	5.05	5.07	5.08	± 0.25	
HB	HGB	g/dl	6.3	6.3	6.3	6.3	6.3	± 0.4	13.3	13.3	13.3	13.3	13.3	± 0.5	16.1	16.1	16.1	16.1	16.1	± 0.6	
		g/l	63	63	63	63	63	± 4	133	133	133	133	133	± 5	161	161	161	161	161	± 6	
HT	HCT	mmol/l	3.91	3.91	3.91	3.91	3.91	± 0.25	8.26	8.26	8.26	8.26	8.26	± 0.31	10.00	10.00	10.00	10.00	10.00	± 0.37	
		%	18.0	18.1	17.7	18.1	18.1	± 1.5	37.1	37.3	36.5	37.3	36.9	± 2.0	44.4	44.6	43.4	44.6	44.7	± 2.5	
VGM	MCV	l/l	0.180	0.181	0.177	0.181	0.181	± 0.015	0.371	0.373	0.365	0.373	0.369	± 0.020	0.444	0.446	0.434	0.446	0.447	± 0.025	
TGMH	MCH	µm ³ ; fl	79	81	78	81	80.0	± 5	81	81	80	81	81.0	± 5	88	88	86	88	88.0	± 5	
		pg	27.6	28.3	27.8	28.3	27.9	± 2.0	29.0	28.9	29.2	28.9	29.2	± 2.0	31.9	31.8	31.9	31.8	31.7	± 2.5	
CCMH	MCHC	fmol	1.72	1.75	1.72	1.75	1.73	± 0.12	1.80	1.80	1.81	1.80	1.82	± 0.12	1.98	1.97	1.98	1.97	1.97	± 0.16	
		g/dl	35.0	34.9	35.6	34.9	34.8	± 3.0	35.9	35.7	36.5	35.7	36.1	± 3.0	36.3	36.1	37.1	36.1	36.0	± 3.0	
		g/l	350	349	356	349	348	± 30	359	357	365	357	361	± 30	363	361	371	361	360	± 30	
IDR	RDW	mmol/l	21.72	21.66	22.10	21.66	21.64	± 1.86	22.26	22.17	22.64	22.17	22.41	± 1.86	22.54	22.41	23.02	22.41	22.37	± 1.86	
PLAQ.	PLTS	%	13.7	13.0	12.5	13.0	12.8	± 4.0	12.8	13.0	12.8	13.0	12.8	± 4.0	12.2	12.4	11.6	12.4	12.0	± 4.0	
VPM	MPV	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68	70	68	68	± 20	250	248	253	248	245	± 30	480	480	495	480	475	± 50	
NEUT		µm ³ ; fl	10.1	10.2	10.2	10.2	9.4	± 2.0	9.4	9.9	9.6	9.9	9.0	± 2.0	9.3	9.6	9.5	9.6	8.8	± 2.0	
		#	1.38	1.44	1.37	1.44	1.48	± 0.35	4.33	4.36	4.22	4.36	4.50	± 0.90	12.09	12.25	12.29	12.25	12.74	± 1.90	
LYMPHO		%	57.3	57.5	57.0	57.5	59.2	± 10.0	58.5	58.9	57.8	58.9	60.0	± 10.0	68.3	70.0	70.2	70.0	70.4	± 10.0	
		#	0.67	0.72	0.70	0.72	0.71	± 0.33	1.96	2.02	2.03	2.02	2.06	± 0.70	2.39	2.26	2.31	2.26	2.52	± 1.50	
MONO		%	28.0	28.6	29.0	28.6	28.4	± 12.0	26.5	27.3	27.8	27.3	27.4	± 8.0	13.5	12.9	13.2	12.9	13.9	± 8.0	
		#	0.13	0.11	0.11	0.11	0.10	± 0.10	0.56	0.47	0.51	0.47	0.42	± 0.42	1.24	0.96	1.09	0.96	0.92	± 0.92	
EOS		%	5.3	4.3	4.5	4.3	3.8	± 3.8	7.5	6.3	7.0	6.3	5.6	± 5.6	7.0	5.5	6.2	5.5	5.1	± 5.1	
		#	0.14	0.16	0.14	0.16	0.13	± 0.13	0.28	0.30	0.28	0.30	0.26	± 0.26	1.20	1.26	1.05	1.26	1.14	± 1.05	
BASO		%	6.0	6.3	6.0	6.3	5.1	± 5.1	3.8	4.0	3.8	4.0	3.5	± 3.5	6.8	7.2	6.0	7.2	6.3	± 6.0	
		#	0.08	0.08	0.08	0.08	0.09	± 0.08	0.27	0.26	0.26	0.26	0.26	± 0.26	0.78	0.77	0.77	0.77	0.78	± 0.77	
		%	3.4	3.3	3.5	3.3	3.5	± 3.3	3.7	3.5	3.6	3.5	3.5	± 3.5	4.4	4.4	4.4	4.4	4.3	± 4.3	

ABX Difftrol



LOT PX 036
Rev 4

CONTROL

(Exp.) 2016-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA							
			120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS				
RETIC	DF120	DF NEXUS	DF NEXUS	RETIC	DF120	DF NEXUS		DF NEXUS	RETIC	DF120	DF NEXUS		DF NEXUS							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	7.6	7.6	7.6			± 1.0	18.5	18.5	18.5			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.30	2.30			± 0.16	4.60	4.60	4.60			± 0.20	5.15	5.15	5.15			± 0.25
HB	HGB	g/dl	6.5	6.5	6.5			± 0.4	13.3	13.3	13.3			± 0.5	16.1	16.1	16.1			± 0.6
		g/l	65	65	65			± 4	133	133	133			± 5	161	161	161			± 6
		mmol/l	4.04	4.04	4.04			± 0.25	8.26	8.26	8.26			± 0.31	10.00	10.00	10.00			± 0.37
HT	HCT	%	18.9	18.9	18.9			± 1.5	38.2	38.2	38.2			± 2.0	45.8	45.8	45.8			± 2.5
		l/l	0.189	0.189	0.189			± 0.015	0.382	0.382	0.382			± 0.020	0.458	0.458	0.458			± 0.025
VGM	MCV	µm ³ ; fl	82	82	82			± 5	83	83	83			± 5	89	89	89			± 5
TGMH	MCH	pg	28.3	28.3	28.3			± 2.0	28.9	28.9	28.9			± 2.0	31.3	31.3	31.3			± 2.5
		fmol	1.76	1.76	1.76			± 0.12	1.80	1.80	1.80			± 0.12	1.94	1.94	1.94			± 0.16
CCMH	MCHC	g/dl	34.5	34.5	34.5			± 3.0	34.8	34.8	34.8			± 3.0	35.1	35.1	35.1			± 3.0
		g/l	345	345	345			± 30	348	348	348			± 30	351	351	351			± 30
		mmol/l	21.40	21.40	21.40			± 1.86	21.63	21.63	21.63			± 1.86	21.81	21.81	21.81			± 1.86
IDR	RDW	%	15.8	15.8	15.8			± 4.0	15.6	15.6	15.6			± 4.0	14.0	14.0	14.0			± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	72			± 20	255	255	255			± 30	485	485	485			± 50
VPM	MPV	µm ³ ; fl	10.0	10.0	10.0			± 2.0	9.3	9.3	9.3			± 2.0	9.0	9.0	9.0			± 2.0
NEUT		#	1.45	1.50	1.50			± 0.35	4.44	4.55	4.55			± 0.90	13.41	13.63	13.63			± 1.90
		%	58.1	60.1	60.1			± 10.0	58.4	59.9	59.9			± 10.0	72.5	73.7	73.7			± 10.0
LYMPHO		#	0.69	0.63	0.63			± 0.33	2.00	1.86	1.86			± 0.70	2.24	2.07	2.07			± 1.50
		%	27.6	25.1	25.1			± 12.0	26.3	24.5	24.5			± 8.0	12.1	11.2	11.2			± 8.0
MONO		#	0.16	0.16	0.16			± 0.16	0.70	0.72	0.72			± 0.70	1.67	1.65	1.65			± 1.65
		%	6.3	6.4	6.4			± 6.3	9.2	9.5	9.5			± 9.2	9.0	8.9	8.9			± 8.9
EOS		#	0.14	0.15	0.15			± 0.14	0.22	0.24	0.24			± 0.22	0.67	0.68	0.68			± 0.67
		%	5.5	5.9	5.9			± 5.5	2.9	3.1	3.1			± 2.9	3.6	3.7	3.7			± 3.6
BASO		#	0.06	0.06	0.06			± 0.06	0.24	0.23	0.23			± 0.23	0.52	0.46	0.46			± 0.46
		%	2.5	2.5	2.5			± 2.5	3.2	3.0	3.0			± 3.0	2.8	2.5	2.5			± 2.5

Ref: TEMP-0821 Rev.37 FRONT / RECTO 1300019740-A