


LOT PX 408

CONTROL

 (Exp.) 2018-01-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
			CONTROL				L	CONTROL				N	CONTROL				H		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7	2.7		± 0.4	8.3	8.3	8.3		± 1.0	19.9	19.9	19.9		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.29	2.29		± 0.16	4.47	4.47	4.47		± 0.20	5.12	5.12	5.12		± 0.25			
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	12.7	12.7	12.7		± 0.5	15.7	15.7	15.7		± 0.6			
	g/l	66	66	66		± 4	127	127	127		± 5	157	157	157		± 6			
	mmol/l	4.10	4.10	4.10		± 0.25	7.89	7.89	7.89		± 0.31	9.75	9.75	9.75		± 0.37			
HT HCT	%	19.0	19.0	19.2		± 1.5	36.2	36.2	36.7		± 2.0	44.5	44.5	45.1		± 2.5			
	l/l	0.190	0.190	0.192		± 0.015	0.362	0.362	0.367		± 0.020	0.445	0.445	0.451		± 0.025			
VGM MCV	µm ³ ; fl	83	83	84		± 5	81	81	82		± 5	87	87	88		± 5			
TGMH MCH	pg	28.8	28.8	28.8		± 2.0	28.4	28.4	28.4		± 2.0	30.7	30.7	30.7		± 2.5			
	fmol	1.79	1.79	1.79		± 0.12	1.76	1.76	1.76		± 0.12	1.90	1.90	1.90		± 0.16			
CCMH MCHC	g/dl	34.7	34.7	34.3		± 3.0	35.1	35.1	34.6		± 3.0	35.2	35.2	34.8		± 3.0			
	g/l	347	347	343		± 30	351	351	346		± 30	352	352	348		± 30			
	mmol/l	21.56	21.56	21.31		± 1.86	21.78	21.78	21.52		± 1.86	21.89	21.89	21.64		± 1.86			
IDR RDW	%	15.8	15.8	15.8		± 4.0	16.2	16.2	16.2		± 4.0	15.4	15.4	15.4		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78	78	78		± 20	250	250	250		± 30	490	490	490		± 50			
VPM MPV	µm ³ ; fl	10.1	10.1	10.1		± 2.0	9.8	9.8	9.8		± 2.0	9.0	9.0	9.0		± 2.0			
NEUT	#	1.50	1.56	1.56		± 0.35	4.66	4.80	4.80		± 0.90	14.10	14.30	14.30		± 1.90			
	%	55.5	57.8	57.8		± 10.0	56.2	57.8	57.8		± 10.0	70.6	71.7	71.7		± 10.0			
LYMPHO	#	0.81	0.73	0.73		± 0.33	2.58	2.41	2.41		± 0.70	2.87	2.65	2.65		± 1.50			
	%	29.9	27.2	27.2		± 12.0	31.1	29.0	29.0		± 8.0	14.4	13.3	13.3		± 8.0			
MONO	#	0.18	0.19	0.19		± 0.18	0.54	0.56	0.56		± 0.54	1.51	1.49	1.49		± 1.49			
	%	6.8	6.9	6.9		± 6.8	6.5	6.7	6.7		± 6.5	7.6	7.5	7.5		± 7.5			
EOS	#	0.14	0.15	0.15		± 0.14	0.27	0.29	0.29		± 0.27	0.98	1.00	1.00		± 0.98			
	%	5.3	5.6	5.6		± 5.3	3.2	3.5	3.5		± 3.2	4.9	5.0	5.0		± 4.9			
BASO	#	0.07	0.07	0.07		± 0.07	0.25	0.25	0.25		± 0.25	0.50	0.50	0.50		± 0.50			
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5			

Ref: TEMP-0821 Rev.41 BACK / VERSO 1300040596