

ABX Difftrol



LOT PX 409
Rev 1

CONTROL

(Exp.) 2018-03-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6	± 0.4	7.7	7.7	7.7	± 1.0	19.0	19.0	19.0	± 2.2					
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.33	2.33	2.33	± 0.16	4.58	4.58	4.58	± 0.20	5.18	5.18	5.18	± 0.25					
HB HGB	g/dl	6.8	6.8	6.8	± 0.4	13.4	13.4	13.4	± 0.5	16.4	16.4	16.4	± 0.6					
	g/l	68	68	68	± 4	134	134	134	± 5	164	164	164	± 6					
HT HCT	mmol/l	4.22	4.22	4.22	± 0.25	8.32	8.32	8.32	± 0.31	10.18	10.18	10.18	± 0.37					
	%	19.1	19.1	19.3	± 1.5	37.1	37.1	37.6	± 2.0	45.1	45.1	45.6	± 2.5					
VGM MCV	l/l	0.191	0.191	0.193	± 0.015	0.371	0.371	0.376	± 0.020	0.451	0.451	0.456	± 0.025					
	µm ³ /fl	82	82	83	± 5	81	81	82	± 5	87	87	88	± 5					
TGMH MCH	pg	29.2	29.2	29.2	± 2.0	29.3	29.3	29.3	± 2.0	31.7	31.7	31.7	± 2.5					
	fmol	1.81	1.81	1.81	± 0.12	1.82	1.82	1.82	± 0.12	1.97	1.97	1.97	± 0.16					
CCMH MCHC	g/dl	35.6	35.6	35.2	± 3.0	36.1	36.1	35.7	± 3.0	36.4	36.4	36.0	± 3.0					
	g/l	356	356	352	± 30	361	361	357	± 30	364	364	360	± 30					
IDR RDW	mmol/l	22.10	22.10	21.84	± 1.86	22.43	22.43	22.16	± 1.86	22.60	22.60	22.34	± 1.86					
	%	15.7	15.7	15.7	± 4.0	16.5	16.5	16.5	± 4.0	14.8	14.8	14.8	± 4.0					
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70	± 20	250	250	250	± 30	492	492	492	± 50					
VPM MPV	µm ³ /fl	9.4	9.4	9.4	± 2.0	9.3	9.3	9.3	± 2.0	8.6	8.6	8.6	± 2.0					
NEUT	#	1.47	1.53	1.53	± 0.35	4.32	4.44	4.44	± 0.90	13.40	13.60	13.60	± 1.90					
	%	56.7	58.9	58.9	± 10.0	56.1	57.6	57.6	± 10.0	70.1	71.4	71.4	± 10.0					
LYMPHO	#	0.75	0.68	0.68	± 0.33	2.33	2.18	2.18	± 0.70	2.98	2.76	2.76	± 1.50					
	%	28.7	26.1	26.1	± 12.0	30.3	28.3	28.3	± 8.0	15.7	14.5	14.5	± 8.0					
MONO	#	0.20	0.20	0.20	± 0.20	0.60	0.62	0.62	± 0.60	1.41	1.39	1.39	± 1.39					
	%	7.6	7.8	7.8	± 7.6	7.8	8.1	8.1	± 7.8	7.4	7.3	7.3	± 7.3					
EOS	#	0.12	0.12	0.12	± 0.12	0.22	0.23	0.23	± 0.22	0.82	0.82	0.82	± 0.82					
	%	4.5	4.7	4.7	± 4.5	2.8	3.0	3.0	± 2.8	4.3	4.3	4.3	± 4.3					
BASO	#	0.07	0.07	0.07	± 0.07	0.23	0.23	0.23	± 0.23	0.48	0.48	0.48	± 0.48					
	%	2.5	2.5	2.5	± 2.5	3.0	3.0	3.0	± 3.0	2.5	2.5	2.5	± 2.5					

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596