


LOT PX 419

CONTROL

 (Exp.) 2019-11-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	$10^9/mm^3; 10^9/l$	2.5	2.5	2.5		± 0.4	7.9	7.9	7.9		± 1.0	18.9	18.9	18.9		± 2.2		
GR RBC	$10^9/mm^3; 10^{12}/l$	2.38	2.38	2.38		± 0.16	4.61	4.61	4.61		± 0.20	5.23	5.23	5.23		± 0.25		
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.8	13.8	13.8		± 0.5	16.8	16.8	16.8		± 0.6		
	g/l	69	69	69		± 4	138	138	138		± 5	168	168	168		± 6		
	mmol/l	4.28	4.28	4.28		± 0.25	8.57	8.57	8.57		± 0.31	10.43	10.43	10.43		± 0.37		
HT HCT	%	19.8	19.8	19.8		± 1.5	38.7	38.7	38.7		± 2.0	47.1	47.1	47.1		± 2.5		
	l/l	0.198	0.198	0.198		± 0.015	0.387	0.387	0.387		± 0.020	0.471	0.471	0.471		± 0.025		
VGM MCV	$\mu m^3; fl$	83	83	83		± 5	84	84	84		± 5	90	90	90		± 5		
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.9	29.9	29.9		± 2.0	32.1	32.1	32.1		± 2.5		
	fmol	1.80	1.80	1.80		± 0.12	1.86	1.86	1.86		± 0.12	1.99	1.99	1.99		± 0.16		
CCMH MCHC	g/dl	34.9	34.9	34.9		± 3.0	35.6	35.6	35.6		± 3.0	35.7	35.7	35.7		± 3.0		
	g/l	349	349	349		± 30	356	356	356		± 30	357	357	357		± 30		
	mmol/l	21.69	21.69	21.69		± 1.86	22.13	22.13	22.13		± 1.86	22.16	22.16	22.16		± 1.86		
IDR RDW	%	15.0	15.0	15.0		± 4.0	15.5	15.5	15.5		± 4.0	15.5	15.5	15.5		± 4.0		
PLAQ. PLTS	$10^3/mm^3; 10^9/l$	72	72	72		± 20	257	257	257		± 30	500	500	500		± 50		
VPM MPV	$\mu m^3; fl$	10.1	10.1	10.1		± 2.0	9.6	9.6	9.6		± 2.0	10.1	10.1	10.1		± 2.0		
NEUT	#	1.47	1.52	1.52		± 0.35	4.54	4.68	4.68		± 0.90	13.60	13.80	13.80		± 1.90		
	%	58.8	60.7	60.7		± 10.0	57.5	59.2	59.2		± 10.0	72.0	73.1	73.1		± 10.0		
LYMPHO	#	0.68	0.62	0.62		± 0.33	2.44	2.28	2.28		± 0.70	2.74	2.55	2.55		± 1.50		
	%	27.3	24.8	24.8		± 12.0	30.9	28.9	28.9		± 8.0	14.5	13.5	13.5		± 8.0		
MONO	#	0.12	0.12	0.12		± 0.12	0.36	0.37	0.37		± 0.36	1.02	1.00	1.00		± 1.00		
	%	4.7	4.8	4.8		± 4.7	4.6	4.7	4.7		± 4.6	5.4	5.3	5.3		± 5.3		
EOS	#	0.16	0.17	0.17		± 0.16	0.24	0.25	0.25		± 0.24	0.81	0.81	0.81		± 0.81		
	%	6.3	6.7	6.7		± 6.3	3.0	3.2	3.2		± 3.0	4.3	4.3	4.3		± 4.3		
BASO	#	0.07	0.08	0.08		± 0.07	0.32	0.32	0.32		± 0.32	0.72	0.72	0.72		± 0.72		
	%	2.9	3.0	3.0		± 2.9	4.0	4.0	4.0		± 4.0	3.8	3.8	3.8		± 3.8		

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596