


**LOT** PX 410

**CONTROL**

 (Exp.) 2018-05-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
			CONTROL				L	CONTROL				N	CONTROL				H		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.9	7.9	7.9		± 1.0	19.3	19.3	19.3		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.33	2.33	2.33		± 0.16	4.68	4.68	4.68		± 0.20	5.10	5.10	5.10		± 0.25			
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	14.1	14.1	14.1		± 0.5	15.8	15.8	15.8		± 0.6			
	g/l	67	67	67		± 4	141	141	141		± 5	158	158	158		± 6			
	mmol/l	4.16	4.16	4.16		± 0.25	8.76	8.76	8.76		± 0.31	9.81	9.81	9.81		± 0.37			
HT HCT	%	19.3	19.3	19.6		± 1.5	39.3	39.3	39.8		± 2.0	44.4	44.4	44.9		± 2.5			
	l/l	0.193	0.193	0.196		± 0.015	0.393	0.393	0.398		± 0.020	0.444	0.444	0.449		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	83	84		± 5	84	84	85		± 5	87	87	88		± 5			
TGMH MCH	pg	28.8	28.8	28.8		± 2.0	30.1	30.1	30.1		± 2.0	31.0	31.0	31.0		± 2.5			
	fmol	1.79	1.79	1.79		± 0.12	1.87	1.87	1.87		± 0.12	1.92	1.92	1.92		± 0.16			
CCMH MCHC	g/dl	34.6	34.6	34.2		± 3.0	35.9	35.9	35.4		± 3.0	35.6	35.6	35.2		± 3.0			
	g/l	346	346	342		± 30	359	359	354		± 30	356	356	352		± 30			
	mmol/l	21.51	21.51	21.26		± 1.86	22.27	22.27	22.01		± 1.86	22.11	22.11	21.86		± 1.86			
IDR RDW	%	15.1	15.1	15.1		± 4.0	15.8	15.8	15.8		± 4.0	15.0	15.0	15.0		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	252	252	252		± 30	480	480	480		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.6	9.6	9.6		± 2.0	9.6	9.6	9.6		± 2.0	9.1	9.1	9.1		± 2.0			
NEUT	#	1.39	1.45	1.45		± 0.35	4.45	4.58	4.58		± 0.90	13.80	14.00	14.00		± 1.90			
	%	55.7	57.8	57.8		± 10.0	56.3	58.0	58.0		± 10.0	71.4	72.6	72.6		± 10.0			
LYMPHO	#	0.75	0.68	0.68		± 0.33	2.50	2.33	2.33		± 0.70	3.07	2.84	2.84		± 1.50			
	%	29.9	27.2	27.2		± 12.0	31.6	29.5	29.5		± 8.0	15.9	14.7	14.7		± 8.0			
MONO	#	0.12	0.12	0.12		± 0.12	0.47	0.49	0.49		± 0.47	1.25	1.25	1.25		± 1.25			
	%	4.8	4.9	4.9		± 4.8	6.0	6.2	6.2		± 6.0	6.5	6.5	6.5		± 6.5			
EOS	#	0.18	0.19	0.19		± 0.18	0.24	0.26	0.26		± 0.24	0.71	0.71	0.71		± 0.71			
	%	7.1	7.6	7.6		± 7.1	3.1	3.3	3.3		± 3.1	3.7	3.7	3.7		± 3.7			
BASO	#	0.06	0.06	0.06		± 0.06	0.24	0.24	0.24		± 0.24	0.48	0.48	0.48		± 0.48			
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5			

Ref. TEMP-0821 Rev.42 BACK / VERSO 1300040596