

LOT		PX 033		CONTROL		(Exp.) 2013-05-05										
		Rev 2				(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse														
		CONTROL				L	CONTROL				N	CONTROL				H
		PENTRA					PENTRA					PENTRA				
60 60C+ ES60	80 XL80	TOLERANCES TOLERANCE		60 60C+ ES60	80 XL80	TOLERANCES TOLERANCE		60 60C+ ES60	80 XL80	TOLERANCES TOLERANCE						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	± 0.4	7.4	7.4	± 1.0	17.7	17.5	± 2.2					
GR	RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.38	± 0.12	4.62	4.60	± 0.15	5.07	5.09	± 0.20					
HB	HGB	g/dl	6.7	6.7	± 0.4	13.6	13.6	± 0.5	15.8	15.8	± 0.6					
		g/l	67	67	± 4	136	136	± 5	158	158	± 6					
		mmol/l	4.16	4.16	± 0.25	8.45	8.45	± 0.31	9.81	9.81	± 0.37					
HT	HCT	%	19.4	19.3	± 1.5	37.9	38.2	± 2.0	44.6	44.8	± 2.5					
		l/l	0.194	0.193	± 0.015	0.379	0.382	± 0.020	0.446	0.448	± 0.025					
VGM	MCV	µm <sup>3</sup> ; fl	79	81	± 5	82	83	± 5	88	88	± 5					
TGMH	MCH	pg	27.3	28.2	± 2.0	29.4	29.6	± 2.0	31.2	31.0	± 2.5					
		fmol	1.70	1.75	± 0.12	1.83	1.84	± 0.12	1.94	1.93	± 0.16					
		g/dl	34.6	34.8	± 3.0	35.9	35.6	± 3.0	35.4	35.3	± 3.0					
CCMH	MCHC	g/l	346	348	± 30	359	356	± 30	354	353	± 30					
		mmol/l	21.50	21.58	± 1.86	22.29	22.12	± 1.86	21.99	21.91	± 1.86					
		%	13.5	13.2	± 4.0	13.5	13.8	± 4.0	13.0	13.7	± 4.0					
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	± 20	255	255	± 30	505	505	± 50					
VPM	MPV	µm <sup>3</sup> ; fl	8.9	8.9	± 2.0	9.1	9.2	± 2.0	8.6	8.8	± 2.0					
NEUT	#		1.37	1.36	± 0.30	4.04	4.03	± 0.80	12.16	11.73	± 1.80					
		%	57.0	56.5	± 10.0	54.6	54.5	± 10.0	68.7	67.0	± 10.0					
LYMPHO	#		0.71	0.72	± 0.20	2.44	2.44	± 0.60	3.10	3.41	± 1.50					
		%	29.5	30.0	± 8.0	33.0	33.0	± 8.0	17.5	19.5	± 8.0					
MONO	#		0.12	0.12	± 0.12	0.39	0.39	± 0.39	0.80	0.79	± 0.75					
		%	5.0	5.0	± 5.0	5.0	5.0	± 5.0	4.5	4.5	± 4.0					
EOS	#		0.13	0.13	± 0.13	0.31	0.33	± 0.31	0.89	0.88	± 0.74					
		%	5.5	5.5	± 5.0	4.0	4.5	± 4.0	5.0	5.0	± 4.0					
BASO	#		0.07	0.07	± 0.06	0.25	0.23	± 0.23	0.76	0.70	± 0.46					
		%	3.0	3.0	± 2.5	3.4	3.0	± 3.0	4.3	4.0	± 2.5					

LOT		PX 033		CONTROL		(Exp.) 2013-05-05										
		Rev 2				(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse														
		CONTROL				L	CONTROL				N	CONTROL				H
		PENTRA					PENTRA					PENTRA				
120 120 RETIC	DX120 DF120	TOLERANCES TOLERANCE		120 120 RETIC	DX120 DF120	TOLERANCES TOLERANCE		120 120 RETIC	DX120 DF120	TOLERANCES TOLERANCE						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	± 0.4	7.7	7.7	± 1.0	18.5	18.5	± 2.2					
GR	RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.44	± 0.12	4.64	4.64	± 0.15	5.17	5.17	± 0.20					
HB	HGB	g/dl	6.8	6.8	± 0.4	13.7	13.7	± 0.5	15.9	15.9	± 0.6					
		g/l	68	68	± 4	137	137	± 5	159	159	± 6					
		mmol/l	4.22	4.22	± 0.25	8.51	8.51	± 0.31	9.87	9.87	± 0.37					
HT	HCT	%	19.8	19.8	± 1.5	39.0	39.0	± 2.0	45.5	45.5	± 2.5					
		l/l	0.198	0.198	± 0.015	0.390	0.390	± 0.020	0.455	0.455	± 0.025					
VGM	MCV	µm <sup>3</sup> ; fl	81	81	± 5	84	84	± 5	88	88	± 5					
TGMH	MCH	pg	27.9	27.9	± 2.0	29.5	29.5	± 2.0	30.8	30.8	± 2.5					
		fmol	1.73	1.73	± 0.12	1.83	1.83	± 0.12	1.91	1.91	± 0.16					
		g/dl	34.4	34.4	± 3.0	35.1	35.1	± 3.0	34.9	34.9	± 3.0					
CCMH	MCHC	g/l	344	344	± 30	351	351	± 30	349	349	± 30					
		mmol/l	21.37	21.37	± 1.86	21.83	21.83	± 1.86	21.70	21.70	± 1.86					
		%	15.8	15.8	± 4.0	16.0	16.0	± 4.0	14.5	14.5	± 4.0					
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	± 20	260	260	± 30	505	505	± 50					
VPM	MPV	µm <sup>3</sup> ; fl	8.9	8.9	± 2.0	9.1	9.1	± 2.0	8.6	8.6	± 2.0					
NEUT	#		1.37	1.39	± 0.30	4.24	4.39	± 0.80	12.80	12.80	± 1.80					
		%	57.0	58.0	± 10.0	55.0	57.0	± 10.0	69.0	69.0	± 10.0					
LYMPHO	#		0.73	0.65	± 0.20	2.54	2.27	± 0.60	3.42	3.15	± 1.50					
		%	30.5	27.0	± 8.0	33.0	29.5	± 8.0	18.5	17.0	± 8.0					
MONO	#		0.12	0.16	± 0.12	0.39	0.46	± 0.39	1.11	1.30	± 0.75					
		%	5.0	6.5	± 5.0	5.0	6.0	± 5.0	6.0	7.0	± 4.0					
EOS	#		0.13	0.13	± 0.13	0.31	0.31	± 0.31	0.74	0.74	± 0.74					
		%	5.0	5.5	± 5.0	4.0	4.0	± 4.0	4.0	4.0	± 4.0					
BASO	#		0.06	0.07	± 0.06	0.23	0.27	± 0.23	0.46	0.56	± 0.46					
		%	2.5	3.0	± 2.5	3.0	3.5	± 3.0	2.5	3.0	± 2.5					

Ref: TEMP-0821 Rev.30 FRONT / RECTO 9930073-A