


LOT
PX 053
CONTROL

(Exp.) 2013-07-05
 (YYYY-MM-DD)

Rev 3

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																				
			CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE			
			PENTRA						PENTRA						PENTRA								
			60 60C+ ES60	80 XL80						60 60C+ ES60	80 XL80						60 60C+ ES60	80 XL80					
GB	WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.4	2.5				± 0.4	7.4	7.4				± 1.0	17.2	17.2			± 2.2				
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.44				± 0.12	4.56	4.57				± 0.15	5.02	5.05			± 0.20				
HB	HGB	g/dl	7.0	7.0				± 0.4	13.7	13.8				± 0.5	15.9	15.9			± 0.6				
		g/l	70	70				± 4	137	138				± 5	159	159			± 6				
		mmol/l	4.35	4.35				± 0.25	8.51	8.57				± 0.31	9.87	9.87			± 0.37				
HT	HCT	%	19.8	20.3				± 1.5	37.8	38.4				± 2.0	43.7	43.9			± 2.5				
		l/l	0.198	0.203				± 0.015	0.378	0.384				± 0.020	0.437	0.439			± 0.025				
VGM	MCV	µm ³ ; fl	80	83				± 5	83	84				± 5	87	87			± 5				
TGMH	MCH	pg	28.2	28.7				± 2.0	30.0	30.2				± 2.0	31.7	31.5			± 2.5				
		fmol	1.75	1.78				± 0.12	1.87	1.88				± 0.12	1.97	1.96			± 0.16				
CCMH	MCHC	g/dl	35.3	34.6				± 3.0	36.2	35.9				± 3.0	36.4	36.2			± 3.0				
		g/l	353	346				± 30	362	359				± 30	364	362			± 30				
		mmol/l	21.91	21.46				± 1.86	22.48	22.32				± 1.86	22.61	22.47			± 1.86				
IDR	RDW	%	13.2	13.6				± 4.0	12.5	13.6				± 4.0	12.3	13.2			± 4.0				
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	70				± 20	258	260				± 30	505	515			± 50				
VPM	MPV	µm ³ ; fl	9.2	9.3				± 2.0	8.9	9.1				± 2.0	8.8	9.0			± 2.0				
NEUT		#	1.49	1.60				± 0.30	4.37	4.40				± 0.80	11.87	11.78			± 1.90				
		%	62.0	64.0				± 10.0	59.0	59.5				± 10.0	69.0	68.5			± 10.0				
LYMPHO		#	0.56	0.55				± 0.30	2.18	2.11				± 0.70	2.92	2.92			± 1.50				
		%	23.5	22.0				± 9.0	29.5	28.5				± 8.0	17.0	17.0			± 8.0				
MONO		#	0.13	0.11				± 0.11	0.33	0.33				± 0.33	0.77	0.77			± 0.77				
		%	5.5	4.5				± 4.5	4.5	4.5				± 4.5	4.5	4.5			± 4.5				
EOS		#	0.13	0.15				± 0.13	0.27	0.30				± 0.27	0.86	0.95			± 0.82				
		%	5.5	6.0				± 5.0	3.5	4.0				± 3.5	5.0	5.5			± 4.5				
BASO		#	0.08	0.09				± 0.05	0.26	0.26				± 0.16	0.77	0.77			± 0.36				
		%	3.5	3.5				± 2.0	3.5	3.5				± 2.0	4.5	4.5			± 2.0				

LOT PX 053

CONTROL

 (Exp.) **2013-07-05**
 (YYYY - MM - DD)

Rev 3

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse																				
		CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA	PENTRA									
		120 120 RETIC	DX120 DF120				120 120 RETIC		DX120 DF120					120 120 RETIC		DX120 DF120						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5				± 0.4	7.8	7.8					± 1.0	18.2	18.2				± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.48				± 0.12	4.60	4.60					± 0.15	5.14	5.14				± 0.20		
HB HGB	g/dl	7.1	7.1				± 0.4	13.8	13.8					± 0.5	15.9	15.9				± 0.6		
	g/l	71	71				± 4	138	138					± 5	159	159				± 6		
	mmol/l	4.41	4.41				± 0.25	8.57	8.57					± 0.31	9.87	9.87				± 0.37		
HT HCT	%	20.3	20.3				± 1.5	38.6	38.6					± 2.0	44.7	44.7				± 2.5		
	l/l	0.203	0.203				± 0.015	0.386	0.386					± 0.020	0.447	0.447				± 0.025		
VGM MCV	µm ³ ; fl	82	82				± 5	84	84					± 5	87	87				± 5		
TGMH MCH	pg	28.6	28.6				± 2.0	30.0	30.0					± 2.0	30.9	30.9				± 2.5		
	fmol	1.78	1.78				± 0.12	1.86	1.86					± 0.12	1.92	1.92				± 0.16		
CCMH MCHC	g/dl	34.9	34.9				± 3.0	35.7	35.7					± 3.0	35.6	35.6				± 3.0		
	g/l	349	349				± 30	357	357					± 30	356	356				± 30		
	mmol/l	21.68	21.68				± 1.86	22.18	22.18					± 1.86	22.08	22.08				± 1.86		
IDR RDW	%	16.0	16.0				± 4.0	16.0	16.0					± 4.0	15.2	15.2				± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74				± 20	258	258					± 30	508	508				± 50		
VPM MPV	µm ³ ; fl	9.1	9.1				± 2.0	9.0	9.0					± 2.0	8.9	8.9				± 2.0		
NEUT	#	1.60	1.55				± 0.30	4.60	4.60					± 0.80	12.69	12.65				± 1.90		
	%	64.0	62.0				± 10.0	59.0	59.0					± 10.0	69.7	69.5				± 10.0		
LYMPHO	#	0.60	0.56				± 0.30	2.34	2.18					± 0.70	3.24	3.00				± 1.50		
	%	24.0	22.5				± 9.0	30.0	28.0					± 8.0	17.8	16.5				± 8.0		
MONO	#	0.11	0.19				± 0.11	0.39	0.51					± 0.33	1.00	1.18				± 0.77		
	%	4.5	7.5				± 4.5	5.0	6.5					± 4.5	5.5	6.5				± 4.5		
EOS	#	0.13	0.13				± 0.13	0.27	0.27					± 0.27	0.82	0.82				± 0.82		
	%	5.0	5.0				± 5.0	3.5	3.5					± 3.5	4.5	4.5				± 4.5		
BASO	#	0.06	0.08				± 0.05	0.20	0.23					± 0.16	0.46	0.55				± 0.36		
	%	2.5	3.0				± 2.0	2.5	3.0					± 2.0	2.5	3.0				± 2.0		

Ref: TEMP-0821 Rev.32 FRONT / RECTO 9930080-A