



LOT PX 014

CONTROL

(Exp.) 2014-03-05
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																	
				CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
				PENTRA				PENTRA				PENTRA									
				60 60C+ ES60	80 XL80			60 60C+ ES60		80 XL80			60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3				± 0.4	7.2	7.2				± 1.0	17.6	17.2				± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.30				± 0.12	4.38	4.35				± 0.15	5.20	5.20				± 0.20	
HB	HGB	g/dl	6.6	6.6				± 0.4	13.1	13.1				± 0.5	16.7	16.6				± 0.6	
		g/l	66	66				± 4	131	131				± 5	167	166				± 6	
		mmol/l	4.10	4.10				± 0.25	8.14	8.14				± 0.31	10.37	10.31				± 0.37	
HT	HCT	%	18.6	18.9				± 1.5	35.9	35.7				± 2.0	45.8	45.8				± 2.5	
		l/l	0.186	0.189				± 0.015	0.359	0.357				± 0.020	0.458	0.458				± 0.025	
VGM	MCV	µm ³ ; fl	79	82				± 5	82	82				± 5	88	88				± 5	
TGMH	MCH	pg	28.1	28.7				± 2.0	29.9	30.1				± 2.0	32.1	31.9				± 2.5	
		fmol	1.74	1.78				± 0.12	1.86	1.87				± 0.12	1.99	1.98				± 0.16	
CCMH	MCHC	g/dl	35.6	35.0				± 3.0	36.5	36.7				± 3.0	36.5	36.3				± 3.0	
		g/l	356	350				± 30	365	367				± 30	365	363				± 30	
		mmol/l	22.08	21.73				± 1.86	22.65	22.81				± 1.86	22.66	22.53				± 1.86	
IDR	RDW	%	12.7	13.3				± 4.0	12.0	13.0				± 4.0	11.6	13.0				± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	72				± 20	258	260				± 30	510	520				± 50	
VPM	MPV	µm ³ ; fl	9.2	9.4				± 2.0	8.6	8.8				± 2.0	8.9	9.1				± 2.0	
NEUT	#		1.27	1.27				± 0.25	4.14	4.18				± 0.80	12.32	12.04				± 1.90	
	%		55.0	55.3				± 10.0	57.5	58.0				± 10.0	70.0	70.0				± 10.0	
LYMPHO	#		0.70	0.70				± 0.20	2.23	2.20				± 0.60	2.90	2.84				± 1.50	
	%		30.5	30.5				± 8.0	31.0	30.5				± 8.0	16.5	16.5				± 8.0	
MONO	#		0.12	0.10				± 0.10	0.32	0.29				± 0.29	0.79	0.69				± 0.69	
	%		5.0	4.5				± 4.5	4.5	4.0				± 4.0	4.5	4.0				± 4.0	
EOS	#		0.15	0.15				± 0.12	0.25	0.29				± 0.21	0.88	0.86				± 0.73	
	%		6.5	6.5				± 5.0	3.5	4.0				± 2.8	5.0	5.0				± 4.0	
BASO	#		0.07	0.07				± 0.07	0.25	0.25				± 0.19	0.70	0.77				± 0.46	
	%		3.0	3.2				± 3.0	3.5	3.5				± 2.5	4.0	4.5				± 2.5	



HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

ABX Difftrol



LOT PX 014

CONTROL



(Exp.) 2014-03-05
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																	
				CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
				PENTRA				PENTRA				PENTRA	PENTRA								
				120 120 RETIC	DX120 DF120			120 120 RETIC		DX120 DF120				120 120 RETIC		DX120 DF120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20		
HB HGB	g/dl						± 0.4						± 0.5						± 0.6		
	g/l						± 4						± 5					± 6			
	mmol/l						± 0.25						± 0.31					± 0.37			
HT HCT	%						± 1.5						± 2.0					± 2.5			
	l/l						± 0.015						± 0.020					± 0.025			
VGM MCV	µm ³ ; fl						± 5						± 5					± 5			
TGMH MCH	pg						± 2.0						± 2.0					± 2.5			
	fmol						± 0.12						± 0.12					± 0.16			
CCMH MCHC	g/dl						± 3.0						± 3.0					± 3.0			
	g/l						± 30						± 30					± 30			
	mmol/l						± 1.86						± 1.86					± 1.86			
IDR RDW	%						± 4.0						± 4.0					± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l						± 20						± 30					± 50			
VPM MPV	µm ³ ; fl						± 2.0						± 2.0					± 2.0			
NEUT	#						± 0.25						± 0.80					± 1.90			
	%						± 10.0						± 10.0					± 10.0			
LYMPHO	#						± 0.20						± 0.60					± 1.50			
	%						± 8.0						± 8.0					± 8.0			
MONO	#						± 0.10						± 0.29					± 0.69			
	%						± 4.5						± 4.0					± 4.0			
EOS	#						± 0.12						± 0.21					± 0.73			
	%						± 5.0						± 2.8					± 4.0			
BASO	#						± 0.07						± 0.19					± 0.46			
	%						± 3.0						± 2.5					± 2.5			

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B