


**LOT** PX 424

**CONTROL**

 (Exp.) 2020-09-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE	
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8			± 1.0	18.9	18.9	18.9		± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40	2.40	2.40		± 0.16	4.66	4.66	4.66			± 0.20	5.13	5.13	5.13		± 0.25		
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.8	13.8	13.8			± 0.5	16.1	16.1	16.1		± 0.6		
	g/l	67	67	67		± 4	138	138	138			± 5	161	161	161		± 6		
	mmol/l	4.16	4.16	4.16		± 0.25	8.57	8.57	8.57			± 0.31	10.00	10.00	10.00		± 0.37		
HT HCT	%	19.4	19.4	19.4		± 1.5	38.7	38.7	38.7			± 2.0	44.6	44.6	44.6		± 2.5		
	l/l	0.194	0.194	0.194		± 0.015	0.387	0.387	0.387			± 0.020	0.446	0.446	0.446		± 0.025		
VGM MCV	µm <sup>3</sup> ; fl	81	81	81		± 5	83	83	83			± 5	87	87	87		± 5		
TGMH MCH	pg	27.9	27.9	27.9		± 2.0	29.6	29.6	29.6			± 2.0	31.4	31.4	31.4		± 2.5		
	fmol	1.73	1.73	1.73		± 0.12	1.84	1.84	1.84			± 0.12	1.95	1.95	1.95		± 0.16		
CCMH MCHC	g/dl	34.5	34.5	34.5		± 3.0	35.7	35.7	35.7			± 3.0	36.1	36.1	36.1		± 3.0		
	g/l	345	345	345		± 30	357	357	357			± 30	361	361	361		± 30		
	mmol/l	21.40	21.40	21.40		± 1.86	22.16	22.16	22.16			± 1.86	22.40	22.40	22.40		± 1.86		
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.5	16.5	16.5			± 4.0	15.0	15.0	15.0		± 4.0		
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	69	69		± 20	251	251	251			± 30	500	500	500		± 50		
VPM MPV	µm <sup>3</sup> ; fl	10.0	10.0	10.0		± 2.0	9.5	9.5	9.5			± 2.0	9.5	9.5	9.5		± 2.0		
NEUT	#	1.35	1.41	1.41		± 0.35	4.24	4.38	4.38			± 0.90	13.40	13.70	13.70		± 1.90		
	%	54.1	56.3	56.3		± 10.0	54.4	56.2	56.2			± 10.0	71.0	72.3	72.3		± 10.0		
LYMPHO	#	0.77	0.70	0.70		± 0.33	2.67	2.50	2.50			± 0.70	3.04	2.82	2.82		± 1.50		
	%	30.7	27.9	27.9		± 12.0	34.2	32.0	32.0			± 8.0	16.1	14.9	14.9		± 8.0		
MONO	#	0.14	0.15	0.15		± 0.14	0.44	0.46	0.46			± 0.44	1.13	1.12	1.12		± 1.12		
	%	5.6	5.8	5.8		± 5.6	5.7	5.9	5.9			± 5.7	6.0	5.9	5.9		± 5.9		
EOS	#	0.17	0.18	0.18		± 0.17	0.20	0.22	0.22			± 0.20	0.76	0.76	0.76		± 0.76		
	%	6.6	7.0	7.0		± 6.6	2.6	2.8	2.8			± 2.6	4.0	4.0	4.0		± 4.0		
BASO	#	0.08	0.08	0.08		± 0.08	0.24	0.24	0.24			± 0.24	0.55	0.55	0.55		± 0.55		
	%	3.0	3.0	3.0		± 3.0	3.1	3.1	3.1			± 3.1	2.9	2.9	2.9		± 2.9		

Ref. TEMP-0821 Rev.44 BACK / VERSO 1300040596