

 (Exp.) **2020-03-05**  
(YYYY-MM-DD)

**LOT** PX 421

**CONTROL**

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.7	2.7	2.7		± 0.4	8.0	8.0	8.0		± 1.0	19.2	19.2	19.2		± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.41	2.41		± 0.16	4.48	4.48	4.48		± 0.20	5.19	5.19	5.19		± 0.25		
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	13.2	13.2	13.2		± 0.5	16.4	16.4	16.4		± 0.6		
	g/l	70	70	70		± 4	132	132	132		± 5	164	164	164		± 6		
	mmol/l	4.35	4.35	4.35		± 0.25	8.20	8.20	8.20		± 0.31	10.18	10.18	10.18		± 0.37		
HT HCT	%	20.0	20.0	20.0		± 1.5	36.3	36.3	36.3		± 2.0	45.2	45.2	45.2		± 2.5		
	l/l	0.200	0.200	0.200		± 0.015	0.363	0.363	0.363		± 0.020	0.452	0.452	0.452		± 0.025		
VGM MCV	µm <sup>3</sup> ; fl	83	83	83		± 5	81	81	81		± 5	87	87	87		± 5		
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.5	29.5	29.5		± 2.0	31.6	31.6	31.6		± 2.5		
	fmol	1.80	1.80	1.80		± 0.12	1.83	1.83	1.83		± 0.12	1.96	1.96	1.96		± 0.16		
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	36.4	36.4	36.4		± 3.0	36.3	36.3	36.3		± 3.0		
	g/l	350	350	350		± 30	364	364	364		± 30	363	363	363		± 30		
	mmol/l	21.73	21.73	21.73		± 1.86	22.59	22.59	22.59		± 1.86	22.56	22.56	22.56		± 1.86		
IDR RDW	%	14.5	14.5	14.5		± 4.0	15.5	15.5	15.5		± 4.0	14.5	14.5	14.5		± 4.0		
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	73	73		± 20	245	245	245		± 30	500	500	500		± 50		
VPM MPV	µm <sup>3</sup> ; fl	10.1	10.1	10.1		± 2.0	9.8	9.8	9.8		± 2.0	9.2	9.2	9.2		± 2.0		
NEUT	#	1.47	1.53	1.53		± 0.35	4.78	4.90	4.90		± 0.90	13.50	13.70	13.70		± 1.90		
	%	54.6	56.7	56.7		± 10.0	59.8	61.2	61.2		± 10.0	70.1	71.4	71.4		± 10.0		
LYMPHO	#	0.81	0.74	0.74		± 0.33	2.35	2.20	2.20		± 0.70	3.00	2.76	2.76		± 1.50		
	%	30.0	27.3	27.3		± 12.0	29.4	27.5	27.5		± 8.0	15.6	14.4	14.4		± 8.0		
MONO	#	0.14	0.14	0.14		± 0.14	0.33	0.34	0.34		± 0.33	0.96	0.94	0.94		± 0.94		
	%	5.2	5.3	5.3		± 5.2	4.1	4.3	4.3		± 4.1	5.0	4.9	4.9		± 4.9		
EOS	#	0.21	0.22	0.22		± 0.21	0.30	0.33	0.33		± 0.30	1.13	1.13	1.13		± 1.13		
	%	7.7	8.2	8.2		± 7.7	3.8	4.1	4.1		± 3.8	5.9	5.9	5.9		± 5.9		
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.65	0.65	0.65		± 0.65		
	%	2.5	2.5	2.5		± 2.5	2.9	2.9	2.9		± 2.9	3.4	3.4	3.4		± 3.4		

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596