


**LOT** PX 431

**CONTROL**

 (Exp.) 2021-11-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6			± 0.4	8.5	8.5			± 1.0	18.7	18.7			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.31			± 0.16	4.54	4.54			± 0.20	5.20	5.20			± 0.25			
HB HGB	g/dl	6.6	6.6			± 0.4	13.5	13.5			± 0.5	16.3	16.3			± 0.6			
	g/l	66	66			± 4	135	135			± 5	163	163			± 6			
	mmol/l	4.10	4.10			± 0.25	8.38	8.38			± 0.31	10.12	10.12			± 0.37			
HT HCT	%	19.2	19.2			± 1.5	38.1	38.1			± 2.0	45.8	45.8			± 2.5			
	l/l	0.192	0.192			± 0.015	0.381	0.381			± 0.020	0.458	0.458			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	83			± 5	84	84			± 5	88	88			± 5			
TGMH MCH	pg	28.6	28.6			± 2.0	29.7	29.7			± 2.0	31.3	31.3			± 2.5			
	fmol	1.77	1.77			± 0.12	1.85	1.85			± 0.12	1.95	1.95			± 0.16			
CCMH MCHC	g/dl	34.4	34.4			± 3.0	35.4	35.4			± 3.0	35.6	35.6			± 3.0			
	g/l	344	344			± 30	354	354			± 30	356	356			± 30			
	mmol/l	21.38	21.38			± 1.86	21.98	21.98			± 1.86	22.12	22.12			± 1.86			
IDR RDW	%	15.0	15.0			± 4.0	15.0	15.0			± 4.0	14.5	14.5			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72			± 20	252	252			± 30	490	490			± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.7	9.7			± 2.0	9.3	9.3			± 2.0	8.6	8.6			± 2.0			
NEUT	#	1.44	1.44			± 0.35	4.99	4.99			± 0.90	13.20	13.20			± 1.90			
	%	55.5	55.5			± 10.0	58.7	58.7			± 10.0	70.4	70.4			± 10.0			
LYMPHO	#	0.71	0.71			± 0.33	2.30	2.30			± 0.70	2.58	2.58			± 1.50			
	%	27.3	27.3			± 12.0	27.0	27.0			± 8.0	13.8	13.8			± 8.0			
MONO	#	0.13	0.13			± 0.13	0.58	0.58			± 0.58	1.18	1.18			± 1.18			
	%	5.0	5.0			± 5.0	6.8	6.8			± 6.8	6.3	6.3			± 6.3			
EOS	#	0.20	0.20			± 0.20	0.37	0.37			± 0.37	1.22	1.22			± 1.22			
	%	7.7	7.7			± 7.7	4.4	4.4			± 4.4	6.5	6.5			± 6.5			
BASO	#	0.12	0.12			± 0.12	0.26	0.26			± 0.26	0.56	0.56			± 0.56			
	%	4.5	4.5			± 4.5	3.1	3.1			± 3.1	3.0	3.0			± 3.0			

Ref: TEMP-0821 Rev.45 BACK / VERSO 1300090306