


**LOT** PX 422

**CONTROL**

 (Exp.) 2020-05-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
			CONTROL				L	CONTROL				N	CONTROL				H		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6	2.6		± 0.4	7.7	7.7	7.7		± 1.0	18.7	18.7	18.7		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.38	2.38		± 0.16	4.70	4.70	4.70		± 0.20	5.35	5.35	5.35		± 0.25			
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.9	13.9	13.9		± 0.5	16.6	16.6	16.6		± 0.6			
	g/l	67	67	67		± 4	139	139	139		± 5	166	166	166		± 6			
	mmol/l	4.16	4.16	4.16		± 0.25	8.63	8.63	8.63		± 0.31	10.31	10.31	10.31		± 0.37			
HT HCT	%	19.5	19.5	19.5		± 1.5	39.0	39.0	39.0		± 2.0	46.5	46.5	46.5		± 2.5			
	l/l	0.195	0.195	0.195		± 0.015	0.390	0.390	0.390		± 0.020	0.465	0.465	0.465		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	82	82	82		± 5	83	83	83		± 5	87	87	87		± 5			
TGMH MCH	pg	28.2	28.2	28.2		± 2.0	29.6	29.6	29.6		± 2.0	31.0	31.0	31.0		± 2.5			
	fmol	1.75	1.75	1.75		± 0.12	1.84	1.84	1.84		± 0.12	1.93	1.93	1.93		± 0.16			
CCMH MCHC	g/dl	34.3	34.3	34.3		± 3.0	35.6	35.6	35.6		± 3.0	35.7	35.7	35.7		± 3.0			
	g/l	343	343	343		± 30	356	356	356		± 30	357	357	357		± 30			
	mmol/l	21.32	21.32	21.32		± 1.86	22.13	22.13	22.13		± 1.86	22.15	22.15	22.15		± 1.86			
IDR RDW	%	14.5	14.5	14.5		± 4.0	15.5	15.5	15.5		± 4.0	15.6	15.6	15.6		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	248	248	248		± 30	495	495	495		± 50			
VPM MPV	µm <sup>3</sup> ; fl	10.2	10.2	10.2		± 2.0	9.4	9.4	9.4		± 2.0	9.6	9.6	9.6		± 2.0			
NEUT	#	1.49	1.54	1.54		± 0.35	4.37	4.51	4.51		± 0.90	13.30	13.50	13.50		± 1.90			
	%	57.2	59.3	59.3		± 10.0	56.8	58.6	58.6		± 10.0	71.1	72.2	72.2		± 10.0			
LYMPHO	#	0.76	0.69	0.69		± 0.33	2.48	2.31	2.31		± 0.70	2.81	2.60	2.60		± 1.50			
	%	29.1	26.5	26.5		± 12.0	32.2	30.0	30.0		± 8.0	15.0	13.9	13.9		± 8.0			
MONO	#	0.11	0.11	0.11		± 0.11	0.36	0.38	0.38		± 0.36	1.10	1.10	1.10		± 1.10			
	%	4.3	4.4	4.4		± 4.3	4.7	4.9	4.9		± 4.7	5.9	5.9	5.9		± 5.9			
EOS	#	0.17	0.18	0.18		± 0.17	0.23	0.25	0.25		± 0.23	0.88	0.90	0.90		± 0.88			
	%	6.7	7.1	7.1		± 6.7	3.0	3.2	3.2		± 3.0	4.7	4.8	4.8		± 4.7			
BASO	#	0.07	0.07	0.07		± 0.07	0.25	0.25	0.25		± 0.25	0.62	0.60	0.60		± 0.60			
	%	2.7	2.7	2.7		± 2.7	3.3	3.3	3.3		± 3.3	3.3	3.2	3.2		± 3.2			

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596