


LOT PX 413

CONTROL

 (Exp.) 2018-11-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				120 RETIC		PENTRA				120 RETIC		PENTRA					
		120	DX120	DX NEXUS	DF NEXUS			120	DX120	DX NEXUS	DF NEXUS			120	DX120	DX NEXUS	DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.4	7.4	7.4		± 1.0	18.6	18.6	18.6		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38	2.38		± 0.16	4.64	4.64	4.64		± 0.20	5.25	5.25	5.25		± 0.25			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	14.0	14.0	14.0		± 0.5	16.5	16.5	16.5		± 0.6			
	g/l	70	70	70		± 4	140	140	140		± 5	165	165	165		± 6			
	mmol/l	4.35	4.35	4.35		± 0.25	8.69	8.69	8.69		± 0.31	10.25	10.25	10.25		± 0.37			
HT HCT	%	20.0	20.0	20.0		± 1.5	39.0	39.0	39.0		± 2.0	45.7	45.7	45.7		± 2.5			
	l/l	0.200	0.200	0.200		± 0.015	0.390	0.390	0.390		± 0.020	0.457	0.457	0.457		± 0.025			
VGM MCV	µm ³ ; fl	84	84	84		± 5	84	84	84		± 5	87	87	87		± 5			
TGMH MCH	pg	29.4	29.4	29.4		± 2.0	30.2	30.2	30.2		± 2.0	31.4	31.4	31.4		± 2.5			
	fmol	1.83	1.83	1.83		± 0.12	1.87	1.87	1.87		± 0.12	1.95	1.95	1.95		± 0.16			
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	35.9	35.9	35.9		± 3.0	36.1	36.1	36.1		± 3.0			
	g/l	350	350	350		± 30	359	359	359		± 30	361	361	361		± 30			
	mmol/l	21.74	21.74	21.74		± 1.86	22.31	22.31	22.31		± 1.86	22.43	22.43	22.43		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.0	16.0	16.0		± 4.0	14.8	14.8	14.8		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	252	252	252		± 30	490	490	490		± 50			
VPM MPV	µm ³ ; fl	9.0	9.0	9.0		± 2.0	8.8	8.8	8.8		± 2.0	8.8	8.8	8.8		± 2.0			
NEUT	#	1.46	1.49	1.49		± 0.35	4.39	4.40	4.40		± 0.90	13.70	13.70	13.70		± 1.90			
	%	60.8	62.1	62.1		± 10.0	59.3	59.4	59.4		± 10.0	73.6	73.6	73.6		± 10.0			
LYMPHO	#	0.62	0.56	0.56		± 0.33	2.23	2.09	2.09		± 0.70	2.72	2.51	2.51		± 1.50			
	%	25.8	23.4	23.4		± 12.0	30.1	28.2	28.2		± 8.0	14.6	13.5	13.5		± 8.0			
MONO	#	0.10	0.10	0.10		± 0.10	0.29	0.30	0.30		± 0.29	0.86	0.86	0.86		± 0.86			
	%	4.2	4.3	4.3		± 4.2	3.9	4.0	4.0		± 3.9	4.6	4.6	4.6		± 4.6			
EOS	#	0.16	0.17	0.17		± 0.16	0.27	0.30	0.30		± 0.27	0.87	0.89	0.89		± 0.87			
	%	6.7	7.1	7.1		± 6.7	3.7	4.0	4.0		± 3.7	4.7	4.8	4.8		± 4.7			
BASO	#	0.06	0.07	0.07		± 0.06	0.22	0.33	0.33		± 0.22	0.47	0.65	0.65		± 0.47			
	%	2.5	3.1	3.1		± 2.5	3.0	4.4	4.4		± 3.0	2.5	3.5	3.5		± 2.5			

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596