


LOT PX 423

CONTROL

 (Exp.) 2020-07-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.8	7.8	7.8		± 1.0	18.3	18.3	18.3		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.34	2.34		± 0.16	4.66	4.66	4.66		± 0.20	5.22	5.22	5.22		± 0.25		
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	13.7	13.7	13.7		± 0.5	16.2	16.2	16.2		± 0.6		
	g/l	66	66	66		± 4	137	137	137		± 5	162	162	162		± 6		
	mmol/l	4.10	4.10	4.10		± 0.25	8.51	8.51	8.51		± 0.31	10.06	10.06	10.06		± 0.37		
HT HCT	%	19.0	19.0	19.0		± 1.5	38.7	38.7	38.7		± 2.0	45.4	45.4	45.4		± 2.5		
	l/l	0.190	0.190	0.190		± 0.015	0.387	0.387	0.387		± 0.020	0.454	0.454	0.454		± 0.025		
VGM MCV	µm ³ ; fl	81	81	81		± 5	83	83	83		± 5	87	87	87		± 5		
TGMH MCH	pg	28.2	28.2	28.2		± 2.0	29.4	29.4	29.4		± 2.0	31.0	31.0	31.0		± 2.5		
	fmol	1.75	1.75	1.75		± 0.12	1.83	1.83	1.83		± 0.12	1.93	1.93	1.93		± 0.16		
CCMH MCHC	g/dl	34.8	34.8	34.8		± 3.0	35.4	35.4	35.4		± 3.0	35.7	35.7	35.7		± 3.0		
	g/l	348	348	348		± 30	354	354	354		± 30	357	357	357		± 30		
	mmol/l	21.62	21.62	21.62		± 1.86	22.00	22.00	22.00		± 1.86	22.15	22.15	22.15		± 1.86		
IDR RDW	%	15.5	15.5	15.5		± 4.0	15.3	15.3	15.3		± 4.0	15.2	15.2	15.2		± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	70		± 20	251	251	251		± 30	475	475	475		± 50		
VPM MPV	µm ³ ; fl	9.5	9.5	9.5		± 2.0	8.8	8.8	8.8		± 2.0	8.8	8.8	8.8		± 2.0		
NEUT	#	1.45	1.51	1.51		± 0.35	4.34	4.47	4.47		± 0.90	13.10	13.30	13.30		± 1.90		
	%	55.9	58.0	58.0		± 10.0	55.7	57.3	57.3		± 10.0	71.7	72.9	72.9		± 10.0		
LYMPHO	#	0.74	0.68	0.68		± 0.33	2.47	2.31	2.31		± 0.70	2.82	2.60	2.60		± 1.50		
	%	28.6	26.0	26.0		± 12.0	31.7	29.6	29.6		± 8.0	15.4	14.2	14.2		± 8.0		
MONO	#	0.14	0.14	0.14		± 0.14	0.48	0.50	0.50		± 0.48	1.08	1.08	1.08		± 1.08		
	%	5.3	5.4	5.4		± 5.3	6.2	6.4	6.4		± 6.2	5.9	5.9	5.9		± 5.9		
EOS	#	0.20	0.21	0.21		± 0.20	0.27	0.29	0.29		± 0.27	0.75	0.75	0.75		± 0.75		
	%	7.6	8.0	8.0		± 7.6	3.4	3.7	3.7		± 3.4	4.1	4.1	4.1		± 4.1		
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.53	0.53	0.53		± 0.53		
	%	2.6	2.6	2.6		± 2.6	3.0	3.0	3.0		± 3.0	2.9	2.9	2.9		± 2.9		

Ref: TEMP-0821 Rev.44 BACK / VERSO 1300040596