


LOT PX 433

CONTROL

 (Exp.) 2022-03-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6			± 0.4	7.7	7.7			± 1.0	18.5	18.5			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38			± 0.16	4.57	4.57			± 0.20	5.20	5.20			± 0.25			
HB HGB	g/dl	6.9	6.9			± 0.4	13.9	13.9			± 0.5	16.6	16.6			± 0.6			
	g/l	69	69			± 4	139	139			± 5	166	166			± 6			
	mmol/l	4.28	4.28			± 0.25	8.63	8.63			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	19.8	19.8			± 1.5	38.4	38.4			± 2.0	46.3	46.3			± 2.5			
	l/l	0.198	0.198			± 0.015	0.384	0.384			± 0.020	0.463	0.463			± 0.025			
VGM MCV	µm ³ ; fl	83	83			± 5	84	84			± 5	89	89			± 5			
TGMH MCH	pg	29.0	29.0			± 2.0	30.4	30.4			± 2.0	31.9	31.9			± 2.5			
	fmol	1.80	1.80			± 0.12	1.89	1.89			± 0.12	1.98	1.98			± 0.16			
CCMH MCHC	g/dl	34.9	34.9			± 3.0	36.2	36.2			± 3.0	35.9	35.9			± 3.0			
	g/l	349	349			± 30	362	362			± 30	359	359			± 30			
	mmol/l	21.69	21.69			± 1.86	22.49	22.49			± 1.86	22.27	22.27			± 1.86			
IDR RDW	%	16.0	16.0			± 4.0	15.5	15.5			± 4.0	14.5	14.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75			± 20	262	262			± 30	492	492			± 50			
VPM MPV	µm ³ ; fl	9.9	9.9			± 2.0	9.5	9.5			± 2.0	9.3	9.3			± 2.0			
NEUT	#	1.54	1.54			± 0.35	4.68	4.68			± 0.90	14.00	14.00			± 1.90			
	%	59.2	59.2			± 10.0	60.8	60.8			± 10.0	75.4	75.4			± 10.0			
LYMPHO	#	0.72	0.72			± 0.33	2.36	2.36			± 0.70	2.55	2.55			± 1.50			
	%	27.8	27.8			± 12.0	30.7	30.7			± 8.0	13.8	13.8			± 8.0			
MONO	#	0.08	0.08			± 0.08	0.19	0.19			± 0.19	0.54	0.54			± 0.54			
	%	3.1	3.1			± 3.1	2.5	2.5			± 2.5	2.9	2.9			± 2.9			
EOS	#	0.20	0.20			± 0.20	0.28	0.28			± 0.28	1.00	1.00			± 1.00			
	%	7.6	7.6			± 7.6	3.6	3.6			± 3.6	5.4	5.4			± 5.4			
BASO	#	0.06	0.06			± 0.06	0.18	0.18			± 0.18	0.46	0.46			± 0.46			
	%	2.3	2.3			± 2.3	2.4	2.4			± 2.4	2.5	2.5			± 2.5			

1300090306
BACK / VERSO
Ref. TEMP-0821 Rev.45