


LOT PX 414

CONTROL

 (Exp.) 2019-01-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
			CONTROL				L	CONTROL				N	CONTROL				H		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	8.1	8.1	8.1		± 1.0	19.2	19.2	19.2		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.41		± 0.16	4.70	4.70	4.70		± 0.20	5.24	5.24	5.24		± 0.25			
HB HGB	g/dl	7.1	7.1	7.1		± 0.4	13.9	13.9	13.9		± 0.5	16.2	16.2	16.2		± 0.6			
	g/l	71	71	71		± 4	139	139	139		± 5	162	162	162		± 6			
	mmol/l	4.41	4.41	4.41		± 0.25	8.63	8.63	8.63		± 0.31	10.06	10.06	10.06		± 0.37			
HT HCT	%	20.2	20.2	20.2		± 1.5	38.5	38.5	38.5		± 2.0	45.6	45.6	45.6		± 2.5			
	l/l	0.202	0.202	0.202		± 0.015	0.385	0.385	0.385		± 0.020	0.456	0.456	0.456		± 0.025			
VGM MCV	µm ³ ; fl	84	84	84		± 5	82	82	82		± 5	87	87	87		± 5			
TGMH MCH	pg	29.5	29.5	29.5		± 2.0	29.6	29.6	29.6		± 2.0	30.9	30.9	30.9		± 2.5			
	fmol	1.83	1.83	1.83		± 0.12	1.84	1.84	1.84		± 0.12	1.92	1.92	1.92		± 0.16			
CCMH MCHC	g/dl	35.1	35.1	35.1		± 3.0	36.1	36.1	36.1		± 3.0	35.5	35.5	35.5		± 3.0			
	g/l	351	351	351		± 30	361	361	361		± 30	355	355	355		± 30			
	mmol/l	21.78	21.78	21.78		± 1.86	22.40	22.40	22.40		± 1.86	22.07	22.07	22.07		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.2	16.2	16.2		± 4.0	15.2	15.2	15.2		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	73	73		± 20	252	252	252		± 30	490	490	490		± 50			
VPM MPV	µm ³ ; fl	10.3	10.3	10.3		± 2.0	10.3	10.3	10.3		± 2.0	9.8	9.8	9.8		± 2.0			
NEUT	#	1.41	1.45	1.45		± 0.35	4.63	4.71	4.71		± 0.90	14.20	14.20	14.20		± 1.90			
	%	56.2	58.1	58.1		± 10.0	57.1	58.2	58.2		± 10.0	73.7	73.5	73.5		± 10.0			
LYMPHO	#	0.73	0.68	0.68		± 0.33	2.62	2.44	2.44		± 0.70	2.86	2.63	2.63		± 1.50			
	%	29.2	27.2	27.2		± 12.0	32.3	30.1	30.1		± 8.0	14.9	13.7	13.7		± 8.0			
MONO	#	0.12	0.11	0.11		± 0.11	0.33	0.33	0.33		± 0.33	0.75	0.75	0.75		± 0.75			
	%	4.9	4.5	4.5		± 4.5	4.1	4.1	4.1		± 4.1	3.9	3.9	3.9		± 3.9			
EOS	#	0.18	0.20	0.20		± 0.18	0.28	0.33	0.33		± 0.28	0.96	1.08	1.08		± 0.96			
	%	7.2	7.8	7.8		± 7.2	3.5	4.1	4.1		± 3.5	5.0	5.6	5.6		± 5.0			
BASO	#	0.06	0.06	0.06		± 0.06	0.24	0.28	0.28		± 0.24	0.48	0.63	0.63		± 0.48			
	%	2.5	2.4	2.4		± 2.4	3.0	3.5	3.5		± 3.0	2.5	3.3	3.3		± 2.5			

Ref. TEMP-0821 Rev.42 BACK / VERSO 1300040596