

 (Exp.) **2022-05-05**  
(YYYY-MM-DD)

LOT		PX 434		CONTROL															
Rev 2		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5			± 0.4	7.9	7.9			± 1.0	18.3	18.3			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.32	2.32			± 0.16	4.63	4.63			± 0.20	5.25	5.25			± 0.25			
HB HGB	g/dl	6.6	6.6			± 0.4	13.9	13.9			± 0.5	16.6	16.6			± 0.6			
	g/l	66	66			± 4	139	139			± 5	166	166			± 6			
	mmol/l	4.10	4.10			± 0.25	8.63	8.63			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	19.0	19.0			± 1.5	38.9	38.9			± 2.0	46.2	46.2			± 2.5			
	l/l	0.190	0.190			± 0.015	0.389	0.389			± 0.020	0.462	0.462			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	82	82			± 5	84	84			± 5	88	88			± 5			
TGMH MCH	pg	28.4	28.4			± 2.0	30.0	30.0			± 2.0	31.6	31.6			± 2.5			
	fmol	1.77	1.77			± 0.12	1.86	1.86			± 0.12	1.96	1.96			± 0.16			
CCMH MCHC	g/dl	34.7	34.7			± 3.0	35.7	35.7			± 3.0	35.9	35.9			± 3.0			
	g/l	347	347			± 30	357	357			± 30	359	359			± 30			
	mmol/l	21.54	21.54			± 1.86	22.19	22.19			± 1.86	22.31	22.31			± 1.86			
IDR RDW	%	16.5	16.5			± 4.0	16.0	16.0			± 4.0	14.5	14.5			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70			± 20	263	263			± 30	490	490			± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.2	9.2			± 2.0	9.0	9.0			± 2.0	9.0	9.0			± 2.0			
NEUT	#	1.50	1.50			± 0.35	4.88	4.88			± 0.90	13.60	13.60			± 1.90			
	%	60.0	60.0			± 10.0	61.8	61.8			± 10.0	74.4	74.4			± 10.0			
LYMPHO	#	0.61	0.61			± 0.33	2.22	2.22			± 0.70	2.53	2.53			± 1.50			
	%	24.3	24.3			± 12.0	28.1	28.1			± 8.0	13.8	13.8			± 8.0			
MONO	#	0.12	0.12			± 0.12	0.21	0.21			± 0.21	0.55	0.55			± 0.55			
	%	4.9	4.9			± 4.9	2.6	2.6			± 2.6	3.0	3.0			± 3.0			
EOS	#	0.21	0.21			± 0.21	0.36	0.36			± 0.36	1.04	1.04			± 1.04			
	%	8.2	8.2			± 8.2	4.6	4.6			± 4.6	5.7	5.7			± 5.7			
BASO	#	0.07	0.07			± 0.07	0.23	0.23			± 0.23	0.57	0.57			± 0.57			
	%	2.6	2.6			± 2.6	2.9	2.9			± 2.9	3.1	3.1			± 3.1			

1300090306  
BACK / VERSO  
Ref: TEMP-0821 Rev.45