


**LOT** PX 401  
Rev 1


**CONTROL**

 (Exp.) 2016-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4	2.4	± 0.4	7.3	7.2	7.2	7.2	7.2	± 1.0	18.1	17.7	17.9	17.7	17.9	± 2.2		
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.25	2.26	2.25	2.28	± 0.16	4.50	4.51	4.46	4.51	4.50	± 0.20	5.05	5.10	5.06	5.10	5.10	± 0.25		
HB	HGB	g/dl	6.6	6.6	6.5	6.6	6.6	± 0.4	13.3	13.3	13.3	13.3	13.2	± 0.5	15.6	15.7	15.6	15.7	15.6	± 0.6		
		g/l	66	66	65	66	66	± 4	133	133	133	133	132	± 5	156	157	156	157	156	± 6		
		mmol/l	4.10	4.10	4.04	4.10	4.10	± 0.25	8.26	8.26	8.26	8.26	8.20	± 0.31	9.69	9.75	9.69	9.75	9.69	± 0.37		
HT	HCT	%	18.7	18.7	18.1	18.7	18.2	± 1.5	36.5	37.0	36.1	37.0	36.5	± 2.0	43.4	43.9	42.5	43.9	42.8	± 2.5		
		l/l	0.187	0.187	0.181	0.187	0.182	± 0.015	0.365	0.370	0.361	0.370	0.365	± 0.020	0.434	0.439	0.425	0.439	0.428	± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	81	83	80	83	80.0	± 5	81	82	81	82	81.0	± 5	86	86	84	86	84.0	± 5		
TGMH	MCH	pg	28.6	29.3	28.8	29.3	28.9	± 2.0	29.6	29.5	29.8	29.5	29.3	± 2.0	30.9	30.8	30.8	30.8	30.6	± 2.5		
		fmol	1.77	1.82	1.79	1.82	1.80	± 0.12	1.84	1.83	1.85	1.83	1.82	± 0.12	1.92	1.91	1.91	1.91	1.90	± 0.16		
CCMH	MCHC	g/dl	35.3	35.3	36.0	35.3	36.2	± 3.0	36.5	36.0	36.8	36.0	36.2	± 3.0	35.9	35.8	36.7	35.8	36.4	± 3.0		
		g/l	353	353	360	353	362	± 30	365	360	368	360	362	± 30	359	358	367	358	364	± 30		
		mmol/l	21.90	21.95	22.33	21.95	22.47	± 1.86	22.66	22.33	22.86	22.33	22.49	± 1.86	22.31	22.23	22.79	22.23	22.61	± 1.86		
IDR	RDW	%	13.2	13.0	12.8	13.0	12.0	± 4.0	13.3	13.5	12.4	13.5	12.5	± 4.0	13.0	13.5	12.0	13.5	12.4	± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	68	68	68	68	± 20	243	240	240	240	234	± 30	495	495	500	495	480	± 50		
VPM	MPV	µm <sup>3</sup> ; fl	9.1	9.4	9.2	9.4	8.8	± 2.0	8.6	9.2	9.0	9.2	8.6	± 2.0	9.0	9.4	9.2	9.4	8.7	± 2.0		
NEUT	#		1.37	1.41	1.38	1.41	1.40	± 0.35	4.12	4.03	4.05	4.03	4.05	± 0.90	12.96	12.69	12.94	12.69	12.98	± 1.90		
		%	57.2	58.9	57.5	58.9	58.2	± 10.0	56.4	56.0	56.2	56.0	56.2	± 10.0	71.6	71.7	72.3	71.7	72.5	± 10.0		
LYMPHO	#		0.65	0.64	0.64	0.64	0.66	± 0.33	2.40	2.39	2.43	2.39	2.45	± 0.70	2.70	2.69	2.72	2.69	2.83	± 1.50		
		%	27.2	26.5	26.5	26.5	27.5	± 12.0	32.9	33.2	33.8	33.2	34.0	± 8.0	14.9	15.2	15.2	15.2	15.8	± 8.0		
MONO	#		0.15	0.13	0.16	0.13	0.14	± 0.13	0.31	0.29	0.22	0.29	0.24	± 0.22	0.74	0.73	0.63	0.73	0.63	± 0.63		
		%	6.2	5.3	6.5	5.3	6.0	± 5.3	4.2	4.0	3.0	4.0	3.4	± 3.0	4.1	4.1	3.5	4.1	3.5	± 3.5		
EOS	#		0.14	0.14	0.14	0.14	0.12	± 0.12	0.23	0.25	0.25	0.25	0.22	± 0.22	0.91	0.83	0.81	0.83	0.72	± 0.72		
		%	6.0	6.0	6.0	6.0	4.8	± 4.8	3.1	3.5	3.5	3.5	3.0	± 3.0	5.0	4.7	4.5	4.7	4.0	± 4.0		
BASO	#		0.08	0.08	0.08	0.08	0.08	± 0.08	0.25	0.24	0.25	0.24	0.24	± 0.24	0.80	0.76	0.81	0.76	0.75	± 0.75		
		%	3.4	3.3	3.5	3.3	3.5	± 3.3	3.4	3.3	3.5	3.3	3.4	± 3.3	4.4	4.3	4.5	4.3	4.2	± 4.2		

**LOT** PX 401  
Rev 1

**CONTROL**

 (Exp.) 2016-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA					PENTRA					PENTRA							
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS					120
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4			± 0.4	7.5	7.5	7.5			± 1.0	18.7	18.7	18.7			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.31	2.31			± 0.16	4.53	4.53	4.53			± 0.20	5.16	5.16	5.16			± 0.25
		g/dl	6.6	6.6	6.6			± 0.4	13.2	13.2	13.2			± 0.5	15.6	15.6	15.6			± 0.6
HB	HGB	g/l	66	66	66			± 4	132	132	132			± 5	156	156	156			± 6
		mmol/l	4.10	4.10	4.10			± 0.25	8.20	8.20	8.20			± 0.31	9.69	9.69	9.69			± 0.37
HT	HCT	%	19.2	19.2	19.2			± 1.5	37.1	37.1	37.1			± 2.0	44.4	44.4	44.4			± 2.5
		l/l	0.192	0.192	0.192			± 0.015	0.371	0.371	0.371			± 0.020	0.444	0.444	0.444			± 0.025
VGM	MCV	µm <sup>3</sup> /fl	83	83	83			± 5	82	82	82			± 5	86	86	86			± 5
TGMH	MCH	pg	28.6	28.6	28.6			± 2.0	29.1	29.1	29.1			± 2.0	30.2	30.2	30.2			± 2.5
		fmol	1.77	1.77	1.77			± 0.12	1.81	1.81	1.81			± 0.12	1.88	1.88	1.88			± 0.16
		g/dl	34.4	34.4	34.4			± 3.0	35.5	35.5	35.5			± 3.0	35.2	35.2	35.2			± 3.0
CCMH	MCHC	g/l	344	344	344			± 30	355	355	355			± 30	352	352	352			± 30
		mmol/l	21.38	21.38	21.38			± 1.86	22.07	22.07	22.07			± 1.86	21.83	21.83	21.83			± 1.86
IDR	RDW	%	14.6	14.6	14.6			± 4.0	16.6	16.6	16.6			± 4.0	15.0	15.0	15.0			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72			± 20	245	245	245			± 30	502	502	502			± 50
VPM	MPV	µm <sup>3</sup> /fl	9.5	9.5	9.5			± 2.0	9.2	9.2	9.2			± 2.0	9.2	9.2	9.2			± 2.0
		#	1.49	1.53	1.53			± 0.35	4.22	4.35	4.35			± 0.90	13.70	13.90	13.90			± 1.90
NEUT		%	62.2	63.9	63.9			± 10.0	56.2	58.0	58.0			± 10.0	73.2	74.3	74.3			± 10.0
		#	0.56	0.51	0.51			± 0.33	2.47	2.30	2.30			± 0.70	2.69	2.49	2.49			± 1.50
LYMPHO		%	23.4	21.3	21.3			± 12.0	32.9	30.7	30.7			± 8.0	14.4	13.3	13.3			± 8.0
		#	0.15	0.15	0.15			± 0.15	0.35	0.37	0.37			± 0.35	1.03	1.01	1.01			± 1.01
MONO		%	6.3	6.4	6.4			± 6.3	4.7	4.9	4.9			± 4.7	5.5	5.4	5.4			± 5.4
		#	0.13	0.14	0.14			± 0.13	0.24	0.26	0.26			± 0.24	0.82	0.84	0.84			± 0.82
EOS		%	5.6	5.9	5.9			± 5.6	3.2	3.4	3.4			± 3.2	4.4	4.5	4.5			± 4.4
		#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.47	0.47	0.47			± 0.47
BASO		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

Ref: TEMP-0821 Rev.37 FRONT / RECTO 1300019740-A