


LOT PX 405
Rev 1


CONTROL

 (Exp.) 2017-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.6		± 0.4	7.5	7.4	7.5	7.4		7.6	± 1.0	18.2	17.9	18.1		
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.37	2.31	2.37	2.31	2.36	± 0.16	4.62	4.63	4.64	4.63	4.65	± 0.20	5.14	5.19	5.19	5.19	5.20	± 0.25	
HB	HGB	g/dl	6.9	6.9	6.9	6.9	6.9	± 0.4	13.6	13.6	13.6	13.6	13.5	± 0.5	16.7	16.7	16.8	16.7	16.6	± 0.6	
		g/l	69	69	69	69	69	± 4	136	136	136	136	135	± 5	167	167	168	167	166	± 6	
		mmol/l	4.28	4.28	4.28	4.28	4.28	± 0.25	8.45	8.45	8.45	8.45	8.38	± 0.31	10.37	10.37	10.43	10.37	10.31	± 0.37	
HT	HCT	%	19.2	19.4	19.2	19.4	19.4	± 1.5	37.0	37.5	37.1	37.0	37.7	± 2.0	45.7	46.2	45.7	45.7	45.8	± 2.5	
		l/l	0.192	0.194	0.192	0.194	0.194	± 0.015	0.370	0.375	0.371	0.370	0.377	± 0.020	0.457	0.462	0.457	0.457	0.458	± 0.025	
VGM	MCV	µm ³ ; fl	81	84	81	84	82.0	± 5	80	81	80	80	81.0	± 5	89	89	88	88	88.0	± 5	
TGMH	MCH	pg	29.1	29.9	29.1	29.9	29.2	± 2.0	29.4	29.4	29.3	29.4	29.0	± 2.0	32.5	32.2	32.4	32.2	31.9	± 2.5	
		fmol	1.81	1.85	1.81	1.85	1.82	± 0.12	1.83	1.82	1.82	1.82	1.80	± 0.12	2.02	2.00	2.01	2.00	1.98	± 0.16	
CCMH	MCHC	g/dl	35.9	35.6	35.9	35.6	35.7	± 3.0	36.8	36.3	36.6	36.7	35.8	± 3.0	36.5	36.2	36.8	36.6	36.3	± 3.0	
		g/l	359	356	359	356	357	± 30	368	363	366	367	358	± 30	365	362	368	366	363	± 30	
		mmol/l	22.32	22.08	22.32	22.08	22.14	± 1.86	22.85	22.52	22.75	22.80	22.26	± 1.86	22.67	22.45	22.84	22.71	22.53	± 1.86	
IDR	RDW	%	13.0	12.8	12.7	12.8	12.1	± 4.0	12.6	13.1	12.5	13.1	12.5	± 4.0	12.8	13.1	12.4	13.1	12.6	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	75	72	75	± 20	250	248	255	248	250	± 30	497	497	515	497	490	± 50	
VPM	MPV	µm ³ ; fl	9.6	9.8	9.6	9.8	9.4	± 2.0	9.5	9.8	9.6	9.8	9.1	± 2.0	8.6	8.9	8.8	8.9	8.4	± 2.0	
NEUT		#	1.40	1.39	1.39	1.39	1.45	± 0.35	4.05	4.01	4.09	4.01	4.03	± 0.90	12.65	12.53	12.85	12.53	13.01	± 1.90	
		%	56.0	55.5	55.7	55.5	55.6	± 10.0	54.0	54.2	54.5	54.2	53.0	± 10.0	69.5	70.0	71.0	70.0	71.5	± 10.0	
LYMPHO		#	0.75	0.78	0.79	0.78	0.85	± 0.33	2.51	2.54	2.63	2.54	2.81	± 0.70	2.82	2.94	2.90	2.94	3.06	± 1.50	
		%	30.1	31.3	31.5	31.3	32.7	± 12.0	33.5	34.3	35.0	34.3	37.0	± 8.0	15.5	16.4	16.0	16.4	16.8	± 8.0	
MONO		#	0.11	0.10	0.08	0.10	0.09	± 0.08	0.45	0.36	0.30	0.36	0.30	± 0.30	1.00	0.79	0.63	0.79	0.64	± 0.63	
		%	4.2	3.9	3.2	3.9	3.4	± 3.2	6.0	4.9	4.0	4.9	4.0	± 4.0	5.5	4.4	3.5	4.4	3.5	± 3.5	
EOS		#	0.16	0.15	0.16	0.15	0.13	± 0.13	0.23	0.24	0.23	0.24	0.23	± 0.23	0.91	0.86	0.91	0.86	0.73	± 0.73	
		%	6.5	6.1	6.3	6.1	5.0	± 5.0	3.1	3.2	3.0	3.2	3.0	± 3.0	5.0	4.8	5.0	4.8	4.0	± 4.0	
BASO		#	0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.25	0.26	0.25	0.23	± 0.23	0.82	0.79	0.81	0.79	0.76	± 0.76	
		%	3.2	3.2	3.3	3.2	3.3	± 3.2	3.4	3.4	3.5	3.4	3.0	± 3.0	4.5	4.4	4.5	4.4	4.2	± 4.2	

LOT PX 405
Rev 1

CONTROL

 (Exp.) 2017-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS			
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS			120 RETIC	DF120	DF NEXUS							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.7	7.7	7.7		± 1.0	18.8	18.8	18.8		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.36	2.36		± 0.16	4.66	4.66	4.66		± 0.20	5.25	5.25	5.25		± 0.25	
		g/dl	7.0	7.0	7.0		± 0.4	13.5	13.5	13.5		± 0.5	16.6	16.6	16.6		± 0.6	
HB	HGB	g/l	70	70	70		± 4	135	135	135		± 5	166	166	166		± 6	
		mmol/l	4.35	4.35	4.35		± 0.25	8.38	8.38	8.38		± 0.31	10.31	10.31	10.31		± 0.37	
HT	HCT	%	19.8	19.8	19.8		± 1.5	38.2	38.2	38.2		± 2.0	46.7	46.7	46.7		± 2.5	
		l/l	0.198	0.198	0.198		± 0.015	0.382	0.382	0.382		± 0.020	0.467	0.467	0.467		± 0.025	
VGM	MCV	µm ³ ; fl	84	84	84		± 5	82	82	82		± 5	89	89	89		± 5	
TGMH	MCH	pg	29.7	29.7	29.7		± 2.0	29.0	29.0	29.0		± 2.0	31.6	31.6	31.6		± 2.5	
		fmol	1.84	1.84	1.84		± 0.12	1.80	1.80	1.80		± 0.12	1.96	1.96	1.96		± 0.16	
		g/dl	35.3	35.3	35.3		± 3.0	35.3	35.3	35.3		± 3.0	35.5	35.5	35.5		± 3.0	
CCMH	MCHC	g/l	353	353	353		± 30	353	353	353		± 30	355	355	355		± 30	
		mmol/l	21.93	21.93	21.93		± 1.86	21.94	21.94	21.94		± 1.86	22.06	22.06	22.06		± 1.86	
IDR	RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	15.0	15.0	15.0		± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	77	77		± 20	252	252	252		± 30	495	495	495		± 50	
VPM	MPV	µm ³ ; fl	9.8	9.8	9.8		± 2.0	9.6	9.6	9.6		± 2.0	8.7	8.7	8.7		± 2.0	
		#	1.52	1.57	1.57		± 0.35	4.22	4.34	4.34		± 0.90	13.40	13.60	13.60		± 1.90	
NEUT		%	58.4	60.5	60.5		± 10.0	54.8	56.4	56.4		± 10.0	71.4	72.5	72.5		± 10.0	
		#	0.75	0.68	0.68		± 0.33	2.53	2.36	2.36		± 0.70	2.80	2.59	2.59		± 1.50	
LYMPHO		%	28.8	26.2	26.2		± 12.0	32.8	30.7	30.7		± 8.0	14.9	13.8	13.8		± 8.0	
		#	0.12	0.12	0.12		± 0.12	0.49	0.50	0.50		± 0.49	1.17	1.15	1.15		± 1.15	
MONO		%	4.5	4.6	4.6		± 4.5	6.3	6.5	6.5		± 6.3	6.2	6.1	6.1		± 6.1	
		#	0.15	0.16	0.16		± 0.15	0.24	0.26	0.26		± 0.24	0.94	0.96	0.96		± 0.94	
EOS		%	5.8	6.2	6.2		± 5.8	3.1	3.4	3.4		± 3.1	5.0	5.1	5.1		± 5.0	
		#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.47	0.47	0.47		± 0.47	
BASO		%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5	

1300032435-A
FRONT / RECTO
Ref: TEMP-0821 Rev.40