


LOT PX 415
Rev 2

CONTROL

 (Exp.) 2019-03-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE	
			PENTRA					PENTRA					PENTRA					
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	± 0.4	7.7	7.7	7.7	± 1.0	19.1	19.1	19.1	± 2.2					
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.41	± 0.16	4.63	4.63	4.63	± 0.20	5.23	5.23	5.23	± 0.25					
HB HGB	g/dl	7.0	7.0	7.0	± 0.4	13.6	13.6	13.6	± 0.5	15.9	15.9	15.9	± 0.6					
	g/l	70	70	70	± 4	136	136	136	± 5	159	159	159	± 6					
HT HCT	mmol/l	4.35	4.35	4.35	± 0.25	8.45	8.45	8.45	± 0.31	9.87	9.87	9.87	± 0.37					
	%	20.2	20.2	20.2	± 1.5	38.0	38.0	38.0	± 2.0	45.5	45.5	45.5	± 2.5					
VGM MCV	l/l	0.202	0.202	0.202	± 0.015	0.380	0.380	0.380	± 0.020	0.455	0.455	0.455	± 0.025					
	µm ³ ·fl	84	84	84	± 5	82	82	82	± 5	87	87	87	± 5					
TGMH MCH	pg	29.0	29.0	29.0	± 2.0	29.4	29.4	29.4	± 2.0	30.4	30.4	30.4	± 2.5					
	fmoi	1.80	1.80	1.80	± 0.12	1.82	1.82	1.82	± 0.12	1.89	1.89	1.89	± 0.16					
CCMH MCHC	g/dl	34.6	34.6	34.6	± 3.0	35.8	35.8	35.8	± 3.0	34.9	34.9	34.9	± 3.0					
	g/l	346	346	346	± 30	358	358	358	± 30	349	349	349	± 30					
IDR RDW	mmol/l	21.47	21.47	21.47	± 1.86	22.25	22.25	22.25	± 1.86	21.70	21.70	21.70	± 1.86					
	%	15.2	15.2	15.2	± 4.0	16.0	16.0	16.0	± 4.0	15.0	15.0	15.0	± 4.0					
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70	± 20	255	255	255	± 30	500	500	500	± 50					
VPM MPV	µm ³ ·fl	9.0	9.0	9.0	± 2.0	9.3	9.3	9.3	± 2.0	9.6	9.6	9.6	± 2.0					
NEUT	#	1.43	1.48	1.48	± 0.35	4.50	4.62	4.62	± 0.90	13.90	14.10	14.10	± 1.90					
	%	57.0	59.0	59.0	± 10.0	58.4	60.0	60.0	± 10.0	73.0	74.0	74.0	± 10.0					
LYMPHO	#	0.72	0.66	0.66	± 0.33	2.39	2.24	2.24	± 0.70	2.69	2.52	2.52	± 1.50					
	%	28.7	26.2	26.2	± 12.0	31.0	29.1	29.1	± 8.0	14.1	13.2	13.2	± 8.0					
MONO	#	0.09	0.09	0.09	± 0.09	0.28	0.29	0.29	± 0.28	0.82	0.80	0.80	± 0.80					
	%	3.5	3.5	3.5	± 3.5	3.7	3.8	3.8	± 3.7	4.3	4.2	4.2	± 4.2					
EOS	#	0.20	0.21	0.21	± 0.20	0.26	0.28	0.28	± 0.26	0.97	0.97	0.97	± 0.97					
	%	7.8	8.3	8.3	± 7.8	3.4	3.6	3.6	± 3.4	5.1	5.1	5.1	± 5.1					
BASO	#	0.08	0.08	0.08	± 0.08	0.27	0.27	0.27	± 0.27	0.67	0.67	0.67	± 0.67					
	%	3.0	3.0	3.0	± 3.0	3.5	3.5	3.5	± 3.5	3.5	3.5	3.5	± 3.5					

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596