


LOT PX 435

CONTROL

 (Exp.) 2022-07-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE	
		CONTROL				L	CONTROL				N	CONTROL					H
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1			± 0.4	8.8	8.8			± 1.0	18.3	18.3			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.46	2.46			± 0.16	4.65	4.65			± 0.20	5.15	5.15			± 0.25	
HB HGB	g/dl	6.5	6.5			± 0.4	13.7	13.7			± 0.5	15.9	15.9			± 0.6	
	g/l	65	65			± 4	137	137			± 5	159	159			± 6	
	mmol/l	4.04	4.04			± 0.25	8.51	8.51			± 0.31	9.87	9.87			± 0.37	
HT HCT	%	20.2	20.2			± 1.5	40.9	40.9			± 2.0	47.4	47.4			± 2.5	
	l/l	0.202	0.202			± 0.015	0.409	0.409			± 0.020	0.474	0.474			± 0.025	
VGM MCV	µm ³ ; fl	82	82			± 5	88	88			± 5	92	92			± 5	
TGMH MCH	pg	26.4	26.4			± 2.0	29.5	29.5			± 2.0	30.9	30.9			± 2.5	
	fmol	1.64	1.64			± 0.12	1.83	1.83			± 0.12	1.92	1.92			± 0.16	
CCMH MCHC	g/dl	32.2	32.2			± 3.0	33.5	33.5			± 3.0	33.6	33.6			± 3.0	
	g/l	322	322			± 30	335	335			± 30	336	336			± 30	
	mmol/l	20.01	20.01			± 1.86	20.79	20.79			± 1.86	20.84	20.84			± 1.86	
IDR RDW	%	16.5	16.5			± 4.0	16.0	16.0			± 4.0	15.0	15.0			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	78			± 20	234	234			± 30	480	480			± 50	
VPM MPV	µm ³ ; fl	10.5	10.5			± 2.0	11.0	11.0			± 2.0	10.5	10.5			± 2.0	
NEUT	#	1.62	1.62			± 0.35	4.85	4.85			± 0.90	13.20	13.20			± 1.90	
	%	52.4	52.4			± 10.0	55.1	55.1			± 10.0	72.0	72.0			± 10.0	
LYMPHO	#	0.93	0.93			± 0.33	3.03	3.03			± 0.70	3.02	3.02			± 1.50	
	%	30.1	30.1			± 12.0	34.4	34.4			± 8.0	16.5	16.5			± 8.0	
MONO	#	0.23	0.23			± 0.23	0.37	0.37			± 0.37	0.42	0.42			± 0.42	
	%	7.5	7.5			± 7.5	4.2	4.2			± 4.2	2.3	2.3			± 2.3	
EOS	#	0.18	0.18			± 0.18	0.33	0.33			± 0.33	0.86	0.86			± 0.86	
	%	5.9	5.9			± 5.9	3.7	3.7			± 3.7	4.7	4.7			± 4.7	
BASO	#	0.13	0.13			± 0.13	0.23	0.23			± 0.23	0.82	0.82			± 0.82	
	%	4.1	4.1			± 4.1	2.6	2.6			± 2.6	4.5	4.5			± 4.5	

Ref: TEMP-0821 Rev.45 BACK / VERSO 1300090306