

 (Exp.) **2021-01-05**
(YYYY-MM-DD)

LOT PX 426

CONTROL

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.8		± 0.4	8.2	8.2	8.2		± 1.0	19.4	19.4	19.4		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.27	2.27	2.27		± 0.16	4.53	4.53	4.53		± 0.20	5.08	5.08	5.08		± 0.25	
HB	HGB	g/dl	6.5	6.5	6.5		± 0.4	13.6	13.6	13.6		± 0.5	16.2	16.2	16.2		± 0.6	
		g/l	65	65	65		± 4	136	136	136		± 5	162	162	162		± 6	
		mmol/l	4.04	4.04	4.04		± 0.25	8.45	8.45	8.45		± 0.31	10.06	10.06	10.06		± 0.37	
HT	HCT	%	18.6	18.6	18.6		± 1.5	37.6	37.6	37.6		± 2.0	45.2	45.2	45.2		± 2.5	
		l/l	0.186	0.186	0.186		± 0.015	0.376	0.376	0.376		± 0.020	0.452	0.452	0.452		± 0.025	
VGM	MCV	µm ³ ; fl	82	82	82		± 5	83	83	83		± 5	89	89	89		± 5	
TGMH	MCH	pg	28.6	28.6	28.6		± 2.0	30.0	30.0	30.0		± 2.0	31.9	31.9	31.9		± 2.5	
		fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86		± 0.12	1.98	1.98	1.98		± 0.16	
CCMH	MCHC	g/dl	34.9	34.9	34.9		± 3.0	36.2	36.2	36.2		± 3.0	35.8	35.8	35.8		± 3.0	
		g/l	349	349	349		± 30	362	362	362		± 30	358	358	358		± 30	
		mmol/l	21.69	21.69	21.69		± 1.86	22.46	22.46	22.46		± 1.86	22.25	22.25	22.25		± 1.86	
IDR	RDW	%	15.5	15.5	15.5		± 4.0	15.0	15.0	15.0		± 4.0	14.2	14.2	14.2		± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69	69		± 20	248	248	248		± 30	487	487	487		± 50	
VPM	MPV	µm ³ ; fl	9.7	9.7	9.7		± 2.0	8.8	8.8	8.8		± 2.0	8.6	8.6	8.6		± 2.0	
NEUT	#		1.46	1.51	1.51		± 0.35	4.34	4.44	4.44		± 0.90	13.40	13.60	13.60		± 1.90	
		%	52.1	54.0	54.0		± 10.0	52.9	54.1	54.1		± 10.0	68.8	69.9	69.9		± 10.0	
LYMPHO	#		0.80	0.72	0.72		± 0.33	2.38	2.22	2.22		± 0.70	2.83	2.62	2.62		± 1.50	
		%	28.4	25.8	25.8		± 12.0	29.0	27.1	27.1		± 8.0	14.6	13.5	13.5		± 8.0	
MONO	#		0.25	0.25	0.25		± 0.25	0.83	0.85	0.85		± 0.83	1.94	1.94	1.94		± 1.94	
		%	8.9	9.1	9.1		± 8.9	10.1	10.4	10.4		± 10.1	10.0	10.0	10.0		± 10.0	
EOS	#		0.21	0.22	0.22		± 0.21	0.41	0.44	0.44		± 0.41	0.72	0.72	0.72		± 0.72	
		%	7.5	8.0	8.0		± 7.5	5.0	5.4	5.4		± 5.0	3.7	3.7	3.7		± 3.7	
BASO	#		0.09	0.09	0.09		± 0.09	0.25	0.25	0.25		± 0.25	0.56	0.56	0.56		± 0.56	
		%	3.1	3.1	3.1		± 3.1	3.0	3.0	3.0		± 3.0	2.9	2.9	2.9		± 2.9	

1300040596
BACK / VERSO
Ref. TEMP-0821 Rev.44