


LOT PX 417

CONTROL

 (Exp.) 2019-07-05
(YYYY-MM-DD)

Rev 3

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				120 RETIC		PENTRA				120 RETIC		PENTRA				120 RETIC	
		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS	120		DX120	DX NEXUS	DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.5	7.5	7.5		± 1.0	18.9	18.9	18.9		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.32	2.32		± 0.16	4.62	4.62	4.62		± 0.20	5.24	5.24	5.24		± 0.25			
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.9	13.9	13.9		± 0.5	16.4	16.4	16.4		± 0.6			
	g/l	67	67	67		± 4	139	139	139		± 5	164	164	164		± 6			
	mmol/l	4.16	4.16	4.16		± 0.25	8.63	8.63	8.63		± 0.31	10.18	10.18	10.18		± 0.37			
HT HCT	%	19.3	19.3	19.3		± 1.5	38.8	38.8	38.8		± 2.0	46.1	46.1	46.1		± 2.5			
	l/l	0.193	0.193	0.193		± 0.015	0.388	0.388	0.388		± 0.020	0.461	0.461	0.461		± 0.025			
VGM MCV	µm ³ ; fl	83	83	83		± 5	84	84	84		± 5	88	88	88		± 5			
TGMH MCH	pg	28.9	28.9	28.9		± 2.0	30.1	30.1	30.1		± 2.0	31.3	31.3	31.3		± 2.5			
	fmol	1.79	1.79	1.79		± 0.12	1.87	1.87	1.87		± 0.12	1.94	1.94	1.94		± 0.16			
CCMH MCHC	g/dl	34.8	34.8	34.8		± 3.0	35.8	35.8	35.8		± 3.0	35.6	35.6	35.6		± 3.0			
	g/l	348	348	348		± 30	358	358	358		± 30	356	356	356		± 30			
	mmol/l	21.61	21.61	21.61		± 1.86	22.24	22.24	22.24		± 1.86	22.09	22.09	22.09		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.0	16.0	16.0		± 4.0	16.0	16.0	16.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	70		± 20	250	250	250		± 30	495	495	495		± 50			
VPM MPV	µm ³ ; fl	10.0	10.0	10.0		± 2.0	9.9	9.9	9.9		± 2.0	9.6	9.6	9.6		± 2.0			
NEUT	#	1.45	1.51	1.51		± 0.35	4.27	4.40	4.40		± 0.90	13.80	14.00	14.00		± 1.90			
	%	58.1	60.2	60.2		± 10.0	56.9	58.6	58.6		± 10.0	72.8	73.7	73.7		± 10.0			
LYMPHO	#	0.70	0.64	0.64		± 0.33	2.36	2.21	2.21		± 0.70	2.48	2.31	2.31		± 1.50			
	%	28.0	25.4	25.4		± 12.0	31.5	29.4	29.4		± 8.0	13.1	12.2	12.2		± 8.0			
MONO	#	0.13	0.14	0.14		± 0.13	0.38	0.39	0.39		± 0.38	1.17	1.15	1.15		± 1.15			
	%	5.2	5.4	5.4		± 5.2	5.0	5.2	5.2		± 5.0	6.2	6.1	6.1		± 6.1			
EOS	#	0.14	0.15	0.15		± 0.14	0.17	0.19	0.19		± 0.17	0.74	0.76	0.76		± 0.74			
	%	5.6	5.9	5.9		± 5.6	2.3	2.5	2.5		± 2.3	3.9	4.0	4.0		± 3.9			
BASO	#	0.08	0.08	0.08		± 0.08	0.32	0.32	0.32		± 0.32	0.76	0.76	0.76		± 0.76			
	%	3.1	3.1	3.1		± 3.1	4.3	4.3	4.3		± 4.3	4.0	4.0	4.0		± 4.0			

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596