


**LOT** PX 427

**CONTROL**

 (Exp.) 2021-03-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				120 RETIC		PENTRA				120 RETIC		PENTRA				120 RETIC	
		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS	120		DX120	DX NEXUS	DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.7	2.7	2.7		± 0.4	8.1	8.1	8.1		± 1.0	19.5	19.5	19.5		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.34	2.34		± 0.16	4.54	4.54	4.54		± 0.20	5.20	5.20	5.20		± 0.25			
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.6	13.6	13.6		± 0.5	16.7	16.7	16.7		± 0.6			
	g/l	67	67	67		± 4	136	136	136		± 5	167	167	167		± 6			
	mmol/l	4.16	4.16	4.16		± 0.25	8.45	8.45	8.45		± 0.31	10.37	10.37	10.37		± 0.37			
HT HCT	%	19.4	19.4	19.4		± 1.5	38.1	38.1	38.1		± 2.0	46.8	46.8	46.8		± 2.5			
	l/l	0.194	0.194	0.194		± 0.015	0.381	0.381	0.381		± 0.020	0.468	0.468	0.468		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	83	83		± 5	84	84	84		± 5	90	90	90		± 5			
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	30.0	30.0	30.0		± 2.0	32.1	32.1	32.1		± 2.5			
	fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86		± 0.12	1.99	1.99	1.99		± 0.16			
CCMH MCHC	g/dl	34.5	34.5	34.5		± 3.0	35.7	35.7	35.7		± 3.0	35.7	35.7	35.7		± 3.0			
	g/l	345	345	345		± 30	357	357	357		± 30	357	357	357		± 30			
	mmol/l	21.42	21.42	21.42		± 1.86	22.15	22.15	22.15		± 1.86	22.16	22.16	22.16		± 1.86			
IDR RDW	%	14.0	14.0	14.0		± 4.0	16.5	16.5	16.5		± 4.0	14.5	14.5	14.5		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72		± 20	255	255	255		± 30	485	485	485		± 50			
VPM MPV	µm <sup>3</sup> ; fl	8.8	8.8	8.8		± 2.0	8.5	8.5	8.5		± 2.0	9.1	9.1	9.1		± 2.0			
NEUT	#	1.40	1.46	1.46		± 0.35	4.10	4.23	4.23		± 0.90	13.10	13.30	13.30		± 1.90			
	%	51.8	54.0	54.0		± 10.0	50.6	52.2	52.2		± 10.0	66.8	68.0	68.0		± 10.0			
LYMPHO	#	0.83	0.76	0.76		± 0.33	2.69	2.51	2.51		± 0.70	3.02	2.81	2.81		± 1.50			
	%	30.9	28.1	28.1		± 12.0	33.2	31.0	31.0		± 8.0	15.5	14.4	14.4		± 8.0			
MONO	#	0.23	0.23	0.23		± 0.23	0.82	0.84	0.84		± 0.82	1.93	1.91	1.91		± 1.91			
	%	8.5	8.7	8.7		± 8.5	10.1	10.4	10.4		± 10.1	9.9	9.8	9.8		± 9.8			
EOS	#	0.17	0.18	0.18		± 0.17	0.26	0.28	0.28		± 0.26	0.98	0.98	0.98		± 0.98			
	%	6.2	6.6	6.6		± 6.2	3.2	3.5	3.5		± 3.2	5.0	5.0	5.0		± 5.0			
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.55	0.55	0.55		± 0.55			
	%	2.6	2.6	2.6		± 2.6	2.9	2.9	2.9		± 2.9	2.8	2.8	2.8		± 2.8			

1300040596  
BACK / VERSO  
Ref. TEMP-0821 Rev.44