


LOT PX 418

CONTROL

 (Exp.) **2019-09-05**
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	$10^9/mm^3; 10^9/l$	2.5	2.5	2.5	± 0.4	7.7	7.7	7.7	± 1.0	19.1	19.1	19.1	± 2.2					
GR RBC	$10^9/mm^3; 10^{12}/l$	2.33	2.33	2.33	± 0.16	4.61	4.61	4.61	± 0.20	5.26	5.26	5.26	± 0.25					
HB HGB	g/dl	6.9	6.9	6.9	± 0.4	13.9	13.9	13.9	± 0.5	16.6	16.6	16.6	± 0.6					
	g/l	69	69	69	± 4	139	139	139	± 5	166	166	166	± 6					
	mmol/l	4.28	4.28	4.28	± 0.25	8.63	8.63	8.63	± 0.31	10.31	10.31	10.31	± 0.37					
HT HCT	%	19.8	19.8	19.8	± 1.5	38.7	38.7	38.7	± 2.0	46.8	46.8	46.8	± 2.5					
	l/l	0.198	0.198	0.198	± 0.015	0.387	0.387	0.387	± 0.020	0.468	0.468	0.468	± 0.025					
VGM MCV	$\mu m^3; fl$	85	85	85	± 5	84	84	84	± 5	89	89	89	± 5					
TGMH MCH	pg	29.6	29.6	29.6	± 2.0	30.2	30.2	30.2	± 2.0	31.6	31.6	31.6	± 2.5					
	fmol	1.84	1.84	1.84	± 0.12	1.87	1.87	1.87	± 0.12	1.96	1.96	1.96	± 0.16					
CCMH MCHC	g/dl	34.8	34.8	34.8	± 3.0	35.9	35.9	35.9	± 3.0	35.5	35.5	35.5	± 3.0					
	g/l	348	348	348	± 30	359	359	359	± 30	355	355	355	± 30					
	mmol/l	21.64	21.64	21.64	± 1.86	22.29	22.29	22.29	± 1.86	22.02	22.02	22.02	± 1.86					
IDR RDW	%	15.0	15.0	15.0	± 4.0	16.0	16.0	16.0	± 4.0	15.5	15.5	15.5	± 4.0					
PLAQ. PLTS	$10^3/mm^3; 10^9/l$	72	72	72	± 20	255	255	255	± 30	510	510	510	± 50					
VPM MPV	$\mu m^3; fl$	9.8	9.8	9.8	± 2.0	9.5	9.5	9.5	± 2.0	9.5	9.5	9.5	± 2.0					
NEUT	#	1.50	1.54	1.54	± 0.35	4.28	4.41	4.41	± 0.90	13.40	13.60	13.60	± 1.90					
	%	59.8	61.7	61.7	± 10.0	55.6	57.3	57.3	± 10.0	70.1	71.3	71.3	± 10.0					
LYMPHO	#	0.64	0.58	0.58	± 0.33	2.38	2.22	2.22	± 0.70	2.87	2.65	2.65	± 1.50					
	%	25.7	23.3	23.3	± 12.0	30.9	28.8	28.8	± 8.0	15.0	13.9	13.9	± 8.0					
MONO	#	0.13	0.14	0.14	± 0.13	0.48	0.49	0.49	± 0.48	1.22	1.20	1.20	± 1.20					
	%	5.3	5.4	5.4	± 5.3	6.2	6.4	6.4	± 6.2	6.4	6.3	6.3	± 6.3					
EOS	#	0.16	0.17	0.17	± 0.16	0.25	0.26	0.26	± 0.25	0.88	0.88	0.88	± 0.88					
	%	6.4	6.8	6.8	± 6.4	3.2	3.4	3.4	± 3.2	4.6	4.6	4.6	± 4.6					
BASO	#	0.07	0.07	0.07	± 0.07	0.32	0.32	0.32	± 0.32	0.74	0.74	0.74	± 0.74					
	%	2.8	2.8	2.8	± 2.8	4.1	4.1	4.1	± 4.1	3.9	3.9	3.9	± 3.9					

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596