


**LOT** PX 428

**CONTROL**

 (Exp.) 2021-05-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.7	2.7			± 0.4	8.4	8.4			± 1.0	19.9	19.9			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40	2.40			± 0.16	4.72	4.72			± 0.20	5.24	5.24			± 0.25			
HB HGB	g/dl	6.9	6.9			± 0.4	14.1	14.1			± 0.5	16.6	16.6			± 0.6			
	g/l	69	69			± 4	141	141			± 5	166	166			± 6			
	mmol/l	4.28	4.28			± 0.25	8.76	8.76			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	19.9	19.9			± 1.5	39.6	39.6			± 2.0	46.6	46.6			± 2.5			
	l/l	0.199	0.199			± 0.015	0.396	0.396			± 0.020	0.466	0.466			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	83			± 5	84	84			± 5	89	89			± 5			
TGMH MCH	pg	28.8	28.8			± 2.0	29.9	29.9			± 2.0	31.7	31.7			± 2.5			
	fmol	1.79	1.79			± 0.12	1.86	1.86			± 0.12	1.97	1.97			± 0.16			
CCMH MCHC	g/dl	34.6	34.6			± 3.0	35.6	35.6			± 3.0	35.6	35.6			± 3.0			
	g/l	346	346			± 30	356	356			± 30	356	356			± 30			
	mmol/l	21.51	21.51			± 1.86	22.08	22.08			± 1.86	22.10	22.10			± 1.86			
IDR RDW	%	16.0	16.0			± 4.0	15.0	15.0			± 4.0	14.0	14.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	75			± 20	255	255			± 30	490	490			± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.2	9.2			± 2.0	8.8	8.8			± 2.0	9.1	9.1			± 2.0			
NEUT	#	1.36	1.36			± 0.35	4.42	4.42			± 0.90	13.41	13.41			± 1.90			
	%	50.4	50.4			± 10.0	52.6	52.6			± 10.0	67.4	67.4			± 10.0			
LYMPHO	#	0.88	0.88			± 0.33	2.69	2.69			± 0.70	3.22	3.22			± 1.50			
	%	32.6	32.6			± 12.0	32.0	32.0			± 8.0	16.2	16.2			± 8.0			
MONO	#	0.18	0.18			± 0.18	0.67	0.67			± 0.67	1.91	1.91			± 1.91			
	%	6.5	6.5			± 6.5	8.0	8.0			± 8.0	9.6	9.6			± 9.6			
EOS	#	0.22	0.22			± 0.22	0.38	0.38			± 0.38	0.72	0.72			± 0.72			
	%	8.1	8.1			± 8.1	4.5	4.5			± 4.5	3.6	3.6			± 3.6			
BASO	#	0.06	0.06			± 0.06	0.24	0.24			± 0.24	0.64	0.64			± 0.64			
	%	2.4	2.4			± 2.4	2.9	2.9			± 2.9	3.2	3.2			± 3.2			

Ref: TEMP-0821 Rev.44 BACK / VERSO 1300040596