

LOT		PX 429		CONTROL		(Exp.) 2021-07-05													
Rev 2						(YYYY-MM-DD)													
		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7			± 0.4	7.8	7.8			± 1.0	18.8	18.8			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.17	2.17			± 0.16	4.64	4.64			± 0.20	5.27	5.27			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	13.9	13.9			± 0.5	16.7	16.7			± 0.6			
	g/l	61	61			± 4	139	139			± 5	167	167			± 6			
	mmol/l	3.79	3.79			± 0.25	8.63	8.63			± 0.31	10.37	10.37			± 0.37			
HT HCT	%	17.6	17.6			± 1.5	39.0	39.0			± 2.0	46.4	46.4			± 2.5			
	l/l	0.176	0.176			± 0.015	0.390	0.390			± 0.020	0.464	0.464			± 0.025			
VGM MCV	µm ³ ; fl	81	81			± 5	84	84			± 5	88	88			± 5			
TGMH MCH	pg	28.1	28.1			± 2.0	30.0	30.0			± 2.0	31.7	31.7			± 2.5			
	fmol	1.75	1.75			± 0.12	1.86	1.86			± 0.12	1.97	1.97			± 0.16			
CCMH MCHC	g/dl	34.7	34.7			± 3.0	35.7	35.7			± 3.0	36.0	36.0			± 3.0			
	g/l	347	347			± 30	357	357			± 30	360	360			± 30			
	mmol/l	21.55	21.55			± 1.86	22.15	22.15			± 1.86	22.36	22.36			± 1.86			
IDR RDW	%	15.8	15.8			± 4.0	16.5	16.5			± 4.0	15.0	15.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72			± 20	250	250			± 30	497	497			± 50			
VPM MPV	µm ³ ; fl	9.6	9.6			± 2.0	9.6	9.6			± 2.0	9.2	9.2			± 2.0			
NEUT	#	1.43	1.43			± 0.35	4.29	4.29			± 0.90	13.10	13.10			± 1.90			
	%	53.1	53.1			± 10.0	55.0	55.0			± 10.0	69.5	69.5			± 10.0			
LYMPHO	#	0.72	0.72			± 0.33	2.40	2.40			± 0.70	2.80	2.80			± 1.50			
	%	26.6	26.6			± 12.0	30.8	30.8			± 8.0	14.9	14.9			± 8.0			
MONO	#	0.18	0.18			± 0.18	0.63	0.63			± 0.63	1.56	1.56			± 1.56			
	%	6.6	6.6			± 6.6	8.1	8.1			± 8.1	8.3	8.3			± 8.3			
EOS	#	0.27	0.27			± 0.27	0.23	0.23			± 0.23	0.88	0.88			± 0.88			
	%	10.1	10.1			± 10.1	3.0	3.0			± 3.0	4.7	4.7			± 4.7			
BASO	#	0.10	0.10			± 0.10	0.24	0.24			± 0.24	0.49	0.49			± 0.49			
	%	3.6	3.6			± 3.6	3.1	3.1			± 3.1	2.6	2.6			± 2.6			

1300040596
BACK / VERSO
Rev.44