

**LOT** PX 015  
 Rev 1

**CONTROL**

(Exp.) 2015-03-05  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4	2.4		± 0.4	7.5	7.5	7.5	7.5		7.4	± 1.0	18.0	17.6	17.7		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.39	2.45	2.39	2.39	± 0.16	4.56	4.57	4.58	4.57	4.51	± 0.20	5.10	5.12	5.12	5.12	5.03	± 0.25	
HB	HGB	g/dl	7.0	7.0	7.0	7.0	6.8	± 0.4	13.5	13.5	13.5	13.5	13.3	± 0.5	16.1	16.2	16.1	16.2	15.9	± 0.6	
		g/l	70	70	70	70	68	± 4	135	135	135	135	133	± 5	161	162	161	162	159	± 6	
		mmol/l	4.35	4.35	4.35	4.35	4.22	± 0.25	8.38	8.38	8.38	8.38	8.26	± 0.31	10.00	10.06	10.00	10.06	9.87	± 0.37	
HT	HCT	%	19.8	19.8	19.6	19.8	19.2	± 1.5	36.9	37.0	36.2	37.0	36.7	± 2.0	44.4	44.5	43.5	44.5	43.9	± 2.5	
		l/l	0.198	0.198	0.196	0.198	0.192	± 0.015	0.369	0.370	0.362	0.370	0.367	± 0.020	0.444	0.445	0.435	0.445	0.439	± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	81	83	80	83	80.3	± 5	81	81	79	81	81.3	± 5	87	87	85	87	87.3	± 5	
TGMH	MCH	pg	28.7	29.3	28.6	29.3	28.5	± 2.0	29.6	29.5	29.5	29.5	29.5	± 2.0	31.6	31.6	31.4	31.6	31.6	± 2.5	
		fmol	1.78	1.82	1.77	1.82	1.77	± 0.12	1.84	1.83	1.83	1.83	1.83	± 0.12	1.96	1.96	1.95	1.96	1.96	± 0.16	
CCMH	MCHC	g/dl	35.4	35.3	35.7	35.3	35.4	± 3.0	36.5	36.5	37.3	36.5	36.3	± 3.0	36.3	36.4	37.0	36.4	36.2	± 3.0	
		g/l	354	353	357	353	354	± 30	365	365	373	365	363	± 30	363	364	370	364	362	± 30	
		mmol/l	21.99	21.91	22.18	21.91	22.00	± 1.86	22.70	22.65	23.17	22.65	22.53	± 1.86	22.53	22.58	22.97	22.58	22.49	± 1.86	
IDR	RDW	%	12.3	13.3	11.8	13.3	10.5	± 4.0	12.0	13.5	12.0	13.5	11.2	± 4.0	11.5	13.5	11.3	13.5	11.1	± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	76	74	74	74	73	± 20	248	245	248	245	252	± 30	470	470	485	470	483	± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.8	10.2	10.0	10.2	9.1	± 2.0	9.7	10.0	9.8	10.0	8.9	± 2.0	8.4	8.7	8.5	8.7	7.7	± 2.0	
NEUT	#	%	1.39	1.45	1.43	1.45	1.46	± 0.35	4.22	4.43	4.35	4.43	4.40	± 0.90	12.69	12.58	12.66	12.58	13.19	± 1.90	
		%	58.0	60.5	59.5	60.5	61.3	± 10.0	56.3	59.0	58.0	59.0	59.7	± 10.0	70.5	71.5	71.5	71.5	73.8	± 10.0	
LYMPHO	#	%	0.65	0.60	0.65	0.60	0.58	± 0.31	2.36	2.25	2.29	2.25	2.12	± 0.70	2.70	2.64	2.57	2.64	2.23	± 1.50	
		%	27.0	25.0	27.0	25.0	24.1	± 12.0	31.5	30.0	30.5	30.0	28.8	± 8.0	15.0	15.0	14.5	15.0	12.5	± 8.0	
MONO	#	%	0.11	0.08	0.07	0.08	0.10	± 0.07	0.32	0.23	0.26	0.23	0.26	± 0.23	0.81	0.53	0.62	0.53	0.63	± 0.53	
		%	4.5	3.5	3.0	3.5	4.0	± 3.0	4.2	3.0	3.5	3.0	3.5	± 3.0	4.5	3.0	3.5	3.0	3.5	± 3.0	
EOS	#	%	0.17	0.18	0.17	0.18	0.17	± 0.17	0.34	0.34	0.34	0.34	0.34	± 0.34	0.99	1.06	1.06	1.06	1.02	± 0.99	
		%	7.0	7.5	7.0	7.5	7.0	± 7.0	4.5	4.5	4.5	4.5	4.5	± 4.5	5.5	6.0	6.0	6.0	5.7	± 5.5	
BASO	#	%	0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.26	0.26	0.26	0.26	± 0.26	0.81	0.79	0.80	0.79	0.80	± 0.79	
		%	3.5	3.5	3.5	3.5	3.6	± 3.5	3.5	3.5	3.5	3.5	3.5	± 3.5	4.5	4.5	4.5	4.5	4.5	± 4.5	

**LOT** PX 015  
 Rev 1

**CONTROL**

(Exp.) 2015-03-05  
 (YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	8.0	8.0	8.0		± 1.0	19.8	19.8	19.8		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.45	2.45		± 0.16	4.58	4.58	4.58		± 0.20	5.15	5.15	5.15		± 0.25			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	13.4	13.4	13.4		± 0.5	16.1	16.1	16.1		± 0.6			
	g/l	70	70	70		± 4	134	134	134		± 5	161	161	161		± 6			
	mmol/l	4.35	4.35	4.35		± 0.25	8.32	8.32	8.32		± 0.31	10.00	10.00	10.00		± 0.37			
HT HCT	%	20.3	20.3	20.3		± 1.5	37.6	37.6	37.6		± 2.0	45.3	45.3	45.3		± 2.5			
	l/l	0.203	0.203	0.203		± 0.015	0.376	0.376	0.376		± 0.020	0.453	0.453	0.453		± 0.025			
VGM MCV	µm <sup>3</sup> ·fl	83	83	83		± 5	82	82	82		± 5	88	88	88		± 5			
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	29.3	29.3	29.3		± 2.0	31.3	31.3	31.3		± 2.5			
	fmol	1.77	1.77	1.77		± 0.12	1.82	1.82	1.82		± 0.12	1.94	1.94	1.94		± 0.16			
CCMH MCHC	g/dl	34.4	34.4	34.4		± 3.0	35.7	35.7	35.7		± 3.0	35.5	35.5	35.5		± 3.0			
	g/l	344	344	344		± 30	357	357	357		± 30	355	355	355		± 30			
	mmol/l	21.38	21.38	21.38		± 1.86	22.16	22.16	22.16		± 1.86	22.06	22.06	22.06		± 1.86			
IDR RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	14.8	14.8	14.8		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	78	78		± 20	255	255	255		± 30	465	465	465		± 50			
VPM MPV	µm <sup>3</sup> ·fl	9.9	9.9	9.9		± 2.0	9.7	9.7	9.7		± 2.0	8.4	8.4	8.4		± 2.0			
NEUT	#	1.53	1.58	1.58		± 0.35	4.74	4.88	4.88		± 0.90	14.50	14.70	14.70		± 1.90			
	%	61.2	63.2	63.2		± 10.0	59.2	61.0	61.0		± 10.0	72.9	74.2	74.2		± 10.0			
LYMPHO	#	0.63	0.57	0.57		± 0.31	2.27	2.10	2.10		± 0.70	2.59	2.34	2.34		± 1.50			
	%	25.1	22.6	22.6		± 12.0	28.4	26.2	26.2		± 8.0	13.1	11.8	11.8		± 8.0			
MONO	#	0.11	0.11	0.11		± 0.11	0.43	0.44	0.44		± 0.43	1.19	1.17	1.17		± 1.17			
	%	4.4	4.5	4.5		± 4.4	5.4	5.5	5.5		± 5.4	6.0	5.9	5.9		± 5.9			
EOS	#	0.17	0.18	0.18		± 0.17	0.32	0.34	0.34		± 0.32	1.09	1.11	1.11		± 1.09			
	%	6.8	7.2	7.2		± 6.8	4.0	4.3	4.3		± 4.0	5.5	5.6	5.6		± 5.5			
BASO	#	0.06	0.06	0.06		± 0.06	0.24	0.24	0.24		± 0.24	0.50	0.50	0.50		± 0.50			
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5			

9930091-A  
 Ref. TEMP-0821 Rev.35