

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT PX 053

CONTROL

(Exp.) **2013-07-05**
(YYYY-MM-DD)

Rev 3

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE	
		60 60C+ ES60	80 XL80	MS60	XLR			60 60C+ ES60	80 XL80	MS60	XLR			60 60C+ ES60	80 XL80	MS60	XLR		
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.5	2.5		± 0.4	7.4	7.4	7.5	7.4		± 1.0	17.2	17.2	17.2	17.2		± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.44	2.47	2.44		± 0.12	4.56	4.57	4.57	4.57		± 0.15	5.02	5.05	5.04	5.05		± 0.20
HB HGB	g/dl	7.0	6.9	6.9	6.9		± 0.4	13.7	13.8	13.8	13.8		± 0.5	16.0	16.1	16.0	16.1		± 0.6
	g/l	70	69	69	69		± 4	137	138	138	138		± 5	160	161	160	161		± 6
	mmol/l	4.35	4.28	4.28	4.28		± 0.25	8.51	8.57	8.57	8.57		± 0.31	9.94	10.00	9.94	10.00		± 0.37
HT HCT	%	19.8	20.3	19.5	20.3		± 1.5	37.8	38.4	37.0	38.4		± 2.0	43.7	43.9	42.8	43.9		± 2.5
	l/l	0.198	0.203	0.195	0.203		± 0.015	0.378	0.384	0.370	0.384		± 0.020	0.437	0.439	0.428	0.439		± 0.025
VGM MCV	µm ³ ; fl	80	83	79	83		± 5	83	84	81	84		± 5	87	87	85	87		± 5
TGMH MCH	pg	28.2	28.3	27.9	28.3		± 2.0	30.0	30.2	30.2	30.2		± 2.0	31.9	31.9	31.7	31.9		± 2.5
	fmol	1.75	1.76	1.73	1.76		± 0.12	1.87	1.88	1.88	1.88		± 0.12	1.98	1.98	1.97	1.98		± 0.16
	g/dl	35.3	34.1	35.4	34.1		± 3.0	36.2	35.9	37.3	35.9		± 3.0	36.6	36.6	37.3	36.6		± 3.0
CCMH MCHC	g/l	353	341	354	341		± 30	362	359	373	359		± 30	366	366	373	366		± 30
	mmol/l	21.91	21.16	21.96	21.16		± 1.86	22.48	22.32	23.15	22.32		± 1.86	22.75	22.76	23.19	22.76		± 1.86
IDR RDW	%	13.2	13.6	12.5	13.6		± 4.0	12.5	13.6	12.1	13.6		± 4.0	12.3	13.2	12.0	13.2		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	70	68	70		± 20	258	260	260	260		± 30	505	515	520	515		± 50
VPM MPV	µm ³ ; fl	9.2	9.3	9.1	9.3		± 2.0	8.9	9.1	9.1	9.1		± 2.0	8.8	9.0	8.9	9.0		± 2.0
NEUT	#	1.49	1.60	1.43	1.60		± 0.30	4.37	4.40	4.31	4.40		± 0.80	11.87	11.78	11.75	11.78		± 1.90
	%	62.0	64.0	57.0	64.0		± 10.0	59.0	59.5	57.5	59.5		± 10.0	69.0	68.5	68.3	68.5		± 10.0
LYMPHO	#	0.56	0.55	0.69	0.55		± 0.30	2.18	2.11	2.29	2.11		± 0.70	2.92	2.92	3.06	2.92		± 1.50
	%	23.5	22.0	27.5	22.0		± 9.0	29.5	28.5	30.5	28.5		± 8.0	17.0	17.0	17.8	17.0		± 8.0
MONO	#	0.13	0.11	0.16	0.11		± 0.11	0.33	0.33	0.38	0.33		± 0.33	0.77	0.77	0.77	0.77		± 0.77
	%	5.5	4.5	6.5	4.5		± 4.5	4.5	4.5	5.0	4.5		± 4.5	4.5	4.5	4.5	4.5		± 4.5
EOS	#	0.13	0.15	0.14	0.15		± 0.13	0.27	0.30	0.27	0.30		± 0.27	0.86	0.95	0.86	0.95		± 0.82
	%	5.5	6.0	5.5	6.0		± 5.0	3.5	4.0	3.5	4.0		± 3.5	5.0	5.5	5.0	5.5		± 4.5
BASO	#	0.08	0.09	0.09	0.09		± 0.05	0.26	0.26	0.26	0.26		± 0.16	0.77	0.77	0.76	0.77		± 0.36
	%	3.5	3.5	3.5	3.5		± 2.0	3.5	3.5	3.5	3.5		± 2.0	4.5	4.5	4.4	4.5		± 2.0

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT PX 053

CONTROL

(Exp.) 2013-07-05
(YYYY-MM-DD)

Rev 3

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8		± 1.0	18.2	18.2	18.2		± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.48	2.48		± 0.12	4.60	4.60	4.60		± 0.15	5.14	5.14	5.14		± 0.20
HB HGB	g/dl	7.1	7.1	7.1		± 0.4	13.8	13.8	13.8		± 0.5	16.0	16.0	16.0		± 0.6
	g/l	71	71	71		± 4	138	138	138		± 5	160	160	160		± 6
HT HCT	mmol/l	4.41	4.41	4.41		± 0.25	8.57	8.57	8.57		± 0.31	9.94	9.94	9.94		± 0.37
	%	20.3	20.3	20.3		± 1.5	38.6	38.6	38.6		± 2.0	44.7	44.7	44.7		± 2.5
VGM MCV	l/l	0.203	0.203	0.203		± 0.015	0.386	0.386	0.386		± 0.020	0.447	0.447	0.447		± 0.025
	µm ³ ·fl	82	82	82		± 5	84	84	84		± 5	87	87	87		± 5
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	30.0	30.0	30.0		± 2.0	31.1	31.1	31.1		± 2.5
	fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86		± 0.12	1.93	1.93	1.93		± 0.16
CCMH MCHC	g/dl	34.9	34.9	34.9		± 3.0	35.7	35.7	35.7		± 3.0	35.8	35.8	35.8		± 3.0
	g/l	349	349	349		± 30	357	357	357		± 30	358	358	358		± 30
	mmol/l	21.68	21.68	21.68		± 1.86	22.18	22.18	22.18		± 1.86	22.22	22.22	22.22		± 1.86
IDR RDW	%	16.0	16.0	16.0		± 4.0	16.0	16.0	16.0		± 4.0	15.2	15.2	15.2		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	74	74		± 20	258	258	258		± 30	508	508	508		± 50
VPM MPV	µm ³ ·fl	9.1	9.1	9.1		± 2.0	9.0	9.0	9.0		± 2.0	8.9	8.9	8.9		± 2.0
NEUT	#	1.60	1.55	1.55		± 0.30	4.60	4.60	4.60		± 0.80	12.69	12.65	12.65		± 1.90
	%	64.0	62.0	62.0		± 10.0	59.0	59.0	59.0		± 10.0	69.7	69.5	69.5		± 10.0
LYMPHO	#	0.60	0.56	0.56		± 0.30	2.34	2.18	2.18		± 0.70	3.24	3.00	3.00		± 1.50
	%	24.0	22.5	22.5		± 9.0	30.0	28.0	28.0		± 8.0	17.8	16.5	16.5		± 8.0
MONO	#	0.11	0.19	0.19		± 0.11	0.39	0.51	0.51		± 0.33	1.00	1.18	1.18		± 0.77
	%	4.5	7.5	7.5		± 4.5	5.0	6.5	6.5		± 4.5	5.5	6.5	6.5		± 4.5
EOS	#	0.13	0.13	0.13		± 0.13	0.27	0.27	0.27		± 0.27	0.82	0.82	0.82		± 0.82
	%	5.0	5.0	5.0		± 5.0	3.5	3.5	3.5		± 3.5	4.5	4.5	4.5		± 4.5
BASO	#	0.06	0.08	0.08		± 0.05	0.20	0.23	0.23		± 0.16	0.46	0.55	0.55		± 0.36
	%	2.5	3.0	3.0		± 2.0	2.5	3.0	3.0		± 2.0	2.5	3.0	3.0		± 2.0

Ref: TEMP-0821 Rev.32 BACK / VERSO 9930080-A