

**LOT PX 014**  
 Rev 2

**CONTROL**

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																
			CONTROL				L	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA					PENTRA						PENTRA					
			60 60C+ ES60	80 XL80	MS60	XLR		60 60C+ ES60	80 XL80	MS60	XLR			60 60C+ ES60	80 XL80	MS60	XLR		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.3	2.3	2.3	2.3	± 0.4	7.2	7.2	7.2	7.2	± 1.0	17.6	17.2	17.3	17.2	± 2.2		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.30	2.34	2.30	± 0.12	4.38	4.35	4.37	4.35	± 0.15	5.20	5.20	5.17	5.20	± 0.20		
HB	HGB	g/dl	6.5	6.5	6.5	6.5	± 0.4	13.1	13.1	13.1	13.1	± 0.5	16.8	16.8	16.7	16.8	± 0.6		
		g/l	65	65	65	65	± 4	131	131	131	131	± 5	168	168	167	168	± 6		
		mmol/l	4.04	4.04	4.04	4.04	± 0.25	8.14	8.14	8.14	8.14	± 0.31	10.43	10.43	10.37	10.43	± 0.37		
HT	HCT	%	18.6	18.9	18.3	18.9	± 1.5	35.9	35.7	35.0	35.7	± 2.0	45.8	45.8	44.5	45.8	± 2.5		
		l/l	0.186	0.189	0.183	0.189	± 0.015	0.359	0.357	0.350	0.357	± 0.020	0.458	0.458	0.445	0.458	± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	79	82	78	82	± 5	82	82	80	82	± 5	88	88	86	88	± 5		
TGMH	MCH	pg	27.7	28.3	27.8	28.3	± 2.0	29.9	30.1	30.0	30.1	± 2.0	32.3	32.3	32.3	32.3	± 2.5		
		fmol	1.72	1.76	1.73	1.76	± 0.12	1.86	1.87	1.86	1.87	± 0.12	2.01	2.01	2.01	2.01	± 0.16		
CCMH	MCHC	g/dl	35.0	34.5	35.6	34.5	± 3.0	36.5	36.7	37.5	36.7	± 3.0	36.7	36.7	37.6	36.7	± 3.0		
		g/l	350	345	356	345	± 30	365	367	375	367	± 30	367	367	376	367	± 30		
		mmol/l	21.74	21.40	22.12	21.40	± 1.86	22.65	22.81	23.27	22.81	± 1.86	22.80	22.80	23.32	22.80	± 1.86		
IDR	RDW	%	12.7	13.3	12.5	13.3	± 4.0	12.0	13.0	12.0	13.0	± 4.0	11.6	13.0	11.5	13.0	± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	72	72	72	± 20	258	260	265	260	± 30	510	520	535	520	± 50		
VPM	MPV	µm <sup>3</sup> ; fl	9.2	9.4	9.2	9.4	± 2.0	8.6	8.8	8.6	8.8	± 2.0	8.9	9.1	9.1	9.1	± 2.0		
NEUT		#	1.27	1.27	1.29	1.27	± 0.25	4.14	4.18	4.18	4.18	± 0.80	12.32	12.04	12.20	12.04	± 1.90		
		%	55.0	55.3	56.0	55.3	± 10.0	57.5	58.0	58.0	58.0	± 10.0	70.0	70.0	70.5	70.0	± 10.0		
LYMPHO		#	0.70	0.70	0.69	0.70	± 0.20	2.23	2.20	2.23	2.20	± 0.60	2.90	2.84	2.77	2.84	± 1.50		
		%	30.5	30.5	30.0	30.5	± 8.0	31.0	30.5	31.0	30.5	± 8.0	16.5	16.5	16.0	16.5	± 8.0		
MONO		#	0.12	0.10	0.10	0.10	± 0.10	0.32	0.29	0.29	0.29	± 0.29	0.79	0.69	0.69	0.69	± 0.69		
		%	5.0	4.5	4.5	4.5	± 4.5	4.5	4.0	4.0	4.0	± 4.0	4.5	4.0	4.0	4.0	± 4.0		
EOS		#	0.15	0.15	0.15	0.15	± 0.12	0.25	0.29	0.25	0.29	± 0.21	0.88	0.86	0.95	0.86	± 0.73		
		%	6.5	6.5	6.5	6.5	± 5.0	3.5	4.0	3.5	4.0	± 2.8	5.0	5.0	5.5	5.0	± 4.0		
BASO		#	0.07	0.07	0.07	0.07	± 0.07	0.25	0.25	0.25	0.25	± 0.19	0.70	0.77	0.69	0.77	± 0.46		
		%	3.0	3.2	3.0	3.2	± 3.0	3.5	3.5	3.5	3.5	± 2.5	4.0	4.5	4.0	4.5	± 2.5		

**LOT PX 014**  
 Rev 2

**CONTROL**

(Exp.) **2014-03-05**  
 (YYYY - MM - DD)

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		120 120 RETIC	DX 120 DF 120	DX NEXUS DF NEXUS			120 120 RETIC	DX 120 DF 120	DX NEXUS DF NEXUS			120 120 RETIC	DX 120 DF 120	DX NEXUS DF NEXUS		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4		± 0.4	7.4	7.4	7.4		± 1.0	18.3	18.3	18.3		± 2.2
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35	2.35		± 0.12	4.38	4.38	4.38		± 0.15	5.26	5.26	5.26		± 0.20
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	13.1	13.1	13.1		± 0.5	16.6	16.6	16.6		± 0.6
	g/l	66	66	66		± 4	131	131	131		± 5	166	166	166		± 6
	mmol/l	4.10	4.10	4.10		± 0.25	8.14	8.14	8.14		± 0.31	10.31	10.31	10.31		± 0.37
HT HCT	%	19.0	19.0	19.0		± 1.5	36.4	36.4	36.4		± 2.0	46.8	46.8	46.8		± 2.5
	l/l	0.190	0.190	0.190		± 0.015	0.364	0.364	0.364		± 0.020	0.468	0.468	0.468		± 0.025
VGM MCV	µm <sup>3</sup> fl	81	81	81		± 5	83	83	83		± 5	89	89	89		± 5
TGMH MCH	pg	28.1	28.1	28.1		± 2.0	29.9	29.9	29.9		± 2.0	31.6	31.6	31.6		± 2.5
	fmol	1.74	1.74	1.74		± 0.12	1.86	1.86	1.86		± 0.12	1.96	1.96	1.96		± 0.16
CCMH MCHC	g/dl	34.7	34.7	34.7		± 3.0	36.0	36.0	36.0		± 3.0	35.5	35.5	35.5		± 3.0
	g/l	347	347	347		± 30	360	360	360		± 30	355	355	355		± 30
	mmol/l	21.53	21.53	21.53		± 1.86	22.38	22.38	22.38		± 1.86	22.02	22.02	22.02		± 1.86
IDR RDW	%	16.1	16.1	16.1		± 4.0	16.8	16.8	16.8		± 4.0	16.1	16.1	16.1		± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	75	75		± 20	268	268	268		± 30	525	525	525		± 50
VPM MPV	µm <sup>3</sup> fl	9.1	9.1	9.1		± 2.0	8.5	8.5	8.5		± 2.0	8.8	8.8	8.8		± 2.0
NEUT	#	1.25	1.38	1.38		± 0.25	4.26	4.43	4.43		± 0.80	12.90	13.20	13.20		± 1.90
	%	52.0	57.6	57.6		± 10.0	57.5	59.8	59.8		± 10.0	70.5	72.2	72.2		± 10.0
LYMPHO	#	0.78	0.65	0.65		± 0.35	2.31	2.09	2.09		± 0.60	3.02	2.69	2.69		± 1.50
	%	32.6	26.9	26.9		± 15.0	31.2	28.3	28.3		± 8.0	16.5	14.7	14.7		± 8.0
MONO	#	0.16	0.13	0.13		± 0.10	0.44	0.39	0.39		± 0.29	1.19	1.08	1.08		± 0.69
	%	6.8	5.4	5.4		± 4.5	6.0	5.3	5.3		± 4.0	6.5	5.9	5.9		± 4.0
EOS	#	0.13	0.17	0.17		± 0.13	0.21	0.27	0.27		± 0.21	0.73	0.86	0.86		± 0.73
	%	5.6	7.1	7.1		± 5.6	2.8	3.6	3.6		± 2.8	4.0	4.7	4.7		± 4.0
BASO	#	0.07	0.07	0.07		± 0.07	0.19	0.22	0.22		± 0.19	0.46	0.46	0.46		± 0.46
	%	3.0	3.0	3.0		± 3.0	2.5	3.0	3.0		± 2.5	2.5	2.5	2.5		± 2.5

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B