

**LOT** PX 074  
 Rev 2

**CONTROL**

(Exp.) 2014-09-05  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.5	2.4	2.5	2.4		± 0.4	7.2	7.3	7.2	7.3		7.2	± 1.0	18.0	17.8	17.7		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.39	2.36	2.39	2.36	2.33	± 0.16	4.62	4.65	4.65	4.65	4.56	± 0.20	5.20	5.25	5.22	5.25	5.16	± 0.25	
HB	HGB	g/dl	6.7	6.8	6.8	6.8	6.7	± 0.4	13.7	13.8	13.8	13.8	13.6	± 0.5	16.5	16.6	16.5	16.6	16.4	± 0.6	
		g/l	67	68	68	68	67	± 4	137	138	138	138	136	± 5	165	166	165	166	164	± 6	
		mmol/l	4.16	4.22	4.22	4.22	4.16	± 0.25	8.51	8.57	8.57	8.57	8.45	± 0.31	10.25	10.31	10.25	10.31	10.18	± 0.37	
HT	HCT	%	18.9	19.4	18.6	19.4	18.7	± 1.5	37.0	37.7	36.7	37.7	37.2	± 2.0	45.2	45.7	44.9	45.7	45.0	± 2.5	
		l/l	0.189	0.194	0.186	0.194	0.187	± 0.015	0.370	0.377	0.367	0.377	0.372	± 0.020	0.452	0.457	0.449	0.457	0.450	± 0.025	
VGM	MCV	µm <sup>3</sup> ·fl	79	82	78	82	80.2	± 5	80	81	79	81	81.6	± 5	87	87	86	87	87.3	± 5	
TGMH	MCH	pg	28.0	28.8	28.5	28.8	28.8	± 2.0	29.7	29.7	29.7	29.7	29.8	± 2.0	31.7	31.6	31.6	31.6	31.8	± 2.5	
		fmol	1.74	1.79	1.77	1.79	1.79	± 0.12	1.84	1.84	1.84	1.84	1.85	± 0.12	1.97	1.96	1.96	1.96	1.97	± 0.16	
CCMH	MCHC	g/dl	35.5	35.1	36.5	35.1	35.9	± 3.0	37.1	36.6	37.6	36.6	36.5	± 3.0	36.5	36.3	36.8	36.3	36.4	± 3.0	
		g/l	355	351	365	351	359	± 30	371	366	376	366	365	± 30	365	363	368	363	364	± 30	
		mmol/l	22.04	21.82	22.65	21.82	22.27	± 1.86	23.02	22.75	23.33	22.75	22.70	± 1.86	22.65	22.57	22.82	22.57	22.61	± 1.86	
IDR	RDW	%	12.4	13.3	12.5	13.3	11.1	± 4.0	12.5	13.7	12.5	13.7	10.6	± 4.0	12.0	13.5	12.2	13.5	10.8	± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	76	72	76	75	± 20	258	260	256	260	267	± 30	500	510	510	510	514	± 50	
VPM	MPV	µm <sup>3</sup> ·fl	9.0	9.4	9.0	9.4	8.4	± 2.0	8.7	9.2	8.8	9.2	8.1	± 2.0	8.7	9.2	8.7	9.2	8.1	± 2.0	
NEUT	#		1.36	1.40	1.38	1.40	1.40	± 0.35	4.07	4.12	4.10	4.12	4.15	± 0.90	12.73	12.37	12.58	12.37	12.71	± 1.90	
		%	56.5	56.0	57.5	56.0	58.4	± 10.0	56.5	56.5	57.0	56.5	57.7	± 10.0	70.7	69.5	71.1	69.5	70.6	± 10.0	
LYMPHO	#		0.74	0.79	0.73	0.79	0.71	± 0.40	2.38	2.41	2.38	2.41	2.31	± 0.70	2.88	2.85	2.87	2.85	2.77	± 1.50	
		%	31.0	31.5	30.5	31.5	29.7	± 12.0	33.0	33.0	33.0	33.0	32.1	± 8.0	16.0	16.0	16.2	16.0	15.4	± 8.0	
MONO	#		0.08	0.09	0.07	0.09	0.08	± 0.07	0.25	0.22	0.22	0.22	0.25	± 0.22	0.72	0.62	0.62	0.62	0.72	± 0.62	
		%	3.5	3.5	3.0	3.5	3.5	± 3.0	3.5	3.0	3.0	3.0	3.5	± 3.0	4.0	3.5	3.5	3.5	4.0	± 3.5	
EOS	#		0.14	0.15	0.14	0.15	0.12	± 0.12	0.25	0.29	0.25	0.29	0.24	± 0.24	0.86	1.16	0.85	1.16	1.04	± 0.85	
		%	5.8	6.0	5.8	6.0	5.1	± 5.1	3.5	4.0	3.5	4.0	3.3	± 3.3	4.8	6.5	4.8	6.5	5.8	± 4.8	
BASO	#		0.08	0.08	0.08	0.08	0.08	± 0.08	0.25	0.26	0.25	0.26	0.24	± 0.24	0.81	0.80	0.78	0.80	0.76	± 0.76	
		%	3.2	3.0	3.2	3.0	3.3	± 3.0	3.5	3.5	3.5	3.5	3.4	± 3.4	4.5	4.5	4.4	4.5	4.2	± 4.2	

**LOT** PX 074  
 Rev 2

**CONTROL**

(Exp.) 2014-09-05  
 (YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.5	7.5	7.5		± 1.0	18.8	18.8	18.8		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40	2.40	2.40		± 0.16	4.67	4.67	4.67		± 0.20	5.33	5.33	5.33		± 0.25			
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.7	13.7	13.7		± 0.5	16.4	16.4	16.4		± 0.6			
	g/l	69	69	69		± 4	137	137	137		± 5	164	164	164		± 6			
HT HCT	mmol/l	4.28	4.28	4.28		± 0.25	8.51	8.51	8.51		± 0.31	10.18	10.18	10.18		± 0.37			
	%	19.9	19.9	19.9		± 1.5	38.3	38.3	38.3		± 2.0	46.9	46.9	46.9		± 2.5			
VGM MCV	l/l	0.199	0.199	0.199		± 0.015	0.383	0.383	0.383		± 0.020	0.469	0.469	0.469		± 0.025			
	µm <sup>3</sup> ·fl	83	83	83		± 5	82	82	82		± 5	88	88	88		± 5			
TGMH MCH	pg	28.8	28.8	28.8		± 2.0	29.3	29.3	29.3		± 2.0	30.8	30.8	30.8		± 2.5			
	fmol	1.79	1.79	1.79		± 0.12	1.82	1.82	1.82		± 0.12	1.91	1.91	1.91		± 0.16			
CCMH MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.8	35.8	35.8		± 3.0	35.0	35.0	35.0		± 3.0			
	g/l	346	346	346		± 30	358	358	358		± 30	350	350	350		± 30			
IDR RDW	mmol/l	21.51	21.51	21.51		± 1.86	22.22	22.22	22.22		± 1.86	21.71	21.71	21.71		± 1.86			
	%	15.4	15.4	15.4		± 4.0	15.8	15.8	15.8		± 4.0	15.0	15.0	15.0		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	78	78		± 20	265	265	265		± 30	515	515	515		± 50			
VPM MPV	µm <sup>3</sup> ·fl	9.1	9.1	9.1		± 2.0	8.9	8.9	8.9		± 2.0	8.7	8.7	8.7		± 2.0			
NEUT	#	1.49	1.49	1.49		± 0.35	4.31	4.33	4.33		± 0.90	13.60	13.60	13.60		± 1.90			
	%	59.5	59.5	59.5		± 10.0	57.5	57.7	57.7		± 10.0	72.1	72.3	72.3		± 10.0			
LYMPHO	#	0.70	0.70	0.70		± 0.40	2.45	2.45	2.45		± 0.70	3.10	3.10	3.10		± 1.50			
	%	27.9	27.9	27.9		± 12.0	32.7	32.7	32.7		± 8.0	16.5	16.5	16.5		± 8.0			
MONO	#	0.10	0.10	0.10		± 0.10	0.29	0.29	0.29		± 0.29	0.90	0.90	0.90		± 0.90			
	%	4.0	4.0	4.0		± 4.0	3.8	3.8	3.8		± 3.8	4.8	4.8	4.8		± 4.8			
EOS	#	0.14	0.14	0.14		± 0.14	0.21	0.21	0.21		± 0.21	0.71	0.71	0.71		± 0.71			
	%	5.6	5.6	5.6		± 5.6	2.8	2.8	2.8		± 2.8	3.8	3.8	3.8		± 3.8			
BASO	#	0.08	0.08	0.08		± 0.08	0.24	0.23	0.23		± 0.23	0.53	0.49	0.49		± 0.49			
	%	3.0	3.0	3.0		± 3.0	3.2	3.0	3.0		± 3.0	2.8	2.6	2.6		± 2.6			

9930091-A  
 Ref. TEMP-0821 Rev.35