


LOT PX 408

CONTROL

 (Exp.) 2018-01-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE	
			CONTROL				L	CONTROL				N	CONTROL				H			
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE			
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS					
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS										
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7	2.7			± 0.4	8.3	8.3	8.3			± 1.0	19.9	19.9	19.9			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.29	2.29			± 0.16	4.47	4.47	4.47			± 0.20	5.12	5.12	5.12			± 0.25
HB	HGB	g/dl	6.6	6.6	6.6			± 0.4	12.7	12.7	12.7			± 0.5	15.7	15.7	15.7			± 0.6
		g/l	66	66	66			± 4	127	127	127			± 5	157	157	157			± 6
		mmol/l	4.10	4.10	4.10			± 0.25	7.89	7.89	7.89			± 0.31	9.75	9.75	9.75			± 0.37
HT	HCT	%	19.0	19.0	19.2			± 1.5	36.2	36.2	36.7			± 2.0	44.5	44.5	45.1			± 2.5
		l/l	0.190	0.190	0.192			± 0.015	0.362	0.362	0.367			± 0.020	0.445	0.445	0.451			± 0.025
VGM	MCV	µm ³ ; fl	83	83	84			± 5	81	81	82			± 5	87	87	88			± 5
TGMH	MCH	pg	28.8	28.8	28.8			± 2.0	28.4	28.4	28.4			± 2.0	30.7	30.7	30.7			± 2.5
		fmol	1.79	1.79	1.79			± 0.12	1.76	1.76	1.76			± 0.12	1.90	1.90	1.90			± 0.16
CCMH	MCHC	g/dl	34.7	34.7	34.3			± 3.0	35.1	35.1	34.6			± 3.0	35.2	35.2	34.8			± 3.0
		g/l	347	347	343			± 30	351	351	346			± 30	352	352	348			± 30
		mmol/l	21.56	21.56	21.31			± 1.86	21.78	21.78	21.52			± 1.86	21.89	21.89	21.64			± 1.86
IDR	RDW	%	15.8	15.8	15.8			± 4.0	16.2	16.2	16.2			± 4.0	15.4	15.4	15.4			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78	78	78			± 20	250	250	250			± 30	490	490	490			± 50
VPM	MPV	µm ³ ; fl	10.1	10.1	10.1			± 2.0	9.8	9.8	9.8			± 2.0	9.0	9.0	9.0			± 2.0
NEUT	#	%	1.50	1.56	1.56			± 0.35	4.66	4.80	4.80			± 0.90	14.10	14.30	14.30			± 1.90
		%	55.5	57.8	57.8			± 10.0	56.2	57.8	57.8			± 10.0	70.6	71.7	71.7			± 10.0
LYMPHO	#	%	0.81	0.73	0.73			± 0.33	2.58	2.41	2.41			± 0.70	2.87	2.65	2.65			± 1.50
		%	29.9	27.2	27.2			± 12.0	31.1	29.0	29.0			± 8.0	14.4	13.3	13.3			± 8.0
MONO	#	%	0.18	0.19	0.19			± 0.18	0.54	0.56	0.56			± 0.54	1.51	1.49	1.49			± 1.49
		%	6.8	6.9	6.9			± 6.8	6.5	6.7	6.7			± 6.5	7.6	7.5	7.5			± 7.5
EOS	#	%	0.14	0.15	0.15			± 0.14	0.27	0.29	0.29			± 0.27	0.98	1.00	1.00			± 0.98
		%	5.3	5.6	5.6			± 5.3	3.2	3.5	3.5			± 3.2	4.9	5.0	5.0			± 4.9
BASO	#	%	0.07	0.07	0.07			± 0.07	0.25	0.25	0.25			± 0.25	0.50	0.50	0.50			± 0.50
		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

Ref: TEMP-0821 Rev.41 BACK / VERSO 1300040596