

ABX Difftrol



LOT PX 409
Rev 1

CONTROL

(Exp.) 2018-03-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6	± 0.4	7.7	7.7	7.7	± 1.0	19.0	19.0	19.0	± 2.2				
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.33	2.33	2.33	± 0.16	4.58	4.58	4.58	± 0.20	5.18	5.18	5.18	± 0.25				
HB	HGB	g/dl	6.8	6.8	6.8	± 0.4	13.4	13.4	13.4	± 0.5	16.4	16.4	16.4	± 0.6				
		g/l	68	68	68	± 4	134	134	134	± 5	164	164	164	± 6				
		mmol/l	4.22	4.22	4.22	± 0.25	8.32	8.32	8.32	± 0.31	10.18	10.18	10.18	± 0.37				
HT	HCT	%	19.1	19.1	19.3	± 1.5	37.1	37.1	37.6	± 2.0	45.1	45.1	45.6	± 2.5				
		l/l	0.191	0.191	0.193	± 0.015	0.371	0.371	0.376	± 0.020	0.451	0.451	0.456	± 0.025				
VGM	MCV	µm ³ ; fl	82	82	83	± 5	81	81	82	± 5	87	87	88	± 5				
TGMH	MCH	pg	29.2	29.2	29.2	± 2.0	29.3	29.3	29.3	± 2.0	31.7	31.7	31.7	± 2.5				
		fmol	1.81	1.81	1.81	± 0.12	1.82	1.82	1.82	± 0.12	1.97	1.97	1.97	± 0.16				
CCMH	MCHC	g/dl	35.6	35.6	35.2	± 3.0	36.1	36.1	35.7	± 3.0	36.4	36.4	36.0	± 3.0				
		g/l	356	356	352	± 30	361	361	357	± 30	364	364	360	± 30				
		mmol/l	22.10	22.10	21.84	± 1.86	22.43	22.43	22.16	± 1.86	22.60	22.60	22.34	± 1.86				
IDR	RDW	%	15.7	15.7	15.7	± 4.0	16.5	16.5	16.5	± 4.0	14.8	14.8	14.8	± 4.0				
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70	± 20	250	250	250	± 30	492	492	492	± 50				
VPM	MPV	µm ³ ; fl	9.4	9.4	9.4	± 2.0	9.3	9.3	9.3	± 2.0	8.6	8.6	8.6	± 2.0				
NEUT		#	1.47	1.53	1.53	± 0.35	4.32	4.44	4.44	± 0.90	13.40	13.60	13.60	± 1.90				
		%	56.7	58.9	58.9	± 10.0	56.1	57.6	57.6	± 10.0	70.1	71.4	71.4	± 10.0				
LYMPHO		#	0.75	0.68	0.68	± 0.33	2.33	2.18	2.18	± 0.70	2.98	2.76	2.76	± 1.50				
		%	28.7	26.1	26.1	± 12.0	30.3	28.3	28.3	± 8.0	15.7	14.5	14.5	± 8.0				
MONO		#	0.20	0.20	0.20	± 0.20	0.60	0.62	0.62	± 0.60	1.41	1.39	1.39	± 1.39				
		%	7.6	7.8	7.8	± 7.6	7.8	8.1	8.1	± 7.8	7.4	7.3	7.3	± 7.3				
EOS		#	0.12	0.12	0.12	± 0.12	0.22	0.23	0.23	± 0.22	0.82	0.82	0.82	± 0.82				
		%	4.5	4.7	4.7	± 4.5	2.8	3.0	3.0	± 2.8	4.3	4.3	4.3	± 4.3				
BASO		#	0.07	0.07	0.07	± 0.07	0.23	0.23	0.23	± 0.23	0.48	0.48	0.48	± 0.48				
		%	2.5	2.5	2.5	± 2.5	3.0	3.0	3.0	± 3.0	2.5	2.5	2.5	± 2.5				

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596