


LOT PX 430

CONTROL

 (Exp.) 2021-09-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9			± 0.4	8.0	8.0			± 1.0	18.7	18.7			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.17	2.17			± 0.16	4.62	4.62			± 0.20	5.26	5.26			± 0.25			
HB HGB	g/dl	5.7	5.7			± 0.4	13.5	13.5			± 0.5	16.4	16.4			± 0.6			
	g/l	57	57			± 4	135	135			± 5	164	164			± 6			
	mmol/l	3.54	3.54			± 0.25	8.38	8.38			± 0.31	10.18	10.18			± 0.37			
HT HCT	%	16.3	16.3			± 1.5	37.0	37.0			± 2.0	45.8	45.8			± 2.5			
	l/l	0.163	0.163			± 0.015	0.370	0.370			± 0.020	0.458	0.458			± 0.025			
VGM MCV	µm ³ ; fl	75	75			± 5	80	80			± 5	87	87			± 5			
TGMH MCH	pg	26.3	26.3			± 2.0	29.2	29.2			± 2.0	31.2	31.2			± 2.5			
	fmol	1.63	1.63			± 0.12	1.81	1.81			± 0.12	1.94	1.94			± 0.16			
CCMH MCHC	g/dl	35.0	35.0			± 3.0	36.5	36.5			± 3.0	35.8	35.8			± 3.0			
	g/l	350	350			± 30	365	365			± 30	358	358			± 30			
	mmol/l	21.75	21.75			± 1.86	22.68	22.68			± 1.86	22.26	22.26			± 1.86			
IDR RDW	%	18.0	18.0			± 4.0	17.0	17.0			± 4.0	14.5	14.5			± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	67	67			± 20	248	248			± 30	483	483			± 50			
VPM MPV	µm ³ ; fl	10.2	10.2			± 2.0	9.9	9.9			± 2.0	9.0	9.0			± 2.0			
NEUT	#	1.50	1.50			± 0.35	4.39	4.39			± 0.90	12.90	12.90			± 1.90			
	%	51.6	51.6			± 10.0	54.9	54.9			± 10.0	69.2	69.2			± 10.0			
LYMPHO	#	0.77	0.77			± 0.33	2.35	2.35			± 0.70	2.84	2.84			± 1.50			
	%	26.4	26.4			± 12.0	29.4	29.4			± 8.0	15.2	15.2			± 8.0			
MONO	#	0.26	0.26			± 0.26	0.51	0.51			± 0.51	1.16	1.16			± 1.16			
	%	9.0	9.0			± 9.0	6.4	6.4			± 6.4	6.2	6.2			± 6.2			
EOS	#	0.23	0.23			± 0.23	0.46	0.46			± 0.46	0.92	0.92			± 0.92			
	%	8.0	8.0			± 8.0	5.7	5.7			± 5.7	4.9	4.9			± 4.9			
BASO	#	0.15	0.15			± 0.15	0.29	0.29			± 0.29	0.84	0.84			± 0.84			
	%	5.0	5.0			± 5.0	3.6	3.6			± 3.6	4.5	4.5			± 4.5			

1300090306
BACK / VERSO
Ref: TEMP-0821 Rev.45