

LOT		PX 033		CONTROL				(Exp.) 2013-05-05										
		Rev 2						(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse																
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4		± 0.4	7.4	7.4				± 1.0	17.7	17.5				± 2.2
GR	RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.38		± 0.12	4.62	4.60				± 0.15	5.07	5.09				± 0.20
HB	HGB	g/dl	6.7	6.7		± 0.4	13.6	13.6				± 0.5	15.8	15.8				± 0.6
		g/l	67	67		± 4	136	136				± 5	158	158				± 6
		mmol/l	4.16	4.16		± 0.25	8.45	8.45				± 0.31	9.81	9.81				± 0.37
HT	HCT	%	19.4	19.3		± 1.5	37.9	38.2				± 2.0	44.6	44.8				± 2.5
		l/l	0.194	0.193		± 0.015	0.379	0.382				± 0.020	0.446	0.448				± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	79	81		± 5	82	83				± 5	88	88				± 5
TGMH	MCH	pg	27.3	28.2		± 2.0	29.4	29.6				± 2.0	31.2	31.0				± 2.5
		fmol	1.70	1.75		± 0.12	1.83	1.84				± 0.12	1.94	1.93				± 0.16
		g/dl	34.6	34.8		± 3.0	35.9	35.6				± 3.0	35.4	35.3				± 3.0
CCMH	MCHC	g/l	346	348		± 30	359	356				± 30	354	353				± 30
		mmol/l	21.50	21.58		± 1.86	22.29	22.12				± 1.86	21.99	21.91				± 1.86
		%	13.5	13.2		± 4.0	13.5	13.8				± 4.0	13.0	13.7				± 4.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72		± 20	255	255				± 30	505	505				± 50
VPM	MPV	µm <sup>3</sup> ; fl	8.9	8.9		± 2.0	9.1	9.2				± 2.0	8.6	8.8				± 2.0
NEUT	#		1.37	1.36		± 0.30	4.04	4.03				± 0.80	12.16	11.73				± 1.80
		%	57.0	56.5		± 10.0	54.6	54.5				± 10.0	68.7	67.0				± 10.0
LYMPHO	#		0.71	0.72		± 0.20	2.44	2.44				± 0.60	3.10	3.41				± 1.50
		%	29.5	30.0		± 8.0	33.0	33.0				± 8.0	17.5	19.5				± 8.0
MONO	#		0.12	0.12		± 0.12	0.39	0.39				± 0.39	0.80	0.79				± 0.75
		%	5.0	5.0		± 5.0	5.0	5.0				± 5.0	4.5	4.5				± 4.0
EOS	#		0.13	0.13		± 0.13	0.31	0.33				± 0.31	0.89	0.88				± 0.74
		%	5.5	5.5		± 5.0	4.0	4.5				± 4.0	5.0	5.0				± 4.0
BASO	#		0.07	0.07		± 0.06	0.25	0.23				± 0.23	0.76	0.70				± 0.46
		%	3.0	3.0		± 2.5	3.4	3.0				± 3.0	4.3	4.0				± 2.5

LOT		PX 033		CONTROL				(Exp.) 2013-05-05										
		Rev 2						(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse																
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		120 120 RETIC	DX120 DF120			120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4		± 0.4	7.7	7.7				± 1.0	18.5	18.5				± 2.2
GR	RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.44		± 0.12	4.64	4.64				± 0.15	5.17	5.17				± 0.20
HB	HGB	g/dl	6.8	6.8		± 0.4	13.7	13.7				± 0.5	15.9	15.9				± 0.6
		g/l	68	68		± 4	137	137				± 5	159	159				± 6
		mmol/l	4.22	4.22		± 0.25	8.51	8.51				± 0.31	9.87	9.87				± 0.37
HT	HCT	%	19.8	19.8		± 1.5	39.0	39.0				± 2.0	45.5	45.5				± 2.5
		l/l	0.198	0.198		± 0.015	0.390	0.390				± 0.020	0.455	0.455				± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	81	81		± 5	84	84				± 5	88	88				± 5
TGMH	MCH	pg	27.9	27.9		± 2.0	29.5	29.5				± 2.0	30.8	30.8				± 2.5
		fmol	1.73	1.73		± 0.12	1.83	1.83				± 0.12	1.91	1.91				± 0.16
		g/dl	34.4	34.4		± 3.0	35.1	35.1				± 3.0	34.9	34.9				± 3.0
CCMH	MCHC	g/l	344	344		± 30	351	351				± 30	349	349				± 30
		mmol/l	21.37	21.37		± 1.86	21.83	21.83				± 1.86	21.70	21.70				± 1.86
		%	15.8	15.8		± 4.0	16.0	16.0				± 4.0	14.5	14.5				± 4.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72		± 20	260	260				± 30	505	505				± 50
VPM	MPV	µm <sup>3</sup> ; fl	8.9	8.9		± 2.0	9.1	9.1				± 2.0	8.6	8.6				± 2.0
NEUT	#		1.37	1.39		± 0.30	4.24	4.39				± 0.80	12.80	12.80				± 1.80
		%	57.0	58.0		± 10.0	55.0	57.0				± 10.0	69.0	69.0				± 10.0
LYMPHO	#		0.73	0.65		± 0.20	2.54	2.27				± 0.60	3.42	3.15				± 1.50
		%	30.5	27.0		± 8.0	33.0	29.5				± 8.0	18.5	17.0				± 8.0
MONO	#		0.12	0.16		± 0.12	0.39	0.46				± 0.39	1.11	1.30				± 0.75
		%	5.0	6.5		± 5.0	5.0	6.0				± 5.0	6.0	7.0				± 4.0
EOS	#		0.13	0.13		± 0.13	0.31	0.31				± 0.31	0.74	0.74				± 0.74
		%	5.0	5.5		± 5.0	4.0	4.0				± 4.0	4.0	4.0				± 4.0
BASO	#		0.06	0.07		± 0.06	0.23	0.27				± 0.23	0.46	0.56				± 0.46
		%	2.5	3.0		± 2.5	3.0	3.5				± 3.0	2.5	3.0				± 2.5

Ref: TEMP-0821 Rev.30 FRONT / RECTO 9930073-A