

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 073

CONTROL

(Exp.) **2013-09-05**
 (YYYY-MM-DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	80 XL80			60 60C+ ES60		80 XL80			60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.3	7.3				± 1.0	17.8	17.5			± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.37			± 0.12	4.58	4.58				± 0.15	5.10	5.11			± 0.20		
HB HGB	g/dl	6.8	6.9			± 0.4	13.2	13.3				± 0.5	16.0	16.0			± 0.6		
	g/l	68	69			± 4	132	133				± 5	160	160			± 6		
	mmol/l	4.22	4.28			± 0.25	8.20	8.26				± 0.31	9.94	9.94			± 0.37		
HT HCT	%	19.1	19.7			± 1.5	36.2	36.6				± 2.0	44.4	44.5			± 2.5		
	l/l	0.191	0.197			± 0.015	0.362	0.366				± 0.020	0.444	0.445			± 0.025		
VGM MCV	µm ³ ; fl	79	83			± 5	79	80				± 5	87	87			± 5		
TGMH MCH	pg	28.1	29.1			± 2.0	28.8	29.0				± 2.0	31.4	31.3			± 2.5		
	fmol	1.74	1.81			± 0.12	1.79	1.80				± 0.12	1.95	1.94			± 0.16		
CCMH MCHC	g/dl	35.6	35.1			± 3.0	36.5	36.3				± 3.0	36.1	36.0			± 3.0		
	g/l	356	351			± 30	365	363				± 30	361	360			± 30		
IDR RDW	mmol/l	22.09	21.78			± 1.86	22.66	22.54				± 1.86	22.39	22.35			± 1.86		
	%	12.7	13.7			± 4.0	12.5	13.6				± 4.0	12.3	13.4			± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	72			± 20	250	245				± 30	510	505			± 50		
VPM MPV	µm ³ ; fl	9.2	9.4			± 2.0	9.4	9.6				± 2.0	9.3	9.6			± 2.0		
NEUT	#	1.36	1.36			± 0.30	4.23	4.25				± 0.80	12.82	12.60			± 1.90		
	%	56.7	56.7			± 10.0	57.9	58.2				± 10.0	72.0	72.0			± 10.0		
LYMPHO	#	0.73	0.72			± 0.25	2.28	2.21				± 0.65	2.58	2.54			± 1.50		
	%	30.5	30.0			± 8.0	31.3	30.3				± 8.0	14.5	14.5			± 8.0		
MONO	#	0.08	0.08			± 0.08	0.30	0.30				± 0.30	0.71	0.62			± 0.62		
	%	3.5	3.3			± 3.3	4.0	4.0				± 4.0	4.0	3.5			± 3.5		
EOS	#	0.14	0.16			± 0.14	0.27	0.29				± 0.27	0.89	0.96			± 0.83		
	%	6.0	6.5			± 5.5	3.5	4.0				± 3.5	5.0	5.5			± 4.5		
BASO	#	0.08	0.08			± 0.08	0.24	0.26				± 0.23	0.80	0.79			± 0.46		
	%	3.3	3.5			± 3.0	3.3	3.5				± 3.0	4.5	4.5			± 2.5		

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 073

CONTROL

(Exp.) **2013-09-05**
 (YYYY-MM-DD)

Rev 2

		ABX Lyse																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20	
HB HGB	g/dl						± 0.4						± 0.5						± 0.6	
	g/l						± 4						± 5						± 6	
	mmol/l						± 0.25						± 0.31						± 0.37	
HT HCT	%						± 1.5						± 2.0						± 2.5	
	l/l						± 0.015						± 0.020						± 0.025	
VGM MCV	µm ³ ; fl						± 5						± 5						± 5	
TGMH MCH	pg						± 2.0						± 2.0						± 2.5	
	fmol						± 0.12						± 0.12						± 0.16	
CCMH MCHC	g/dl						± 3.0						± 3.0						± 3.0	
	g/l						± 30						± 30						± 30	
	mmol/l						± 1.86						± 1.86						± 1.86	
IDR RDW	%						± 4.0						± 4.0						± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						± 20						± 30						± 50	
VPM MPV	µm ³ ; fl						± 2.0						± 2.0						± 2.0	
NEUT	#						± 0.35						± 0.90						± 1.90	
	%						± 10.0						± 10.0						± 10.0	
LYMPHO	#						± 0.25						± 0.70						± 1.50	
	%						± 8.0						± 8.0						± 8.0	
MONO	#						± 0.03						± 0.07						± 0.18	
	%						± 1.0						± 1.0						± 1.0	
EOS	#						± 0.15						± 0.24						± 0.70	
	%						± 4.0						± 3.0						± 3.0	
BASO	#						± 0.06						± 0.25						± 0.60	
	%						± 2.5						± 3.0						± 3.0	

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B