



LOT PX 093

CONTROL

(Exp.) **2013-11-05**
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																				
				CONTROL				L	TOLERANCES TOLERANCE		CONTROL				N	TOLERANCES TOLERANCE		CONTROL				H	TOLERANCES TOLERANCE	
				PENTRA				PENTRA				PENTRA												
				60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.2	2.3				± 0.4	7.2	7.3				± 1.0	17.2	17.1				± 2.2				
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.38				± 0.12	4.62	4.60				± 0.15	5.12	5.12				± 0.20				
HB	HGB	g/dl	6.8	6.8				± 0.4	13.5	13.6				± 0.5	16.2	16.3				± 0.6				
		g/l	68	68				± 4	135	136				± 5	162	163				± 6				
		mmol/l	4.22	4.22				± 0.25	8.38	8.45				± 0.31	10.06	10.12				± 0.37				
HT	HCT	%	19.2	19.5				± 1.5	37.4	37.7				± 2.0	45.1	45.6				± 2.5				
		l/l	0.192	0.195				± 0.015	0.374	0.377				± 0.020	0.451	0.456				± 0.025				
VGM	MCV	µm ³ ; fl	79	82				± 5	81	82				± 5	88	89				± 5				
TGMH	MCH	pg	28.0	28.6				± 2.0	29.2	29.6				± 2.0	31.6	31.8				± 2.5				
		fmol	1.74	1.77				± 0.12	1.81	1.84				± 0.12	1.96	1.98				± 0.16				
CCMH	MCHC	g/dl	35.4	34.8				± 3.0	36.1	36.1				± 3.0	36.0	35.8				± 3.0				
		g/l	354	348				± 30	361	361				± 30	360	358				± 30				
		mmol/l	22.00	21.64				± 1.86	22.40	22.39				± 1.86	22.33	22.21				± 1.86				
IDR	RDW	%	13.0	13.8				± 4.0	13.0	13.7				± 4.0	12.5	13.7				± 4.0				
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	65				± 20	240	245				± 30	500	505				± 50				
VPM	MPV	µm ³ ; fl	9.1	9.2				± 2.0	9.0	9.2				± 2.0	9.0	9.2				± 2.0				
NEUT	#	1.18	1.29				± 0.25	4.08	4.16				± 0.75	12.13	12.06				± 1.80					
	%	53.5	56.0				± 10.0	56.6	57.0				± 10.0	70.5	70.5				± 10.0					
LYMPHO	#	0.70	0.70				± 0.20	2.28	2.26				± 0.60	2.84	2.74				± 1.40					
	%	32.0	30.3				± 8.0	31.6	31.0				± 8.0	16.5	16.0				± 8.0					
MONO	#	0.11	0.09				± 0.09	0.29	0.29				± 0.29	0.69	0.69				± 0.69					
	%	5.0	4.0				± 4.0	4.0	4.0				± 4.0	4.0	4.0				± 4.0					
EOS	#	0.15	0.15				± 0.15	0.31	0.33				± 0.30	0.77	0.86				± 0.62					
	%	6.5	6.5				± 6.5	4.3	4.5				± 4.0	4.5	5.0				± 3.5					
BASO	#	0.07	0.07				± 0.07	0.25	0.26				± 0.19	0.77	0.77				± 0.35					
	%	3.0	3.2				± 3.0	3.5	3.5				± 2.5	4.5	4.5				± 2.0					



ABX Difftrol

LOT PX 093

CONTROL



(Exp.) 2013-11-05
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																	
			CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA					
			120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l					± 0.4						± 1.0						± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l					± 0.12						± 0.15						± 0.20		
HB HGB	g/dl					± 0.4						± 0.5						± 0.6		
	g/l					± 4						± 5					± 6			
	mmol/l					± 0.25						± 0.31					± 0.37			
HT HCT	%					± 1.5						± 2.0					± 2.5			
	l/l					± 0.015						± 0.020					± 0.025			
VGM MCV	µm ³ ·fl					± 5						± 5					± 5			
TGMH MCH	pg					± 2.0						± 2.0					± 2.5			
	fmol					± 0.12						± 0.12					± 0.16			
CCMH MCHC	g/dl					± 3.0						± 3.0					± 3.0			
	g/l					± 30						± 30					± 30			
	mmol/l					± 1.86						± 1.86					± 1.86			
IDR RDW	%					± 4.0						± 4.0					± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l					± 20						± 30					± 50			
VPM MPV	µm ³ ·fl					± 2.0						± 2.0					± 2.0			
NEUT	#					± 0.25						± 0.75					± 1.80			
	%					± 10.0						± 10.0					± 10.0			
LYMPHO	#					± 0.20						± 0.60					± 1.40			
	%					± 8.0						± 8.0					± 8.0			
MONO	#					± 0.09						± 0.29					± 0.69			
	%					± 4.0						± 4.0					± 4.0			
EOS	#					± 0.15						± 0.30					± 0.62			
	%					± 6.5						± 4.0					± 3.5			
BASO	#					± 0.07						± 0.19					± 0.35			
	%					± 3.0						± 2.5					± 2.0			

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B