

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 113

CONTROL

(Exp.) **2014-01-05**
 (YYYY-MM-DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5			± 0.4	7.4	7.3				± 1.0	17.7	17.5			± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.33			± 0.12	4.58	4.56				± 0.15	5.15	5.15			± 0.20		
HB HGB	g/dl	6.8	6.7			± 0.4	13.4	13.4				± 0.5	16.3	16.2			± 0.6		
	g/l	68	67			± 4	134	134				± 5	163	162			± 6		
	mmol/l	4.22	4.16			± 0.25	8.32	8.32				± 0.31	10.12	10.06			± 0.37		
HT HCT	%	18.9	19.1			± 1.5	37.1	37.4				± 2.0	44.8	45.3			± 2.5		
	l/l	0.189	0.191			± 0.015	0.371	0.374				± 0.020	0.448	0.453			± 0.025		
VGM MCV	µm ³ ; fl	79	82			± 5	81	82				± 5	87	88			± 5		
TGMH MCH	pg	28.5	28.8			± 2.0	29.3	29.4				± 2.0	31.7	31.5			± 2.5		
	fmol	1.77	1.79			± 0.12	1.82	1.82				± 0.12	1.97	1.95			± 0.16		
CCMH MCHC	g/dl	36.0	35.1			± 3.0	36.1	35.8				± 3.0	36.4	35.7			± 3.0		
	g/l	360	351			± 30	361	358				± 30	364	357			± 30		
	mmol/l	22.37	21.78			± 1.86	22.43	22.25				± 1.86	22.59	22.20			± 1.86		
IDR RDW	%	13.0	14.0			± 4.0	12.5	13.8				± 4.0	11.6	13.3			± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	70			± 20	258	260				± 30	505	515			± 50		
VPM MPV	µm ³ ; fl	9.0	9.2			± 2.0	8.7	9.0				± 2.0	8.8	8.9			± 2.0		
NEUT	#	1.36	1.39			± 0.30	3.96	3.83				± 0.80	12.74	12.43			± 1.90		
	%	56.5	55.5			± 10.0	53.5	52.5				± 10.0	72.0	71.0			± 10.0		
LYMPHO	#	0.73	0.79			± 0.25	2.59	2.63				± 0.60	2.74	2.80			± 1.50		
	%	30.5	31.5			± 8.0	35.0	36.0				± 8.0	15.5	16.0			± 8.0		
MONO	#	0.08	0.08			± 0.08	0.33	0.30				± 0.30	0.72	0.72			± 0.72		
	%	3.0	3.0			± 3.0	4.5	4.0				± 4.0	4.0	4.0			± 4.0		
EOS	#	0.16	0.16			± 0.11	0.26	0.29				± 0.19	0.71	0.79			± 0.45		
	%	6.5	6.5			± 4.5	3.5	4.0				± 2.5	4.0	4.5			± 2.5		
BASO	#	0.08	0.09			± 0.07	0.26	0.26				± 0.23	0.80	0.79			± 0.45		
	%	3.5	3.5			± 3.0	3.5	3.5				± 3.0	4.5	4.5			± 2.5		

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 113

CONTROL

(Exp.) **2014-01-05**
 (YYYY-MM-DD)

Rev 2

		ABX Lyse																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20	
HB HGB	g/dl						± 0.4						± 0.5						± 0.6	
	g/l						± 4						± 5						± 6	
	mmol/l						± 0.25						± 0.31						± 0.37	
HT HCT	%						± 1.5						± 2.0						± 2.5	
	l/l						± 0.015						± 0.020						± 0.025	
VGM MCV	µm ³ ; fl						± 5						± 5						± 5	
TGMH MCH	pg						± 2.0						± 2.0						± 2.5	
	fmol						± 0.12						± 0.12						± 0.16	
CCMH MCHC	g/dl						± 3.0						± 3.0						± 3.0	
	g/l						± 30						± 30						± 30	
	mmol/l						± 1.86						± 1.86						± 1.86	
IDR RDW	%						± 4.0						± 4.0						± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						± 20						± 30						± 50	
VPM MPV	µm ³ ; fl						± 2.0						± 2.0						± 2.0	
NEUT	#						± 0.30						± 0.80						± 1.90	
	%						± 10.0						± 10.0						± 10.0	
LYMPHO	#						± 0.25						± 0.60						± 1.50	
	%						± 8.0						± 8.0						± 8.0	
MONO	#						± 0.08						± 0.30						± 0.72	
	%						± 3.0						± 4.0						± 4.0	
EOS	#						± 0.11						± 0.19						± 0.45	
	%						± 4.5						± 2.5						± 2.5	
BASO	#						± 0.07						± 0.23						± 0.45	
	%						± 3.0						± 3.0						± 2.5	

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B