



HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

ABX Difftrol



LOT PX 034

CONTROL

(Exp.) **2014-05-05**
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																	
				CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
				PENTRA				PENTRA				PENTRA									
				60 60C+ ES60	80 XL80			60 60C+ ES60		80 XL80			60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4				± 0.4	7.4	7.4				± 1.0	17.6	17.3				± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.31	2.25				± 0.12	4.57	4.52				± 0.15	5.08	5.05				± 0.20	
HB	HGB	g/dl	6.6	6.6				± 0.4	13.6	13.5				± 0.5	15.9	15.8				± 0.6	
		g/l	66	66				± 4	136	135				± 5	159	158				± 6	
		mmol/l	4.10	4.10				± 0.25	8.45	8.38				± 0.31	9.87	9.81				± 0.37	
HT	HCT	%	18.5	18.7				± 1.5	37.5	37.1				± 2.0	44.7	44.4				± 2.5	
		l/l	0.185	0.187				± 0.015	0.375	0.371				± 0.020	0.447	0.444				± 0.025	
VGM	MCV	µm ³ ; fl	80	83				± 5	82	82				± 5	88	88				± 5	
TGMH	MCH	pg	28.6	29.3				± 2.0	29.8	29.9				± 2.0	31.3	31.3				± 2.5	
		fmol	1.77	1.82				± 0.12	1.85	1.85				± 0.12	1.94	1.94				± 0.16	
CCMH	MCHC	g/dl	35.7	35.3				± 3.0	36.3	36.4				± 3.0	35.6	35.6				± 3.0	
		g/l	357	353				± 30	363	364				± 30	356	356				± 30	
		mmol/l	22.18	21.95				± 1.86	22.54	22.62				± 1.86	22.09	22.08				± 1.86	
IDR	RDW	%	12.4	13.5				± 4.0	12.3	12.9				± 4.0	12.0	12.9				± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	75				± 20	254	255				± 30	485	495				± 50	
VPM	MPV	µm ³ ; fl	8.7	9.0				± 2.0	8.9	9.3				± 2.0	8.5	8.9				± 2.0	
NEUT	#	1.30	1.33				± 0.30	4.12	4.13				± 0.80	12.32	11.97				± 1.90		
	%	54.0	55.3				± 10.0	55.7	55.8				± 10.0	70.0	69.2				± 10.0		
LYMPHO	#	0.73	0.71				± 0.35	2.44	2.44				± 0.65	2.82	2.91				± 1.50		
	%	30.5	29.7				± 12.0	33.0	33.0				± 8.0	16.0	16.8				± 8.0		
MONO	#	0.13	0.13				± 0.13	0.30	0.30				± 0.30	0.79	0.71				± 0.71		
	%	5.5	5.0				± 5.0	4.0	4.0				± 4.0	4.5	4.0				± 4.0		
EOS	#	0.18	0.18				± 0.18	0.29	0.29				± 0.29	0.91	0.95				± 0.91		
	%	7.0	7.0				± 7.0	3.8	3.8				± 3.8	5.0	5.5				± 5.0		
BASO	#	0.07	0.07				± 0.07	0.26	0.25				± 0.19	0.79	0.78				± 0.36		
	%	3.0	3.0				± 2.5	3.5	3.4				± 2.5	4.5	4.5				± 2.0		



ABX Difftrol

LOT PX 034

CONTROL



(Exp.) 2014-05-05
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse																	
				CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
				PENTRA				PENTRA				PENTRA	PENTRA								
				120 120 RETIC	DX120 DF120			120 120 RETIC		DX120 DF120				120 120 RETIC		DX120 DF120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20		
HB HGB	g/dl						± 0.4						± 0.5						± 0.6		
	g/l						± 4						± 5						± 6		
	mmol/l						± 0.25						± 0.31						± 0.37		
HT HCT	%						± 1.5						± 2.0						± 2.5		
	l/l						± 0.015						± 0.020						± 0.025		
VGM MCV	µm ³ ; fl						± 5						± 5						± 5		
TGMH MCH	pg						± 2.0						± 2.0						± 2.5		
	fmol						± 0.12						± 0.12						± 0.16		
CCMH MCHC	g/dl						± 3.0						± 3.0						± 3.0		
	g/l						± 30						± 30						± 30		
	mmol/l						± 1.86						± 1.86						± 1.86		
IDR RDW	%						± 4.0						± 4.0						± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l						± 20						± 30						± 50		
VPM MPV	µm ³ ; fl						± 2.0						± 2.0						± 2.0		
NEUT	#						± 0.30						± 0.80						± 1.90		
	%						± 10.0						± 10.0						± 10.0		
LYMPHO	#						± 0.35						± 0.65						± 1.50		
	%						± 12.0						± 8.0						± 8.0		
MONO	#						± 0.13						± 0.30						± 0.71		
	%						± 5.0						± 4.0						± 4.0		
EOS	#						± 0.18						± 0.29						± 0.91		
	%						± 7.0						± 3.8						± 5.0		
BASO	#						± 0.07						± 0.19						± 0.36		
	%						± 2.5						± 2.5						± 2.0		

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B