


**LOT** PX 413  
Rev 2


**CONTROL**

 (Exp.) 2018-11-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5			± 0.4	7.2	7.2			± 1.0	17.5	17.5			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.36	2.36			± 0.16	4.64	4.64			± 0.20	5.18	5.18			± 0.25			
HB HGB	g/dl	6.9	6.9			± 0.4	14.1	14.1			± 0.5	16.6	16.6			± 0.6			
	g/l	69	69			± 4	141	141			± 5	166	166			± 6			
	mmol/l	4.28	4.28			± 0.25	8.76	8.76			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	20.1	20.1			± 1.5	39.0	39.0			± 2.0	45.1	46.1			± 2.5			
	l/l	0.201	0.201			± 0.015	0.390	0.390			± 0.020	0.451	0.461			± 0.025			
VGM MCV	µm <sup>3</sup> fl	85	85			± 5	84	84			± 5	87	89			± 5			
TGMH MCH	pg	29.2	29.2			± 2.0	30.4	30.4			± 2.0	32.0	32.0			± 2.5			
	fmol	1.82	1.82			± 0.12	1.89	1.89			± 0.12	1.99	1.99			± 0.16			
CCMH MCHC	g/dl	34.4	34.4			± 3.0	36.2	36.2			± 3.0	36.8	36.0			± 3.0			
	g/l	344	344			± 30	362	362			± 30	368	360			± 30			
	mmol/l	21.36	21.36			± 1.86	22.47	22.47			± 1.86	22.87	22.36			± 1.86			
IDR RDW	%	13.7	12.7			± 4.0	13.5	12.6			± 4.0	13.8	13.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	73			± 20	250	250			± 30	495	495			± 50			
VPM MPV	µm <sup>3</sup> fl	9.2	9.2			± 2.0	8.9	9.1			± 2.0	8.7	9.0			± 2.0			
NEUT	#	1.50	1.50			± 0.35	4.18	4.18			± 0.90	12.20	12.20			± 1.90			
	%	60.1	60.1			± 10.0	58.0	58.0			± 10.0	69.7	69.7			± 10.0			
LYMPHO	#	0.66	0.66			± 0.33	2.20	2.20			± 0.70	2.82	2.82			± 1.50			
	%	26.4	26.4			± 12.0	30.5	30.5			± 8.0	16.1	16.1			± 8.0			
MONO	#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.67	0.67			± 0.67			
	%	3.3	3.3			± 3.3	3.4	3.4			± 3.4	3.8	3.8			± 3.8			
EOS	#	0.17	0.17			± 0.17	0.30	0.30			± 0.30	1.05	1.05			± 1.05			
	%	6.7	6.7			± 6.7	4.2	4.2			± 4.2	6.0	6.0			± 6.0			
BASO	#	0.09	0.09			± 0.09	0.28	0.28			± 0.28	0.77	0.77			± 0.77			
	%	3.5	3.5			± 3.5	3.9	3.9			± 3.9	4.4	4.4			± 4.4			

**LOT** PX 413  
Rev 2

**CONTROL**

 (Exp.) 2018-11-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5		± 0.4	7.2	7.2	7.3		± 1.0	17.6	17.4	18.0		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.37	2.40		± 0.16	4.61	4.64	4.64		± 0.20	5.12	5.15	5.19		± 0.25			
HB HGB	g/dl	6.9	6.8	6.9		± 0.4	14.0	14.0	13.9		± 0.5	16.5	16.5	16.5		± 0.6			
	g/l	69	68	69		± 4	140	140	139		± 5	165	165	165		± 6			
	mmol/l	4.28	4.22	4.28		± 0.25	8.69	8.69	8.63		± 0.31	10.25	10.25	10.25		± 0.37			
HT HCT	%	19.8	19.4	19.4		± 1.5	38.7	38.5	38.3		± 2.0	45.1	45.3	45.7		± 2.5			
	l/l	0.198	0.194	0.194		± 0.015	0.387	0.385	0.383		± 0.020	0.451	0.453	0.457		± 0.025			
VGM MCV	µm <sup>3</sup> ·fl	82	82	81.0		± 5	84	83	82.5		± 5	88	88	88.0		± 5			
TGMH MCH	pg	28.6	28.7	28.8		± 2.0	30.4	30.2	30.0		± 2.0	32.2	32.0	31.8		± 2.5			
	fmol	1.78	1.78	1.79		± 0.12	1.89	1.87	1.86		± 0.12	2.00	1.99	1.97		± 0.16			
CCMH MCHC	g/dl	34.9	35.0	35.5		± 3.0	36.2	36.4	36.3		± 3.0	36.6	36.4	36.1		± 3.0			
	g/l	349	350	355		± 30	362	364	363		± 30	366	364	361		± 30			
	mmol/l	21.68	21.73	22.04		± 1.86	22.45	22.57	22.55		± 1.86	22.74	22.61	22.43		± 1.86			
IDR RDW	%	13.5	12.8	12.8		± 4.0	13.2	12.5	12.4		± 4.0	13.3	12.3	12.3		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	74	72		± 20	250	255	245		± 30	490	505	480		± 50			
VPM MPV	µm <sup>3</sup> ·fl	9.0	9.0	8.8		± 2.0	8.8	8.9	8.3		± 2.0	8.5	8.7	8.1		± 2.0			
NEUT	#	1.44	1.46	1.53		± 0.35	4.17	4.19	4.34		± 0.90	12.32	12.23	12.96		± 1.90			
	%	59.8	60.9	61.0		± 10.0	57.9	58.2	59.5		± 10.0	70.0	70.3	72.0		± 10.0			
LYMPHO	#	0.65	0.62	0.66		± 0.33	2.24	2.18	2.23		± 0.70	2.80	2.64	2.70		± 1.50			
	%	27.1	26.0	26.5		± 12.0	31.1	30.3	30.5		± 8.0	15.9	15.2	15.0		± 8.0			
MONO	#	0.07	0.07	0.08		± 0.07	0.22	0.25	0.22		± 0.22	0.65	0.68	0.54		± 0.54			
	%	3.0	3.1	3.0		± 3.0	3.1	3.5	3.0		± 3.0	3.7	3.9	3.0		± 3.0			
EOS	#	0.16	0.16	0.15		± 0.15	0.30	0.32	0.26		± 0.26	1.04	1.08	0.99		± 0.99			
	%	6.6	6.5	6.0		± 6.0	4.1	4.4	3.5		± 3.5	5.9	6.2	5.5		± 5.5			
BASO	#	0.08	0.08	0.09		± 0.08	0.27	0.26	0.26		± 0.26	0.79	0.77	0.81		± 0.77			
	%	3.5	3.5	3.5		± 3.5	3.8	3.6	3.5		± 3.5	4.5	4.4	4.5		± 4.4			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42