


**LOT** PX 423  
Rev 2


**CONTROL**

 (Exp.) 2020-07-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	17.2	17.2			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30	2.30			± 0.16	4.63	4.63			± 0.20	5.14	5.14			± 0.25			
HB HGB	g/dl	6.5	6.5			± 0.4	13.7	13.7			± 0.5	16.3	16.3			± 0.6			
	g/l	65	65			± 4	137	137			± 5	163	163			± 6			
	mmol/l	4.04	4.04			± 0.25	8.51	8.51			± 0.31	10.12	10.12			± 0.37			
HT HCT	%	19.1	19.1			± 1.5	38.4	38.4			± 2.0	44.7	44.7			± 2.5			
	l/l	0.191	0.191			± 0.015	0.384	0.384			± 0.020	0.447	0.447			± 0.025			
VGM MCV	µm <sup>3</sup> fl	83	83			± 5	83	83			± 5	87	87			± 5			
TGMH MCH	pg	28.3	28.3			± 2.0	29.6	29.6			± 2.0	31.7	31.7			± 2.5			
	fmol	1.76	1.76			± 0.12	1.84	1.84			± 0.12	1.97	1.97			± 0.16			
CCMH MCHC	g/dl	34.0	34.0			± 3.0	35.7	35.7			± 3.0	36.5	36.5			± 3.0			
	g/l	340	340			± 30	357	357			± 30	365	365			± 30			
	mmol/l	21.14	21.14			± 1.86	22.14	22.14			± 1.86	22.64	22.64			± 1.86			
IDR RDW	%	13.8	13.5			± 4.0	14.0	13.0			± 4.0	13.5	12.5			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	66	66			± 20	247	247			± 30	473	473			± 50			
VPM MPV	µm <sup>3</sup> fl	9.5	9.5			± 2.0	8.9	8.9			± 2.0	9.0	9.0			± 2.0			
NEUT	#	1.38	1.38			± 0.35	4.06	4.06			± 0.90	11.94	11.94			± 1.90			
	%	55.2	55.2			± 10.0	54.8	54.8			± 10.0	69.4	69.4			± 10.0			
LYMPHO	#	0.74	0.74			± 0.33	2.37	2.37			± 0.70	2.79	2.79			± 1.50			
	%	29.7	29.7			± 12.0	32.0	32.0			± 8.0	16.2	16.2			± 8.0			
MONO	#	0.12	0.12			± 0.12	0.44	0.44			± 0.44	0.91	0.91			± 0.91			
	%	4.7	4.7			± 4.7	5.9	5.9			± 5.9	5.3	5.3			± 5.3			
EOS	#	0.18	0.18			± 0.18	0.28	0.28			± 0.28	0.81	0.81			± 0.81			
	%	7.1	7.1			± 7.1	3.8	3.8			± 3.8	4.7	4.7			± 4.7			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.76	0.76			± 0.76			
	%	3.3	3.3			± 3.3	3.5	3.5			± 3.5	4.4	4.4			± 4.4			

**LOT** PX 423  
Rev 2

**CONTROL**

 (Exp.) 2020-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP							
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.6			± 0.4	7.5	7.5	7.6			± 1.0	17.6	17.4	17.4			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.36	2.32	2.34			± 0.16	4.65	4.63	4.64			± 0.20	5.14	5.10	5.12			± 0.25
		g/dl	6.6	6.5	6.5			± 0.4	13.7	13.6	13.7			± 0.5	16.3	16.2	16.3			± 0.6
HB	HGB	g/l	66	65	65			± 4	137	136	137			± 5	163	162	163			± 6
		mmol/l	4.10	4.04	4.04			± 0.25	8.51	8.45	8.51			± 0.31	10.12	10.06	10.12			± 0.37
HT	HCT	%	18.9	18.6	18.6			± 1.5	38.1	37.5	37.8			± 2.0	44.7	43.9	44.0			± 2.5
		l/l	0.189	0.186	0.186			± 0.015	0.381	0.375	0.378			± 0.020	0.447	0.439	0.440			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	80	80	79.5			± 5	82	81	81.5			± 5	87	86	86.0			± 5
TGMH	MCH	pg	28.0	28.0	27.8			± 2.0	29.5	29.4	29.5			± 2.0	31.7	31.8	31.8			± 2.5
		fmol	1.74	1.74	1.73			± 0.12	1.83	1.82	1.83			± 0.12	1.97	1.97	1.98			± 0.16
		g/dl	35.0	35.0	34.9			± 3.0	35.9	36.3	36.2			± 3.0	36.5	36.9	37.0			± 3.0
CCMH	MCHC	g/l	350	350	349			± 30	359	363	362			± 30	365	369	370			± 30
		mmol/l	21.71	21.75	21.70			± 1.86	22.31	22.52	22.50			± 1.86	22.64	22.94	22.99			± 1.86
IDR	RDW	%	13.5	13.0	13.0			± 4.0	13.5	13.0	13.0			± 4.0	13.0	13.0	12.0			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	67	68			± 20	244	248	240			± 30	472	477	465			± 50
VPM	MPV	µm <sup>3</sup> ; fl	9.3	9.4	8.9			± 2.0	8.6	8.5	8.4			± 2.0	8.7	8.5	8.4			± 2.0
		#	1.38	1.36	1.44			± 0.35	4.10	4.04	4.18			± 0.90	12.28	12.02	12.35			± 1.90
		%	55.2	54.3	55.5			± 10.0	54.7	53.9	55.0			± 10.0	69.8	69.1	71.0			± 10.0
		#	0.75	0.74	0.79			± 0.33	2.45	2.42	2.51			± 0.70	2.83	2.77	2.70			± 1.50
		%	30.0	29.5	30.5			± 12.0	32.6	32.3	33.0			± 8.0	16.1	15.9	15.5			± 8.0
		#	0.11	0.14	0.10			± 0.10	0.41	0.48	0.38			± 0.38	0.86	0.97	0.78			± 0.78
		%	4.4	5.5	4.0			± 4.0	5.4	6.4	5.0			± 5.0	4.9	5.6	4.5			± 4.5
		#	0.18	0.19	0.17			± 0.17	0.28	0.29	0.27			± 0.27	0.83	0.87	0.78			± 0.78
		%	7.1	7.4	6.5			± 6.5	3.7	3.9	3.5			± 3.5	4.7	5.0	4.5			± 4.5
		#	0.08	0.08	0.09			± 0.08	0.27	0.26	0.27			± 0.26	0.79	0.77	0.78			± 0.77
		%	3.3	3.3	3.5			± 3.3	3.6	3.5	3.5			± 3.5	4.5	4.4	4.5			± 4.4

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44