


**LOT** PX 426  
Rev 2


**CONTROL**

 (Exp.) 2021-01-05  
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.8	2.8				± 0.4	7.8	7.8				± 1.0	18.0	18.0				± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.23	2.23				± 0.16	4.52	4.52				± 0.20	5.00	5.00				± 0.25	
HB HGB	g/dl	6.3	6.3				± 0.4	13.6	13.6				± 0.5	16.2	16.2				± 0.6	
	g/l	63	63				± 4	136	136				± 5	162	162				± 6	
	mmol/l	3.91	3.91				± 0.25	8.45	8.45				± 0.31	10.06	10.06				± 0.37	
HT HCT	%	18.3	18.3				± 1.5	37.5	37.5				± 2.0	44.5	44.5				± 2.5	
	l/l	0.183	0.183				± 0.015	0.375	0.375				± 0.020	0.445	0.445				± 0.025	
VGM MCV	µm <sup>3</sup> fl	82	82				± 5	83	83				± 5	89	89				± 5	
TGMH MCH	pg	28.3	28.3				± 2.0	30.1	30.1				± 2.0	32.4	32.4				± 2.5	
	fmol	1.75	1.75				± 0.12	1.87	1.87				± 0.12	2.01	2.01				± 0.16	
CCMH MCHC	g/dl	34.5	34.5				± 3.0	36.3	36.3				± 3.0	36.4	36.4				± 3.0	
	g/l	345	345				± 30	363	363				± 30	364	364				± 30	
	mmol/l	21.40	21.40				± 1.86	22.51	22.51				± 1.86	22.61	22.61				± 1.86	
IDR RDW	%	14.0	13.0				± 4.0	13.3	12.5				± 4.0	13.0	12.0				± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	66	66				± 20	245	245				± 30	490	490				± 50	
VPM MPV	µm <sup>3</sup> fl	9.9	9.9				± 2.0	8.9	8.9				± 2.0	8.9	8.9				± 2.0	
NEUT	#	1.47	1.47				± 0.35	4.05	4.05				± 0.90	12.02	12.02				± 1.90	
	%	52.4	52.4				± 10.0	51.9	51.9				± 10.0	66.8	66.8				± 10.0	
LYMPHO	#	0.80	0.80				± 0.33	2.35	2.35				± 0.70	2.84	2.84				± 1.50	
	%	28.6	28.6				± 12.0	30.1	30.1				± 8.0	15.8	15.8				± 8.0	
MONO	#	0.23	0.23				± 0.23	0.73	0.73				± 0.73	1.46	1.46				± 1.46	
	%	8.1	8.1				± 8.1	9.3	9.3				± 9.3	8.1	8.1				± 8.1	
EOS	#	0.21	0.21				± 0.21	0.38	0.38				± 0.38	0.86	0.86				± 0.86	
	%	7.5	7.5				± 7.5	4.9	4.9				± 4.9	4.8	4.8				± 4.8	
BASO	#	0.10	0.10				± 0.10	0.30	0.30				± 0.30	0.81	0.81				± 0.81	
	%	3.4	3.4				± 3.4	3.8	3.8				± 3.8	4.5	4.5				± 4.5	

**LOT** PX 426  
Rev 2

**CONTROL**

 (Exp.) 2021-01-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60		MS60	MSCRP		60 60C+ ES60	MS60		MSCRP		60 60C+ ES60	MS60	MSCRP	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.8	2.8	2.7		± 0.4	7.8	7.9	7.8		± 1.0	18.5	18.6	18.5		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.28	2.26	2.26		± 0.16	4.50	4.49	4.50		± 0.20	5.00	4.96	5.00		± 0.25			
	g/dl	6.4	6.3	6.4		± 0.4	13.6	13.4	13.5		± 0.5	16.3	16.0	16.2		± 0.6			
HB HGB	g/l	64	63	64		± 4	136	134	135		± 5	163	160	162		± 6			
	mmol/l	3.97	3.91	3.97		± 0.25	8.45	8.32	8.38		± 0.31	10.12	9.94	10.06		± 0.37			
HT HCT	%	18.2	17.9	17.7		± 1.5	37.4	37.3	37.4		± 2.0	45.0	44.1	44.0		± 2.5			
	l/l	0.182	0.179	0.177		± 0.015	0.374	0.373	0.374		± 0.020	0.450	0.441	0.440		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	80	79	78.5		± 5	83	83	83.0		± 5	90	89	88.0		± 5			
TGMH MCH	pg	28.1	27.9	28.3		± 2.0	30.2	29.8	30.0		± 2.0	32.6	32.3	32.4		± 2.5			
	fmol	1.74	1.73	1.76		± 0.12	1.88	1.85	1.86		± 0.12	2.02	2.00	2.01		± 0.16			
	g/dl	35.1	35.3	36.1		± 3.0	36.4	36.0	36.1		± 3.0	36.2	36.2	36.8		± 3.0			
CCMH MCHC	g/l	351	353	361		± 30	364	360	361		± 30	362	362	368		± 30			
	mmol/l	21.79	21.91	22.40		± 1.86	22.61	22.33	22.45		± 1.86	22.49	22.51	22.86		± 1.86			
IDR RDW	%	13.5	13.0	13.2		± 4.0	12.5	12.0	12.0		± 4.0	12.0	12.0	11.5		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	67	65		± 20	247	247	244		± 30	487	490	486		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.7	9.9	9.2		± 2.0	8.6	8.6	8.3		± 2.0	8.6	8.6	8.3		± 2.0			
NEUT	#	1.48	1.44	1.50		± 0.35	4.17	4.01	4.35		± 0.90	12.56	12.03	12.67		± 1.90			
	%	52.7	51.4	55.5		± 10.0	53.5	50.7	55.8		± 10.0	67.9	64.7	68.5		± 10.0			
LYMPHO	#	0.79	0.80	0.75		± 0.33	2.31	2.31	2.18		± 0.70	2.87	2.85	2.68		± 1.50			
	%	28.3	28.7	27.7		± 12.0	29.6	29.3	28.0		± 8.0	15.5	15.3	14.5		± 8.0			
MONO	#	0.22	0.27	0.18		± 0.18	0.65	0.92	0.62		± 0.62	1.33	1.97	1.41		± 1.33			
	%	7.7	9.5	6.5		± 6.5	8.3	11.6	8.0		± 8.0	7.2	10.6	7.6		± 7.2			
EOS	#	0.22	0.20	0.18		± 0.18	0.37	0.36	0.35		± 0.35	0.89	0.89	0.93		± 0.89			
	%	7.9	7.0	6.8		± 6.8	4.8	4.6	4.5		± 4.5	4.8	4.8	5.0		± 4.8			
BASO	#	0.10	0.10	0.09		± 0.09	0.30	0.30	0.29		± 0.29	0.85	0.86	0.81		± 0.81			
	%	3.4	3.4	3.5		± 3.4	3.8	3.8	3.7		± 3.7	4.6	4.6	4.4		± 4.4			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44