


LOT PX 418
Rev 2


CONTROL

 (Exp.) 2019-09-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5				± 0.4	7.5	7.5				± 1.0	18.0	18.0				± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.27	2.27				± 0.16	4.58	4.58				± 0.20	5.21	5.21				± 0.25	
HB HGB	g/dl	6.8	6.8				± 0.4	14.0	14.0				± 0.5	16.8	16.8				± 0.6	
	g/l	68	68				± 4	140	140				± 5	168	168				± 6	
	mmol/l	4.22	4.22				± 0.25	8.69	8.69				± 0.31	10.43	10.43				± 0.37	
HT HCT	%	19.5	19.5				± 1.5	38.9	38.9				± 2.0	45.8	45.8				± 2.5	
	l/l	0.195	0.195				± 0.015	0.389	0.389				± 0.020	0.458	0.458				± 0.025	
VGM MCV	µm ³ fl	86	86				± 5	85	85				± 5	88	88				± 5	
TGMH MCH	pg	30.0	30.0				± 2.0	30.6	30.6				± 2.0	32.2	32.2				± 2.5	
	fmol	1.86	1.86				± 0.12	1.90	1.90				± 0.12	2.00	2.00				± 0.16	
CCMH MCHC	g/dl	34.8	34.8				± 3.0	36.0	36.0				± 3.0	36.6	36.6				± 3.0	
	g/l	348	348				± 30	360	360				± 30	366	366				± 30	
	mmol/l	21.63	21.63				± 1.86	22.33	22.33				± 1.86	22.76	22.76				± 1.86	
IDR RDW	%	13.5	11.5				± 4.0	13.3	11.5				± 4.0	13.0	11.8				± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67				± 20	246	246				± 30	503	503				± 50	
VPM MPV	µm ³ fl	9.5	9.9				± 2.0	9.2	9.7				± 2.0	9.5	9.9				± 2.0	
NEUT	#	1.45	1.45				± 0.35	4.18	4.18				± 0.90	12.53	12.53				± 1.90	
	%	57.8	57.8				± 10.0	55.7	55.7				± 10.0	69.6	69.6				± 10.0	
LYMPHO	#	0.72	0.72				± 0.33	2.45	2.45				± 0.70	2.93	2.93				± 1.50	
	%	28.7	28.7				± 12.0	32.6	32.6				± 8.0	16.3	16.3				± 8.0	
MONO	#	0.10	0.10				± 0.10	0.35	0.35				± 0.35	0.79	0.79				± 0.79	
	%	3.8	3.8				± 3.8	4.6	4.6				± 4.6	4.4	4.4				± 4.4	
EOS	#	0.16	0.16				± 0.16	0.27	0.27				± 0.27	0.95	0.95				± 0.95	
	%	6.3	6.3				± 6.3	3.6	3.6				± 3.6	5.3	5.3				± 5.3	
BASO	#	0.09	0.09				± 0.09	0.26	0.26				± 0.26	0.79	0.79				± 0.79	
	%	3.4	3.4				± 3.4	3.5	3.5				± 3.5	4.4	4.4				± 4.4	

LOT PX 418
Rev 2

CONTROL

 (Exp.) 2019-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRCP		60 60C+ ES60	MS60	MSCRCP		60 60C+ ES60	MS60	MSCRCP							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.5	7.4	7.5			± 1.0	18.2	17.8	18.0			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.31	2.31			± 0.16	4.58	4.55	4.56			± 0.20	5.16	5.18	5.19			± 0.25
		g/dl	6.8	6.7	6.7			± 0.4	13.9	13.8	13.9			± 0.5	16.7	16.5	16.6			± 0.6
HB	HGB	g/l	68	67	67			± 4	139	138	139			± 5	167	165	166			± 6
		mmol/l	4.22	4.16	4.16			± 0.25	8.63	8.57	8.63			± 0.31	10.37	10.25	10.31			± 0.37
HT	HCT	%	19.3	18.9	18.9			± 1.5	38.0	37.3	38.1			± 2.0	45.4	45.1	45.2			± 2.5
		l/l	0.193	0.189	0.189			± 0.015	0.380	0.373	0.381			± 0.020	0.454	0.451	0.452			± 0.025
VGM	MCV	µm ³ ; fl	83	82	82.0			± 5	83	82	83.5			± 5	88	87	87.0			± 5
TGMH	MCH	pg	29.2	29.0	29.0			± 2.0	30.3	30.3	30.5			± 2.0	32.4	31.9	32.0			± 2.5
		fmol	1.81	1.80	1.80			± 0.12	1.88	1.88	1.89			± 0.12	2.01	1.98	1.99			± 0.16
		g/dl	35.2	35.4	35.4			± 3.0	36.6	37.0	36.5			± 3.0	36.8	36.6	36.8			± 3.0
CCMH	MCHC	g/l	352	354	354			± 30	366	370	365			± 30	368	366	368			± 30
		mmol/l	21.84	21.97	21.97			± 1.86	22.71	22.97	22.67			± 1.86	22.84	22.74	22.83			± 1.86
IDR	RDW	%	12.7	12.7	12.5			± 4.0	12.5	12.0	12.0			± 4.0	12.3	12.0	12.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	71	69	69			± 20	250	250	248			± 30	503	505	495			± 50
VPM	MPV	µm ³ ; fl	9.3	9.5	9.0			± 2.0	9.2	9.2	8.6			± 2.0	9.4	9.3	8.7			± 2.0
		#	1.41	1.39	1.45			± 0.35	4.20	4.12	4.24			± 0.90	12.72	12.34	12.69			± 1.90
		%	58.7	57.9	58.0			± 10.0	56.0	55.7	56.5			± 10.0	69.9	69.3	70.5			± 10.0
		#	0.66	0.66	0.73			± 0.33	2.42	2.31	2.48			± 0.70	2.93	2.74	2.97			± 1.50
		%	27.5	27.4	29.0			± 12.0	32.3	31.2	33.0			± 8.0	16.1	15.4	16.5			± 8.0
		#	0.09	0.11	0.09			± 0.09	0.32	0.42	0.34			± 0.32	0.80	1.01	0.72			± 0.72
		%	3.6	4.6	3.5			± 3.5	4.3	5.7	4.5			± 4.3	4.4	5.7	4.0			± 4.0
		#	0.16	0.16	0.15			± 0.15	0.29	0.30	0.23			± 0.23	0.95	0.94	0.81			± 0.81
		%	6.7	6.6	6.0			± 6.0	3.8	4.0	3.0			± 3.0	5.2	5.3	4.5			± 4.5
		#	0.08	0.08	0.09			± 0.08	0.27	0.25	0.23			± 0.23	0.80	0.77	0.81			± 0.77
		%	3.5	3.5	3.5			± 3.5	3.6	3.4	3.0			± 3.0	4.4	4.3	4.5			± 4.3

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43