


LOT PX 428
Rev 2


CONTROL

 (Exp.) 2021-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio															
		CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		XL80	XLR			XL80	XLR			XL80	XLR			XL80	XLR		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6			± 0.4	7.7	7.7			± 1.0	18.2	18.2			± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.32	2.32			± 0.16	4.65	4.65			± 0.20	5.12	5.12			± 0.25	
HB HGB	g/dl	6.8	6.8			± 0.4	14.1	14.1			± 0.5	16.6	16.6			± 0.6	
	g/l	68	68			± 4	141	141			± 5	166	166			± 6	
	mmol/l	4.22	4.22			± 0.25	8.76	8.76			± 0.31	10.31	10.31			± 0.37	
HT HCT	%	19.5	19.5			± 1.5	39.1	39.1			± 2.0	45.6	45.6			± 2.5	
	l/l	0.195	0.195			± 0.015	0.391	0.391			± 0.020	0.456	0.456			± 0.025	
VGM MCV	µm ³ ; fl	84	84			± 5	84	84			± 5	89	89			± 5	
TGMH MCH	pg	29.3	29.3			± 2.0	30.3	30.3			± 2.0	32.4	32.4			± 2.5	
	fmol	1.82	1.82			± 0.12	1.88	1.88			± 0.12	2.01	2.01			± 0.16	
CCMH MCHC	g/dl	34.9	34.9			± 3.0	36.1	36.1			± 3.0	36.4	36.4			± 3.0	
	g/l	349	349			± 30	361	361			± 30	364	364			± 30	
	mmol/l	21.67	21.67			± 1.86	22.42	22.42			± 1.86	22.62	22.62			± 1.86	
IDR RDW	%	13.5	13.5			± 4.0	12.5	12.5			± 4.0	12.0	12.0			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	243	243			± 30	477	477			± 50	
VPM MPV	µm ³ ; fl	9.5	9.5			± 2.0	9.4	9.4			± 2.0	10.0	10.0			± 2.0	
NEUT	#	1.31	1.31			± 0.35	4.13	4.13			± 0.90	12.30	12.30			± 1.90	
	%	50.4	50.4			± 10.0	53.6	53.6			± 10.0	67.6	67.6			± 10.0	
LYMPHO	#	0.83	0.83			± 0.33	2.38	2.38			± 0.70	2.73	2.73			± 1.50	
	%	32.1	32.1			± 12.0	30.9	30.9			± 8.0	15.0	15.0			± 8.0	
MONO	#	0.16	0.16			± 0.16	0.59	0.59			± 0.59	1.51	1.51			± 1.51	
	%	6.3	6.3			± 6.3	7.6	7.6			± 7.6	8.3	8.3			± 8.3	
EOS	#	0.20	0.20			± 0.20	0.34	0.34			± 0.34	0.86	0.86			± 0.86	
	%	7.8	7.8			± 7.8	4.4	4.4			± 4.4	4.7	4.7			± 4.7	
BASO	#	0.09	0.09			± 0.09	0.27	0.27			± 0.27	0.80	0.80			± 0.80	
	%	3.4	3.4			± 3.4	3.5	3.5			± 3.5	4.4	4.4			± 4.4	

LOT PX 428
Rev 2

CONTROL

 (Exp.) 2021-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE		
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE			
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.7	2.5		± 0.4	7.9	7.9	7.5		± 1.0	19.2	18.7	18.0		± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.37	2.40		± 0.16	4.67	4.63	4.68		± 0.20	5.15	5.07	5.20		± 0.25
		g/dl	6.8	6.7	6.8		± 0.4	14.1	13.9	14.1		± 0.5	16.7	16.3	16.6		± 0.6
HB	HGB	g/l	68	67	68		± 4	141	139	141		± 5	167	163	166		± 6
		mmol/l	4.22	4.16	4.22		± 0.25	8.76	8.63	8.76		± 0.31	10.37	10.12	10.31		± 0.37
HT	HCT	%	19.5	19.0	19.2		± 1.5	38.8	38.0	38.8		± 2.0	45.8	44.6	45.5		± 2.5
		l/l	0.195	0.190	0.192		± 0.015	0.388	0.380	0.388		± 0.020	0.458	0.446	0.455		± 0.025
VGM	MCV	µm ³ ; fl	81	80	80.0		± 5	83	82	83.0		± 5	89	88	87.5		± 5
TGMH	MCH	pg	28.2	28.3	28.3		± 2.0	30.2	30.0	30.1		± 2.0	32.4	32.1	31.9		± 2.5
		fmol	1.75	1.76	1.76		± 0.12	1.87	1.86	1.87		± 0.12	2.01	2.00	1.98		± 0.16
		g/dl	34.8	35.3	35.4		± 3.0	36.4	36.6	36.3		± 3.0	36.4	36.5	36.5		± 3.0
CCMH	MCHC	g/l	348	353	354		± 30	364	366	363		± 30	364	365	365		± 30
		mmol/l	21.63	21.94	21.99		± 1.86	22.59	22.74	22.54		± 1.86	22.63	22.69	22.66		± 1.86
IDR	RDW	%	13.5	13.0	13.0		± 4.0	12.5	12.0	12.0		± 4.0	12.5	12.0	11.5		± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	71	68	69		± 20	250	244	245		± 30	490	490	475		± 50
VPM	MPV	µm ³ ; fl	9.3	9.5	8.8		± 2.0	8.9	9.2	8.5		± 2.0	9.3	9.6	9.1		± 2.0
		#	1.26	1.33	1.39		± 0.35	4.17	4.06	4.24		± 0.90	12.98	12.42	12.78		± 1.90
		%	50.4	49.2	55.5		± 10.0	52.8	51.4	56.5		± 10.0	67.6	66.4	71.0		± 10.0
		#	0.85	0.96	0.75		± 0.33	2.64	2.70	2.33		± 0.70	3.21	3.20	2.61		± 1.50
		%	33.9	35.4	30.0		± 12.0	33.4	34.2	31.0		± 8.0	16.7	17.1	14.5		± 8.0
		#	0.13	0.14	0.13		± 0.13	0.48	0.51	0.41		± 0.41	1.38	1.51	0.99		± 0.99
		%	5.1	5.1	5.0		± 5.0	6.1	6.4	5.5		± 5.5	7.2	8.1	5.5		± 5.5
		#	0.18	0.19	0.15		± 0.15	0.32	0.34	0.26		± 0.26	0.77	0.77	0.81		± 0.77
		%	7.2	6.9	6.0		± 6.0	4.0	4.3	3.5		± 3.5	4.0	4.1	4.5		± 4.0
		#	0.09	0.09	0.09		± 0.09	0.29	0.28	0.26		± 0.26	0.86	0.82	0.81		± 0.81
		%	3.4	3.4	3.5		± 3.4	3.7	3.5	3.5		± 3.5	4.5	4.4	4.5		± 4.4

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44