


LOT PX 429
Rev 2


CONTROL

 (Exp.) 2021-07-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		XL80	XLR					XL80	XLR					XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7				± 0.4	7.5	7.5				± 1.0	17.7	17.7				± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.11	2.11				± 0.16	4.59	4.59				± 0.20	5.14	5.14				± 0.25	
HB HGB	g/dl	6.0	6.0				± 0.4	13.9	13.9				± 0.5	16.8	16.8				± 0.6	
	g/l	60	60				± 4	139	139				± 5	168	168				± 6	
	mmol/l	3.73	3.73				± 0.25	8.63	8.63				± 0.31	10.43	10.43				± 0.37	
HT HCT	%	17.5	17.5				± 1.5	39.0	39.0				± 2.0	45.7	45.7				± 2.5	
	l/l	0.175	0.175				± 0.015	0.390	0.390				± 0.020	0.457	0.457				± 0.025	
VGM MCV	µm ³ fl	83	83				± 5	85	85				± 5	89	89				± 5	
TGMH MCH	pg	28.4	28.4				± 2.0	30.3	30.3				± 2.0	32.7	32.7				± 2.5	
	fmol	1.77	1.77				± 0.12	1.88	1.88				± 0.12	2.03	2.03				± 0.16	
CCMH MCHC	g/dl	34.3	34.3				± 3.0	35.6	35.6				± 3.0	36.7	36.7				± 3.0	
	g/l	343	343				± 30	356	356				± 30	367	367				± 30	
	mmol/l	21.28	21.28				± 1.86	22.12	22.12				± 1.86	22.81	22.81				± 1.86	
IDR RDW	%	14.0	14.0				± 5.0	13.5	13.5				± 5.0	12.7	12.7				± 5.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67				± 20	238	238				± 30	485	485				± 50	
VPM MPV	µm ³ fl	10.0	10.0				± 2.0	10.3	10.3				± 2.0	10.0	10.0				± 2.0	
NEUT	#	1.36	1.36				± 0.35	4.12	4.12				± 0.90	12.09	12.09				± 1.90	
	%	50.5	50.5				± 10.0	54.9	54.9				± 10.0	68.3	68.3				± 10.0	
LYMPHO	#	0.80	0.80				± 0.33	2.33	2.33				± 0.70	2.64	2.64				± 1.50	
	%	29.7	29.7				± 12.0	31.1	31.1				± 8.0	14.9	14.9				± 8.0	
MONO	#	0.15	0.15				± 0.15	0.52	0.52				± 0.52	1.29	1.29				± 1.29	
	%	5.6	5.6				± 5.6	6.9	6.9				± 6.9	7.3	7.3				± 7.3	
EOS	#	0.30	0.30				± 0.30	0.29	0.29				± 0.29	0.92	0.92				± 0.92	
	%	11.1	11.1				± 11.1	3.8	3.8				± 3.8	5.2	5.2				± 5.2	
BASO	#	0.08	0.08				± 0.08	0.25	0.25				± 0.25	0.76	0.76				± 0.76	
	%	3.1	3.1				± 3.1	3.3	3.3				± 3.3	4.3	4.3				± 4.3	

LOT PX 429
Rev 2

CONTROL

 (Exp.) 2021-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE		
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE			
			60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.6	2.6			± 0.4	7.5		7.4	7.3				± 1.0	18.3
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.17	2.13	2.14		± 0.16	4.58	4.55	4.58		± 0.20	5.15	5.08	5.16		± 0.25
		g/dl	6.0	5.9	6.0		± 0.4	13.9	13.7	13.8		± 0.5	16.8	16.5	16.7		± 0.6
HB	HGB	g/l	60	59	60		± 4	139	137	138		± 5	168	165	167		± 6
		mmol/l	3.73	3.66	3.73		± 0.25	8.63	8.51	8.57		± 0.31	10.43	10.25	10.37		± 0.37
HT	HCT	%	17.1	16.6	16.9		± 1.5	38.0	37.3	38.0		± 2.0	45.8	44.7	44.9		± 2.5
		l/l	0.171	0.166	0.169		± 0.015	0.380	0.373	0.380		± 0.020	0.458	0.447	0.449		± 0.025
VGM	MCV	µm ³ ·fl	79	78	79		± 5	83	82	83		± 5	89	88	87		± 5
TGMH	MCH	pg	27.6	27.7	28.0		± 2.0	30.3	30.1	30.1		± 2.0	32.6	32.5	32.4		± 2.5
		fmol	1.72	1.72	1.74		± 0.12	1.88	1.87	1.87		± 0.12	2.03	2.02	2.01		± 0.16
		g/dl	35.0	35.5	35.5		± 3.0	36.6	36.7	36.3		± 3.0	36.7	36.9	37.2		± 3.0
CCMH	MCHC	g/l	350	355	355		± 30	366	367	363		± 30	367	369	372		± 30
		mmol/l	21.73	22.05	22.04		± 1.86	22.71	22.80	22.54		± 1.86	22.76	22.92	23.10		± 1.86
IDR	RDW	%	13.5	12.5	13.5		± 4.0	12.5	12.5	12.5		± 4.0	12.0	11.5	11.5		± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	69	70		± 20	247	239	240		± 30	494	489	482		± 50
VPM	MPV	µm ³ ·fl	9.4	9.6	9.0		± 2.0	9.5	9.8	9.1		± 2.0	9.1	9.3	8.7		± 2.0
		#	1.35	1.29	1.43		± 0.35	4.04	4.00	4.12		± 0.90	12.70	12.28	12.58		± 1.90
		%	50.1	49.8	55.0		± 10.0	53.8	54.0	56.5		± 10.0	69.4	69.0	71.5		± 10.0
		#	0.82	0.82	0.74		± 0.33	2.41	2.52	2.34		± 0.70	2.75	2.83	2.55		± 1.50
		%	30.4	31.4	28.5		± 12.0	32.1	34.0	32.0		± 8.0	15.0	15.9	14.5		± 8.0
		#	0.18	0.14	0.13		± 0.13	0.56	0.38	0.37		± 0.37	1.15	1.07	0.88		± 0.88
		%	6.6	5.2	5.0		± 5.0	7.5	5.1	5.0		± 5.0	6.3	6.0	5.0		± 5.0
		#	0.27	0.27	0.21		± 0.21	0.26	0.27	0.22		± 0.22	0.93	0.89	0.79		± 0.79
		%	9.9	10.5	8.0		± 8.0	3.4	3.7	3.0		± 3.0	5.1	5.0	4.5		± 4.5
		#	0.08	0.08	0.09		± 0.08	0.24	0.24	0.26		± 0.24	0.77	0.73	0.79		± 0.73
		%	3.0	3.1	3.5		± 3.0	3.2	3.2	3.5		± 3.2	4.2	4.1	4.5		± 4.1

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44