


**LOT** PX 432  
Rev 3


**CONTROL**

 (Exp.) 2022-01-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4			± 0.4	7.5	7.5			± 1.0	17.4	17.4			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35			± 0.16	4.50	4.50			± 0.20	5.00	5.00			± 0.25			
HB HGB	g/dl	6.9	6.9			± 0.4	13.2	13.2			± 0.5	16.1	16.1			± 0.6			
	g/l	69	69			± 4	132	132			± 5	161	161			± 6			
	mmol/l	4.28	4.28			± 0.25	8.20	8.20			± 0.31	10.00	10.00			± 0.37			
HT HCT	%	20.0	20.0			± 1.5	36.9	36.9			± 2.0	44.0	44.0			± 2.5			
	l/l	0.200	0.200			± 0.015	0.369	0.369			± 0.020	0.440	0.440			± 0.025			
VGM MCV	µm <sup>3</sup> fl	85	85			± 5	82	82			± 5	88	88			± 5			
TGMH MCH	pg	29.4	29.4			± 2.0	29.3	29.3			± 2.0	32.2	32.2			± 2.5			
	fmol	1.82	1.82			± 0.12	1.82	1.82			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	34.5	34.5			± 3.0	35.8	35.8			± 3.0	36.6	36.6			± 3.0			
	g/l	345	345			± 30	358	358			± 30	366	366			± 30			
	mmol/l	21.45	21.45			± 1.86	22.21	22.21			± 1.86	22.72	22.72			± 1.86			
IDR RDW	%	15.0	15.0			± 4.0	14.0	14.0			± 4.0	13.0	13.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	69			± 20	253	253			± 30	492	492			± 50			
VPM MPV	µm <sup>3</sup> fl	10.5	10.5			± 2.0	9.5	9.5			± 2.0	10.2	10.2			± 2.0			
NEUT	#	1.34	1.34			± 0.35	4.19	4.19			± 0.90	12.16	12.16			± 1.90			
	%	56.0	56.0			± 10.0	55.9	55.9			± 10.0	69.9	69.9			± 10.0			
LYMPHO	#	0.72	0.72			± 0.33	2.42	2.42			± 0.70	2.73	2.73			± 1.50			
	%	30.1	30.1			± 12.0	32.3	32.3			± 8.0	15.7	15.7			± 8.0			
MONO	#	0.09	0.09			± 0.09	0.34	0.34			± 0.34	0.82	0.82			± 0.82			
	%	3.9	3.9			± 3.9	4.5	4.5			± 4.5	4.7	4.7			± 4.7			
EOS	#	0.16	0.16			± 0.16	0.30	0.30			± 0.30	0.94	0.94			± 0.94			
	%	6.6	6.6			± 6.6	4.0	4.0			± 4.0	5.4	5.4			± 5.4			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.75	0.75			± 0.75			
	%	3.4	3.4			± 3.4	3.3	3.3			± 3.3	4.3	4.3			± 4.3			

**LOT** PX 432  
Rev 3

**CONTROL**

 (Exp.) 2022-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5		± 0.4	7.5	7.5	7.5		± 1.0	17.9	17.6	17.8		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43	2.40	2.38		± 0.16	4.54	4.50	4.48		± 0.20	5.05	4.97	5.05		± 0.25	
		g/dl	7.0	6.9	7.0		± 0.4	13.2	13.0	13.1		± 0.5	16.1	15.9	16.1		± 0.6	
HB	HGB	g/l	70	69	70		± 4	132	130	131		± 5	161	159	161		± 6	
		mmol/l	4.35	4.28	4.35		± 0.25	8.20	8.07	8.14		± 0.31	10.00	9.87	10.00		± 0.37	
HT	HCT	%	19.7	19.4	19.5		± 1.5	36.3	36.0	36.3		± 2.0	44.4	43.2	44.2		± 2.5	
		l/l	0.197	0.194	0.195		± 0.015	0.363	0.360	0.363		± 0.020	0.444	0.432	0.442		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	81	81	82.0		± 5	80	80	81.0		± 5	88	87	87.5		± 5	
TGMH	MCH	pg	28.8	28.8	29.4		± 2.0	29.1	28.9	29.2		± 2.0	31.9	32.0	31.9		± 2.5	
		fmol	1.79	1.79	1.83		± 0.12	1.81	1.79	1.82		± 0.12	1.98	1.99	1.98		± 0.16	
		g/dl	35.6	35.5	35.9		± 3.0	36.3	36.1	36.1		± 3.0	36.2	36.8	36.4		± 3.0	
CCMH	MCHC	g/l	356	355	359		± 30	363	361	361		± 30	362	368	364		± 30	
		mmol/l	22.09	22.04	22.27		± 1.86	22.57	22.43	22.42		± 1.86	22.50	22.84	22.63		± 1.86	
IDR	RDW	%	14.5	14.0	13.0		± 4.0	13.5	13.5	13.0		± 4.0	13.0	13.0	12.5		± 4.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	71	70	70		± 20	258	255	253		± 30	499	510	495		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	10.0	10.7	9.8		± 2.0	9.0	9.1	8.6		± 2.0	9.7	10.0	9.2		± 2.0	
		#	1.40	1.40	1.48		± 0.35	4.25	4.25	4.39		± 0.90	12.44	12.23	12.64		± 1.90	
		%	58.5	58.4	59.0		± 10.0	56.6	56.7	58.5		± 10.0	69.5	69.5	71.0		± 10.0	
		#	0.68	0.67	0.69		± 0.33	2.38	2.27	2.21		± 0.70	2.85	2.59	2.58		± 1.50	
		%	28.2	27.8	27.5		± 12.0	31.7	30.2	29.5		± 8.0	15.9	14.7	14.5		± 8.0	
		#	0.09	0.10	0.10		± 0.09	0.34	0.42	0.38		± 0.34	0.91	1.07	0.98		± 0.91	
		%	3.7	4.0	4.0		± 3.7	4.5	5.6	5.0		± 4.5	5.1	6.1	5.5		± 5.1	
		#	0.15	0.15	0.15		± 0.15	0.29	0.31	0.26		± 0.26	0.93	0.95	0.89		± 0.89	
		%	6.2	6.4	6.0		± 6.0	3.8	4.1	3.5		± 3.5	5.2	5.4	5.0		± 5.0	
		#	0.08	0.08	0.09		± 0.08	0.26	0.26	0.26		± 0.26	0.77	0.76	0.71		± 0.71	
		%	3.4	3.4	3.5		± 3.4	3.4	3.4	3.5		± 3.4	4.3	4.3	4.0		± 4.0	

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.45