

**LOT** PX 407  
Rev 1

**CONTROL**

 (Exp.) 2017-11-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4	2.5	± 0.4	7.7	7.6	7.8	7.6	7.8	± 1.0	18.8	18.6	18.8	18.6	19.4	± 2.2		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43	2.39	2.43	2.39	2.40	± 0.16	4.65	4.67	4.67	4.67	4.65	± 0.20	5.05	5.10	5.08	5.10	5.03	± 0.25		
HB	HGB	g/dl	7.0	6.9	6.9	6.9	6.8	± 0.4	13.9	13.8	13.9	13.8	13.7	± 0.5	15.9	16.0	15.9	16.0	15.7	± 0.6		
		g/l	70	69	69	69	68	± 4	139	138	139	138	137	± 5	159	160	159	160	157	± 6		
HT	HCT	mmol/l	4.35	4.28	4.28	4.28	4.22	± 0.25	8.63	8.57	8.63	8.57	8.51	± 0.31	9.87	9.94	9.87	9.94	9.75	± 0.37		
		%	19.7	19.8	19.4	19.8	19.2	± 1.5	38.1	38.8	37.8	38.8	38.1	± 2.0	44.4	44.4	44.2	44.4	43.8	± 2.5		
		l/l	0.197	0.198	0.194	0.198	0.192	± 0.015	0.381	0.388	0.378	0.388	0.381	± 0.020	0.444	0.444	0.442	0.444	0.438	± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	81	83	80	83	80.0	± 5	82	83	81	83	82.0	± 5	88	87	87	87	87.0	± 5		
TGMH	MCH	pg	28.8	28.9	28.4	28.9	28.3	± 2.0	29.9	29.6	29.8	29.6	29.5	± 2.0	31.5	31.4	31.3	31.4	31.2	± 2.5		
		fmol	1.79	1.79	1.76	1.79	1.76	± 0.12	1.86	1.84	1.85	1.84	1.83	± 0.12	1.96	1.95	1.94	1.95	1.94	± 0.16		
CCMH	MCHC	g/dl	35.6	34.8	35.5	34.8	35.4	± 3.0	36.5	35.6	36.7	35.6	35.9	± 3.0	35.8	36.1	36.0	36.1	35.9	± 3.0		
		g/l	356	348	355	348	354	± 30	365	356	367	356	359	± 30	358	361	360	361	359	± 30		
		mmol/l	22.09	21.60	22.04	21.60	21.99	± 1.86	22.64	22.11	22.82	22.11	22.31	± 1.86	22.22	22.39	22.34	22.39	22.28	± 1.86		
IDR	RDW	%	13.2	13.5	13.0	13.0	13.0	± 4.0	13.3	13.3	12.8	12.6	12.5	± 4.0	13.0	13.5	12.5	12.5	12.2	± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	68	68	68	70	± 20	247	245	245	245	240	± 30	506	505	525	505	500	± 50		
VPM	MPV	µm <sup>3</sup> ; fl	9.0	9.0	9.0	9.0	8.5	± 2.0	9.1	9.2	9.0	9.2	8.7	± 2.0	8.4	8.5	8.4	8.5	8.1	± 2.0		
NEUT		#	1.35	1.35	1.35	1.35	1.45	± 0.35	4.43	4.34	4.39	4.34	4.49	± 0.90	13.07	12.72	12.86	12.72	13.58	± 1.90		
		%	56.2	56.1	56.2	56.1	58.0	± 10.0	57.5	57.1	56.3	57.1	57.5	± 10.0	69.5	68.4	68.4	68.4	70.0	± 10.0		
LYMPHO		#	0.71	0.70	0.69	0.70	0.69	± 0.33	2.33	2.26	2.30	2.26	2.22	± 0.70	2.76	2.73	2.58	2.73	2.43	± 1.50		
		%	29.6	29.3	28.7	29.3	27.5	± 12.0	30.3	29.7	29.5	29.7	28.5	± 8.0	14.7	14.7	13.7	14.7	12.5	± 8.0		
MONO		#	0.11	0.11	0.13	0.11	0.14	± 0.11	0.38	0.43	0.52	0.43	0.55	± 0.38	1.00	1.17	1.39	1.17	1.46	± 1.00		
		%	4.4	4.7	5.4	4.7	5.5	± 4.4	4.9	5.7	6.7	5.7	7.0	± 4.9	5.3	6.3	7.4	6.3	7.5	± 5.3		
EOS		#	0.16	0.16	0.15	0.16	0.14	± 0.14	0.28	0.29	0.30	0.29	0.27	± 0.27	1.13	1.13	1.13	1.13	1.07	± 1.07		
		%	6.5	6.5	6.3	6.5	5.5	± 5.5	3.7	3.8	3.9	3.8	3.5	± 3.5	6.0	6.1	6.0	6.1	5.5	± 5.5		
BASO		#	0.08	0.08	0.08	0.08	0.09	± 0.08	0.28	0.28	0.28	0.28	0.27	± 0.27	0.85	0.84	0.85	0.84	0.87	± 0.84		
		%	3.3	3.4	3.4	3.4	3.5	± 3.3	3.6	3.7	3.6	3.7	3.5	± 3.5	4.5	4.5	4.5	4.5	4.5	± 4.5		

# ABX Difftrol



**LOT** PX 407  
Rev 1

**CONTROL**

(Exp.) 2017-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					
			120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS		
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS			120 RETIC	DF120	DF NEXUS							
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6	2.6		± 0.4	8.0	8.0	8.0		± 1.0	20.5	20.5	20.5		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.44	2.44		± 0.16	4.68	4.68	4.68		± 0.20	5.13	5.13	5.13		± 0.25	
HB	HGB	g/dl	7.0	7.0	7.0		± 0.4	13.8	13.8	13.8		± 0.5	15.8	15.8	15.8		± 0.6	
		g/l	70	70	70		± 4	138	138	138		± 5	158	158	158		± 6	
		mmol/l	4.35	4.35	4.35		± 0.25	8.57	8.57	8.57		± 0.31	9.81	9.81	9.81		± 0.37	
HT	HCT	%	20.3	20.3	20.3		± 1.5	38.8	38.8	38.8		± 2.0	45.1	45.1	45.1		± 2.5	
		l/l	0.203	0.203	0.203		± 0.015	0.388	0.388	0.388		± 0.020	0.451	0.451	0.451		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	83	83	83		± 5	83	83	83		± 5	88	88	88		± 5	
TGMH	MCH	pg	28.7	28.7	28.7		± 2.0	29.5	29.5	29.5		± 2.0	30.8	30.8	30.8		± 2.5	
		fmol	1.78	1.78	1.78		± 0.12	1.83	1.83	1.83		± 0.12	1.91	1.91	1.91		± 0.16	
CCMH	MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.5	35.5	35.5		± 3.0	35.0	35.0	35.0		± 3.0	
		g/l	346	346	346		± 30	355	355	355		± 30	350	350	350		± 30	
		mmol/l	21.46	21.46	21.46		± 1.86	22.06	22.06	22.06		± 1.86	21.73	21.73	21.73		± 1.86	
IDR	RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	15.2	15.2	15.2		± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72		± 20	255	255	255		± 30	500	500	500		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.1	9.1	9.1		± 2.0	9.1	9.1	9.1		± 2.0	8.5	8.5	8.5		± 2.0	
NEUT		#	1.46	1.52	1.52		± 0.35	4.66	4.77	4.77		± 0.90	14.50	14.80	14.80		± 1.90	
		%	56.3	58.4	58.4		± 10.0	58.3	59.6	59.6		± 10.0	70.7	71.8	71.8		± 10.0	
LYMPHO		#	0.74	0.68	0.68		± 0.33	2.25	2.10	2.10		± 0.70	2.73	2.52	2.52		± 1.50	
		%	28.6	26.0	26.0		± 12.0	28.1	26.3	26.3		± 8.0	13.3	12.3	12.3		± 8.0	
MONO		#	0.15	0.16	0.16		± 0.15	0.59	0.61	0.61		± 0.59	1.76	1.74	1.74		± 1.74	
		%	5.9	6.0	6.0		± 5.9	7.4	7.6	7.6		± 7.4	8.6	8.5	8.5		± 8.5	
EOS		#	0.17	0.18	0.18		± 0.17	0.26	0.28	0.28		± 0.26	1.00	1.00	1.00		± 1.00	
		%	6.7	7.1	7.1		± 6.7	3.2	3.5	3.5		± 3.2	4.9	4.9	4.9		± 4.9	
BASO		#	0.07	0.07	0.07		± 0.07	0.24	0.24	0.24		± 0.24	0.51	0.51	0.51		± 0.51	
		%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5	

Ref: TEMP-0821 Rev.40 FRONT / RECTO 1300032435-A