

LOT PX 409
Rev 1


CONTROL

 (Exp.) 2018-03-05
(YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR			80 XL80		XLR			80 XL80	XLR				80 XL80	XLR		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	17.7	17.7			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.27	2.27			± 0.16	4.55	4.55			± 0.20	5.09	5.09			± 0.25			
HB HGB	g/dl	6.7	6.7			± 0.4	13.4	13.4			± 0.5	16.4	16.4			± 0.6			
	g/l	67	67			± 4	134	134			± 5	164	164			± 6			
HT HCT	mmol/l	4.16	4.16			± 0.25	8.32	8.32			± 0.31	10.18	10.18			± 0.37			
	%	18.8	18.8			± 1.5	36.4	35.9			± 2.0	44.3	43.8			± 2.5			
VGM MCV	l/l	0.188	0.188			± 0.015	0.364	0.359			± 0.020	0.443	0.438			± 0.025			
	µm ³ ; fl	83	83			± 5	80	79			± 5	87	86			± 5			
TGMH MCH	pg	29.5	29.5			± 2.0	29.5	29.5			± 2.0	32.2	32.2			± 2.5			
	fmol	1.83	1.83			± 0.12	1.83	1.83			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	35.6	35.6			± 3.0	36.8	37.3			± 3.0	37.0	37.5			± 3.0			
	g/l	356	356			± 30	368	373			± 30	370	375			± 30			
	mmol/l	22.08	22.08			± 1.86	22.86	23.15			± 1.86	23.00	23.27			± 1.86			
IDR RDW	%	13.0	12.5			± 4.0	13.0	12.5			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65			± 20	250	246			± 30	500	495			± 50			
VPM MPV	µm ³ ; fl	9.4	9.4			± 2.0	9.4	9.4			± 2.0	8.7	8.7			± 2.0			
NEUT	#	1.41	1.41			± 0.35	4.11	4.11			± 0.90	12.12	12.12			± 1.90			
	%	56.2	56.2			± 10.0	55.5	55.5			± 10.0	68.5	68.5			± 10.0			
LYMPHO	#	0.70	0.70			± 0.33	2.23	2.23			± 0.70	2.92	2.92			± 1.50			
	%	27.8	27.8			± 12.0	30.1	30.1			± 8.0	16.5	16.5			± 8.0			
MONO	#	0.17	0.17			± 0.17	0.52	0.52			± 0.52	1.03	1.03			± 1.03			
	%	6.6	6.6			± 6.6	7.0	7.0			± 7.0	5.8	5.8			± 5.8			
EOS	#	0.15	0.15			± 0.15	0.28	0.28			± 0.28	0.85	0.85			± 0.85			
	%	6.1	6.1			± 6.1	3.8	3.8			± 3.8	4.8	4.8			± 4.8			
BASO	#	0.08	0.08			± 0.08	0.27	0.27			± 0.27	0.78	0.78			± 0.78			
	%	3.3	3.3			± 3.3	3.6	3.6			± 3.6	4.4	4.4			± 4.4			

LOT PX 409
Rev 1

CONTROL

 (Exp.) **2018-03-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP					
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.6		± 0.4	7.5	7.4	7.5		± 1.0	18.3	17.9	18.0		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.33	2.34		± 0.16	4.54	4.58	4.56		± 0.20	5.07	5.12	5.10		± 0.25	
HB	HGB	g/dl	6.7	6.7	6.7		± 0.4	13.4	13.4	13.5		± 0.5	16.3	16.4	16.3		± 0.6	
		g/l	67	67	67		± 4	134	134	135		± 5	163	164	163		± 6	
		mmol/l	4.16	4.16	4.16		± 0.25	8.32	8.32	8.38		± 0.31	10.12	10.18	10.12		± 0.37	
HT	HCT	%	18.7	18.6	18.7		± 1.5	36.3	36.2	36.0		± 2.0	44.1	44.0	43.6		± 2.5	
		l/l	0.187	0.186	0.187		± 0.015	0.363	0.362	0.360		± 0.020	0.441	0.440	0.436		± 0.025	
VGM	MCV	µm ³ ; fl	80	80	80.0		± 5	80	79	79.0		± 5	87	86	85.5		± 5	
TGMH	MCH	pg	28.6	28.8	28.6		± 2.0	29.5	29.3	29.6		± 2.0	32.1	32.0	32.0		± 2.5	
		fmol	1.78	1.79	1.78		± 0.12	1.83	1.82	1.84		± 0.12	2.00	1.99	1.98		± 0.16	
CCMH	MCHC	g/dl	35.8	35.9	35.8		± 3.0	36.9	37.0	37.5		± 3.0	37.0	37.2	37.4		± 3.0	
		g/l	358	359	358		± 30	369	370	375		± 30	370	372	374		± 30	
		mmol/l	22.23	22.32	22.23		± 1.86	22.91	23.00	23.27		± 1.86	22.95	23.13	23.21		± 1.86	
IDR	RDW	%	13.5	13.0	12.5		± 4.0	13.7	13.0	12.5		± 4.0	13.0	12.0	11.8		± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68	70		± 20	255	255	250		± 30	500	510	495		± 50	
VPM	MPV	µm ³ ; fl	9.2	9.4	8.7		± 2.0	9.2	9.4	8.8		± 2.0	8.6	8.7	8.2		± 2.0	
NEUT		#	1.39	1.37	1.46		± 0.35	4.19	4.08	4.24		± 0.90	12.74	12.35	12.71		± 1.90	
		%	55.7	54.7	56.0		± 10.0	55.9	55.2	56.5		± 10.0	69.6	69.0	70.6		± 10.0	
LYMPHO		#	0.72	0.71	0.77		± 0.33	2.29	2.21	2.29		± 0.70	2.91	2.79	2.88		± 1.50	
		%	28.7	28.5	29.5		± 12.0	30.5	29.9	30.5		± 8.0	15.9	15.6	16.0		± 8.0	
MONO		#	0.16	0.19	0.16		± 0.16	0.48	0.56	0.45		± 0.45	1.02	1.11	0.85		± 0.85	
		%	6.3	7.5	6.0		± 6.0	6.4	7.6	6.0		± 6.0	5.6	6.2	4.7		± 4.7	
EOS		#	0.15	0.15	0.13		± 0.13	0.27	0.27	0.26		± 0.26	0.88	0.88	0.81		± 0.81	
		%	6.0	6.0	5.0		± 5.0	3.6	3.7	3.5		± 3.5	4.8	4.9	4.5		± 4.5	
BASO		#	0.08	0.08	0.09		± 0.08	0.27	0.27	0.26		± 0.26	0.75	0.77	0.76		± 0.75	
		%	3.3	3.3	3.5		± 3.3	3.6	3.6	3.5		± 3.5	4.1	4.3	4.2		± 4.1	

Ref: TEMP-0821 Rev.42 FRONT / RECTO 1300040596