


LOT PX 430
Rev 2


CONTROL

 (Exp.) 2021-09-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8			± 0.4	7.7	7.7			± 1.0	17.2	17.2			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.12	2.12			± 0.16	4.59	4.59			± 0.20	5.18	5.18			± 0.25			
HB HGB	g/dl	5.5	5.5			± 0.4	13.5	13.5			± 0.5	16.5	16.5			± 0.6			
	g/l	55	55			± 4	135	135			± 5	165	165			± 6			
	mmol/l	3.42	3.42			± 0.25	8.38	8.38			± 0.31	10.25	10.25			± 0.37			
HT HCT	%	16.1	16.1			± 1.5	37.2	37.2			± 2.0	45.1	45.1			± 2.5			
	l/l	0.161	0.161			± 0.015	0.372	0.372			± 0.020	0.451	0.451			± 0.025			
VGM MCV	µm ³ fl	76	76			± 5	81	81			± 5	87	87			± 5			
TGMH MCH	pg	25.9	25.9			± 2.0	29.4	29.4			± 2.0	31.9	31.9			± 2.5			
	fmol	1.61	1.61			± 0.12	1.83	1.83			± 0.12	1.98	1.98			± 0.16			
CCMH MCHC	g/dl	34.1	34.1			± 3.0	36.3	36.3			± 3.0	36.6	36.6			± 3.0			
	g/l	341	341			± 30	363	363			± 30	366	366			± 30			
	mmol/l	21.20	21.20			± 1.86	22.55	22.55			± 1.86	22.74	22.74			± 1.86			
IDR RDW	%	14.0	14.0			± 5.0	13.5	13.5			± 5.0	13.0	13.0			± 5.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	61	61			± 20	236	236			± 30	475	475			± 50			
VPM MPV	µm ³ fl	10.6	10.6			± 2.0	10.5	10.5			± 2.0	9.8	9.8			± 2.0			
NEUT	#	1.48	1.48			± 0.35	4.13	4.13			± 0.90	11.75	11.75			± 1.90			
	%	52.8	52.8			± 10.0	53.7	53.7			± 10.0	68.3	68.3			± 10.0			
LYMPHO	#	0.82	0.82			± 0.33	2.38	2.38			± 0.70	2.73	2.73			± 1.50			
	%	29.2	29.2			± 12.0	30.9	30.9			± 8.0	15.9	15.9			± 8.0			
MONO	#	0.19	0.19			± 0.19	0.51	0.51			± 0.51	1.14	1.14			± 1.14			
	%	6.8	6.8			± 6.8	6.6	6.6			± 6.6	6.6	6.6			± 6.6			
EOS	#	0.22	0.22			± 0.22	0.42	0.42			± 0.42	0.83	0.83			± 0.83			
	%	7.9	7.9			± 7.9	5.5	5.5			± 5.5	4.8	4.8			± 4.8			
BASO	#	0.09	0.09			± 0.09	0.25	0.25			± 0.25	0.76	0.76			± 0.76			
	%	3.3	3.3			± 3.3	3.3	3.3			± 3.3	4.4	4.4			± 4.4			

LOT PX 430
Rev 2

CONTROL

 (Exp.) 2021-09-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.8			± 0.4	7.6	7.7	7.5			± 1.0	18.2	17.9	17.8			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.19	2.15	2.11			± 0.16	4.60	4.52	4.53			± 0.20	5.18	5.11	5.16			± 0.25	
	g/dl	5.6	5.5	5.5			± 0.4	13.5	13.2	13.2			± 0.5	16.6	16.2	16.4			± 0.6	
HB HGB	g/l	56	55	55			± 4	135	132	132			± 5	166	162	164			± 6	
	mmol/l	3.48	3.42	3.42			± 0.25	8.38	8.20	8.20			± 0.31	10.31	10.06	10.18			± 0.37	
HT HCT	%	16.0	15.5	15.4			± 1.5	36.8	35.7	36.2			± 2.0	45.1	43.9	44.9			± 2.5	
	l/l	0.160	0.155	0.154			± 0.015	0.368	0.357	0.362			± 0.020	0.451	0.439	0.449			± 0.025	
VGM MCV	µm ³ ·fl	73	72	73.0			± 5	80	79	80.0			± 5	87	86	87.0			± 5	
TGMH MCH	pg	25.6	25.6	26.1			± 2.0	29.3	29.2	29.1			± 2.0	32.0	31.7	31.8			± 2.5	
	fmol	1.59	1.59	1.62			± 0.12	1.82	1.81	1.81			± 0.12	1.99	1.97	1.97			± 0.16	
	g/dl	35.0	35.5	35.7			± 3.0	36.7	37.0	36.4			± 3.0	36.8	36.9	36.5			± 3.0	
CCMH MCHC	g/l	350	355	357			± 30	367	370	364			± 30	368	369	365			± 30	
	mmol/l	21.75	22.06	22.17			± 1.86	22.78	22.96	22.62			± 1.86	22.87	22.89	22.69			± 1.86	
IDR RDW	%	14.0	13.5	13.5			± 4.0	13.0	12.5	12.5			± 4.0	12.0	12.0	11.7			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	68	61			± 20	249	245	235			± 30	490	490	470			± 50	
VPM MPV	µm ³ ·fl	10.3	10.4	9.4			± 2.0	9.6	9.8	9.4			± 2.0	9.0	9.1	8.6			± 2.0	
NEUT	#	1.49	1.46	1.50			± 0.35	4.10	4.45	4.31			± 0.90	12.30	12.76	12.73			± 1.90	
	%	53.3	52.1	53.5			± 10.0	54.0	57.8	57.5			± 10.0	67.6	71.3	71.5			± 10.0	
LYMPHO	#	0.79	0.77	0.83			± 0.33	2.36	2.29	2.33			± 0.70	3.09	2.79	2.85			± 1.50	
	%	28.3	27.6	29.5			± 12.0	31.1	29.7	31.0			± 8.0	17.0	15.6	16.0			± 8.0	
MONO	#	0.22	0.27	0.21			± 0.21	0.52	0.28	0.30			± 0.28	1.15	0.75	0.71			± 0.71	
	%	8.0	9.8	7.5			± 7.5	6.8	3.7	4.0			± 3.7	6.3	4.2	4.0			± 4.0	
EOS	#	0.20	0.21	0.17			± 0.17	0.37	0.42	0.30			± 0.30	0.87	0.84	0.80			± 0.80	
	%	7.1	7.6	6.0			± 6.0	4.9	5.5	4.0			± 4.0	4.8	4.7	4.5			± 4.5	
BASO	#	0.09	0.08	0.10			± 0.08	0.24	0.25	0.26			± 0.24	0.78	0.75	0.71			± 0.71	
	%	3.3	2.9	3.5			± 2.9	3.2	3.3	3.5			± 3.2	4.3	4.2	4.0			± 4.0	

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.45